

NOTICE OF VIOLATION

Industrial Gamma Inspection

License No. 24-19850-01

As a result of the inspection conducted on June 12, 1986, and in accordance with 10 CFR Part 2, Appendix C - General Statement of Policy and Procedure for NRC Enforcement Actions (1985), the following violations were identified:

- 1.0 10 CFR 34.24 requires that each survey instrument used to conduct physical radiation surveys be calibrated at intervals not to exceed three months and after each instrument servicing, and a record shall be maintained of the results of each instrument calibration and date thereof for two years after the date of calibration.

Contrary to the above, the licensee failed to maintain the proper instrument calibration records performed on survey instrument Serial No. 7660 during the period from January 11, 1985 to November 25, 1985, and survey instrument Serial No. 1046 during the period from March 14, 1985 to January 22, 1986.

This is a Severity Level V violation (Supplement VI).

- 2.0 10 CFR 20.401(a) requires each licensee to maintain records showing the radiation exposures of all individuals for whom personnel monitoring is required. Such records shall be kept on Form NRC-5 or on clear and legible records containing all the information required by Form NRC-5. 10 CFR 20.401(c)(1) requires the licensee to maintain the above records until the Commission Authorizes disposition.

Contrary to the above, the licensee failed to maintain the exposure records (NRC-5 Forms) of personnel monitored during the period from February 25, 1985 to April 24, 1986.

This is a Severity Level V violation (Supplement IV).

- 3.0 10 CFR 34.33(c) requires pocket dosimeters be checked at periods not to exceed one year for correct response to radiation.

Contrary to the above, no checks of pocket dosimeters were performed during 1985, and to the day of the inspection, June 12, 1986.

This is a Severity Level IV violation (Supplement VI).

- 4.0 10 CFR 20.407(b) requires that you submit to the Commission within the first quarter of each calendar year, a report of exposures recorded for individuals under your licensed program for the preceding calendar year.


Contrary to the above, as of June 12, 1986, the licensee had failed to submit the required report for calendar year 1985.

This is a Severity Level V violation (Supplement IV).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Dated

6/26/86


D. J. Sreniawski, Chief
Nuclear Materials Safety
Section 2