

SEP 24 1986

Docket No. : 030-01221
License No.: 04-01935-03
Control No.: 70433

V. A. Medical Center
2615 East Clinton Avenue
Fresno, California 93703

Attention: Mr. J.J. Mason
Director

Gentlemen:

Enclosed is the NRC license amendment which you requested.

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify us so that we can provide appropriate corrections and answers.

Please note that Items 8.H. and 8.I. do not limit the number of Gd-153 and I-125 sources which you may possess. This will allow you to purchase sources as they are needed. Accordingly, License condition 20. has been added to your license to assure that sealed sources are inventoried every 6 months.

Any future correspondence relating to your license should specifically reference your license and docket numbers to expedite your inquiry.

Sincerely,

Beth A. Riedlinger
Health Physicist (Licensing)
Nuclear Materials Safety Section

Enclosure: Attachment A
Amendment No. 40 to License No. 04-01935-03

8610070336 860924
REG5 LIC30
04-01935-03 PDR

OFFICE	RV <i>BA</i>					
RNAME	Riedlinger:fr					
DATE	9/24/86					

OFFICIAL RECORD COPY

U.S. GPO 1983-400-247