

AUG 6 1986

West Shore Hospital  
ATTN: Burton O. Parks  
Administrator  
1465 East Parkdale Avenue  
Manistee, MI 49660

Gentlemen:

Enclosed is Amendment No. 08 to your NRC License No. 21-16277-01 in accordance with your request.

As stated in letter dated February 23, 1981, please note that you are required to follow procedures listed in Appendix L of Regulatory Guide 10.8, October, 1980 for iodine-125 seeds as listed in Group VI of Schedule A, Section 35.100 of 10 CFR 35.

Also, please be aware that you are required to perform dose calibrator linearity tests on a quarterly basis as stated in License Condition No. 13.

Please review your license carefully to assure that you understand and are in compliance with the terms and conditions contained therein. If you have questions, please contact our office at (312) 790-5625.

Sincerely,

Original Signed By  
Kathy J. Graden  
Materials Licensing Section

Enclosure: Amendment No. 08

8610070400 860806  
REG3 LIC30  
21-16277-01 FDR

R111

KG  
Graden/pd  
8/1/86