



PRESBYTERIAN-UNIVERSITY OF PENNSYLVANIA MEDICAL CENTER

Fifty One North Thirty-ninth Street Philadelphia, Pennsylvania 19104

Area Code 215 662-8000

August 25, 1986

DAVID A. IDDENDEN, M.B., M.R.C.O.G., F.A.C.O.G.

Director

Department of Obstetrics & Gynecology

662-8798

John E. Glenn, Ph.D, Chief  
Nuclear Materials Safety Section B  
Division of Radiation Safety  
and Safeguards  
631 Park Avenue  
King of Prussia, Pa. 19406

Re: Inspection No. 86-01

Dear Mr. Glenn:

In response to your letter of August 4, 1986, I wish to indicate the corrective steps we are taking to insure that there is no repetition of the violation cited in that letter.

We have taken steps to insure that no Iodine-125 source is maintained on our premises longer than six months from the date of its' leak test prior to being dispatched to us. This step in itself should bring us in full compliance with our license. As an additional pre-caution I have arranged with local services to provide a leak test in our facility on demand, should for any reason an Iodine-125 source need to have its' six month leak test performed at our facility.

Many thanks.

Yours sincerely,

David A. Iddenden, M.B., M.R.C.O.G.,  
F.A.C.O.G.

DAI/mm

cc:

8611110472 861107  
REG1 LIC30  
37-20856-01 PDR

SEP 09 1986