

MERCY CATHOLIC MEDICAL CENTER
OF SOUTHEASTERN PENNSYLVANIA

APR 21 P2:14

April 14, 1986

RANCH

Mr. Jack Davis
Nuclear Materials Safety Section A
Division of Radiation Safety and Safeguards
U.S. Nuclear Regulatory Commission
Region I
631 Park Avenue
King of Prussia, Pa. 19406

RE: Amendment to License
#37-00993-04

Dear Mr. Davis:

We would like to amend our License #37-00993-04 for the CGR-MeV Alcyon II Cobalt Unit at the Misericordia Division of the Mercy Catholic Medical Center, as follows:

To permit calibration of the Unit with a water bath, the machine must be operated with the retractable beam stopped retracted. There is not sufficient distance to place the water phantom at 80cm. T.S.D.

We request permission to install a bypass with a key in the gantry.

The Unit would be used in the bypass condition once a year and the key controlled by the Physicist.

During operation without the primary beam stopper, the room directly under the Cobalt Room would be evacuated, locked and posted. It is anticipated that the measurements would be made on weekends or at night.

A survey of surrounding areas with the primary shield and water bath in position, would be made prior to any extensive measurements.

Log	Apr. 13 th
Remitter	
Check No.	37648
Amount	\$230
Fee Category	2A
Type of Fee	Amendment
Date Check Rec'd.	4/21/86
Date Completed	4/24/86

Emberley

"OFFICIAL RECORD COPY"

105347


MISERICORDIA DIVISION
54th Street and Cedar Avenue
Philadelphia, Pa. 19143
(215) 748-9000

8607210132 860515
REG1 LIC30
37-00993-04 PDR

APR 16 1986

Please let me know whether this amendment can be added to our license. If you need additional information of it, giving technical details, please contact James N. Bierly, Ph.D. and myself in writing and we would be glad to supply additional details.

Sincerely yours,

A handwritten signature in dark ink, appearing to read "C. Jules Rominger", followed by a large, stylized flourish or checkmark.

C. Jules Rominger, M.D.

CJR/me
Ck. encl.

RECEIVED: William O. Miller, Chief
License Fee Management Branch
Office of Administration

John E. Glenn, Chief
Nuclear Materials Section B
Division of Engineering and
Technical Programs

03000460
02300
8/9.0

LICENSE FEE TRANSMITTAL

A. REGION I

1. APPLICATION ATTACHED

Applicant/Licensee: Mercy Catholic Medical Center

Application Dated: 4/14/86

Control No.: 105347

License No.: 37-00993-04

2. FEE ATTACHED

Amount: \$ 230.00

Check No.: 37648 / 24201

3. COMMENTS

Signed Brandie Pilchek

Date 4/17/86

B. LICENSE FEE MANAGEMENT BRANCH

1. Fee Category and Amount: 7A (\$230)

2. Correct Fee Paid. Application may be processed for:

Amendment ✓

Renewal

License

Signed D Jackson

Date 4/24/86