mantic

envirolab, incorporated

249 boston post road east lyme, ct 06333

(203) 739-4080

U.S. Nuclear Regulatory Commission 11/26/86 Repion I 631 Park Ave. King of Prussia, PA 19406 Docket No. 030-22092

License No. 06-20800-01

Gentlemen:

In response to the Notice of Violation which we received from your offices 11/3/86 please find below the corrective steps we have taken.

- A) A letter of amendment of license and appropriate fees have been forwarded to US NRC Region I.
- B) We have started a wipe test program with CEP, Inc., Santa Fe. New Mexico which will be on a 6 month basis.
- C) A photo copy of our license and conditions have now been posted above the gas chromatograph as well as a notation that specific operating procedures may be located in the operating manual of the instrument.

If you have any other comments, please notify us.

Sincerely,

William S. Gengin William S. Gedrgian,

Director

RECEIVED-REGION 1 1987 DEC -O PM 2: 00

a keep and

State of the State

Carlo de la compania La compania de la co