

JAN 14 1987

Docket No. 030-13596

License No. 06_17892-01

New Milford Hospital
ATTN: Mr. Richard E. Pugh
President and Executive Director
21 Elm Street
New Milford, Connecticut 06776

Gentlemen:

Subject: Inspection No. 030-13596/86-01

This refers to your letter dated November 26, 1986 in response to our letter dated November 4, 1986.

Thank you for informing us of the corrective and preventive actions documented in your letter. These actions will be examined during a future inspection of your licensed program.

In accordance with the telephone conversation between Dr. White and Mr. Oberg of this office on December 17, 1986, we understand that instrument accuracy calibration checks of your dose calibrator will be performed presently and annually hereafter. Please advise us in writing if our understanding of this matter is not correct.

Your cooperation with us is appreciated.

Sincerely,

Original Signed By:

John R. White, Chief
Nuclear Materials Safety Section C,
Division of Radiation Safety and
Safeguards

cc:
Public Document Room (PDR)
Nuclear Safety Information Center (NSIC)
State of Connecticut

RETURN ORIGINAL TO
REGION I

8701200232 870114
REG1 LIC30
05-17892-01 PDR

"OFFICIAL RECORD COPY"

IE:07

JAN 14 1987

New Milford Hospital

2

bcc:

Region I Docket Room (w/concurrences)

RI:DRSS
Obergruejr

~~12/1/86~~
1/5/87

RI:DRSS
White

~~12/1/86~~
1/1/87

OFFICIAL RECORD COPY

RL NEW MILFORD - 0002.0.0
12/22/86