



PHONE: AREA CODE 717
898 2211

BEAR CREEK CONSTRUCTION CO.
LANDISVILLE, PENNSYLVANIA 17538

December 23, 1986

U.S. Nuclear Regulatory
Commission - Region 1
631 Park Avenue
King of Prussia, Pa., 19406

Attention: John D. Kinneman, Chief
Nuclear Materials

Re: Inspection No. 030-22313/86-01

Gentlemen:

As directed in your December 10, 1986 letter and the October 14, 1986 inspection, we have made the following corrective steps relative the cited violations:

- A. Shipping papers have been prepared and will accompany the gauge in transport to and from future temporary field locations.
- B. A leak test was performed on December 23, 1986, and the sample sent to Troxler for analysis. The result of the leak test will be maintained. Leak tests will be done in the future at 6 month intervals or prior to putting the gauge in service.
- C. An inventory log has been established & inventory records are being maintained.
- D. Copies of the dosimetry reports requested at the inspection have been received from R.S. Landauer, Jr. & Co. Dosimetry records are being maintained.
- E. The required documents have been posted.

Based on the above corrective steps, and steps we will take in the future, we feel we are now in full compliance with our license.

Very truly yours,

BEAR CREEK CONSTRUCTION CO.


Thomas R. Kahler, P.E.

8702030467 870122
REG1 LIC30
37-20901-01 PDR

TRK/ey