

50-10/237/249

DAP 9-1
Revision 13

Date 11/7/86

Item # 86-11-130

PROCEDURES MANAGER DISTRIBUTION ITEM

Distribution Item: EPIP 500-3, Rev. 3.

Instructions to Clerk: DOSR Stamp August 30, 1985.

Xerox and distribute in accordance with PM's
instructions.

Explanation: Correction to procedure to reflect change of
film badges to TLD's and on inventory of
Radiation Disaster cart change 1 bottle of
dilute hydrochloric acid to sodium hypochlorite
(bleach) dilute 10:1.

Dennis A. Ring
Approved by Procedures Manager

Transmittal Receipt

Register No. 72A & 72

(SS) Dm 11-12-86

Remove: EPIP 500-3.

Insert: Attached EPIP 500-3.

NOV171986

Sign and Date

(Return this form to the Procedures Manager)

Form (9-13)

1111b

8611260445 861107
PDR ADOCK 05000010
F PDR

A045
0/1

QUARTERLY
ST. JOSEPH HOSPITAL EMERGENCY
CART INVENTORY
(Primary Responsibility - Radiation-Chemistry Director)

A. PURPOSE

The purpose of this procedure is to describe the actions necessary to inventory the St. Joseph Hospital radiation emergency cart.

B. REFERENCES

None.

C. PREREQUISITES

None.

D. PRECAUTIONS

None.

E. LIMITATIONS AND ACTIONS

An inventory of the hospital emergency cart shall be made after each use of the hospital cart and at least once in each quarter of the year.

F. PROCEDURE

1. Station GSEP Coordinator will forward a copy of this procedure to appropriate Radiation-Chemistry Foremen for completion.
2. Call the hospital and arrange a convenient time for the inventory. (Refer to Station's Director's Phone List in EPIP Manual).
3. Collect the following items to be taken to the hospital for inventory:
 - a. One (1) recently calibrated CP.
 - b. One (1) copy of the hospital inventory checklist.
 - c. Five (5) D cell batteries (biannually).
Date changed ____/____/____.
 - d. One (1) recently calibrated GM.
 - e. Twelve (12) thermoluminescent dosimetry (TLD).
 - f. Twelve (12) finger ring film badges.

APPROVED

AUG 30 '85

D.O.S.R.

0065E
0125b

8601220432 3pp.

g. Eighteen (18) recently calibrated dosimeters.

(1) Twelve (12) O-200MR.

(2) Six (6) O-1R.

4. At the hospital:

a. Replace the GM and CP with the recently calibrated ones. RECORD new calibration dates on the attached Inventory of Radiation Disaster Cart sheet.

b. Replace the old cells with the new ones (biannually).

c. Replace all old TLD's and finger rings with new ones and ensure the expiration date of the new TLD's and finger rings exceeds the next inventory date (Note: the procedure is completed quarterly).

d. Replace the dosimeters (biannually).

Date changed ____/____/____.

e. Inventory all items in the cart using the attached inventory sheet.

f. Inventory the supplies in the Decon and Sample Taking Kits using the lists provided in the kit.

g. Note missing items on the attached inventory sheet and obtain replacements. Supplies missing from the Decon and Sampling Kits may be ordered from the company Medical Department. Other supplies are available in the Dresden Storeroom.

5. Forward completed procedure to Station GSEP Coordinator.

G. CHECKLISTS

Inventory of Radiation Disaster Cart sheet, (attached).

H. TECHNICAL SPECIFICATIONS REFERENCES

None.

APPROVED

AUG 30 '85

D.O.S.B.

INVENTORY OF RADIATION DISASTER CART

NUMBER ON CART

ITEM

1	GM Counter and Probe	Calibration Date _____
1	CP Survey Meter	Calibration Date _____
2	Dosimeter Charger	
6	0-1R Gamma & X-Ray Pencil Dosimeters	
12	0-200mR Gamma & X-Ray Pencil Dosimeters	
5	D Cell Batteries	
100 ft.	Radiation Rope	
4	Signs - High Radiation	
4	Signs - Radiation Area - Authorized Entry Only	
4	Signs - Airborne Radioactivity	
4	Signs - Radioactive Materials	
4	Signs - Radiation Area - Radioactive Materials - Authorized Entry Only	
4	Signs - "DO NOT ENTER"	
1 roll	Wrapping Paper, 36" wide	
6	Step-Off Pads - Check Shoes	
6	Step-Off Pads - Remove Protective Clothing	
6	Step-Off Pads - Remove Protective Footwear	
6 pair	Shoe Covers	
3	Full-Face Particulate Filter Respirators	
6	Half-Face Particulate Filter Respirators	
1 pkg.	Type H Filter Cartridge - 2 per package	
6 rolls	2" Masking Tape	
8	Bioassay Bottles	
Assorted	Plastic Bags - all sizes	
2 doz.	Shower Hooks	
12	Gummed Labels - "Rad. Material"	
1 doz.	Soft Scrubbing Brushes (in decon kit)	
2 gal.	Phisoex	
1 box	Detergent - Tide, Dreft	
1 box (200)	Cotton-Tipped Applicators	
2 pkg. (50)	Cotton Balls	
50	4" x 3" Gauze	
2	5 gallons Jerry Cans	
2 bottles	2000cc Saline - Irrigation (in decon kit)	
1 pkg. (100)	Paper Cups	
1 pkg.	Towels	
2 rolls	Adhesive Tape 2"	
2 pints	Hydrogen Peroxide (in decon kit)	
1 bottle	Sodium Hypochlorite (bleach) Dilute 10:1	
1	Decon Kit	
1	Sample Taking Kit	

APPROVED

AUG 30 '85

D.O.S.R.

NAME

DATE

All the above listed deficiencies have been corrected.

DATE COMPLETED

SIGNATURE

0065E
0125b