



# OsteoscanAffiliates,inc.

August 22, 1986

U.S. Nuclear Regulatory Commission  
Region I  
631 Park Ave.  
King of Prussia, PA 19406


Attn: Dr. Josephine M. Piccone:

Dear Dr. Piccone:

We would like to amend our NRC Materials License No. 29-27856-01 to allow use of the ND1100 Bone Mineral Analyzer at two other New Jersey locations. The new locations are 45 Jones Street, Jersey City, N.J. and 40 Fuld Street, Trenton, N.J. Enclosed is a letter authorizing use of the ND1100 at these locations, diagrams of the facilities, and statements of training for two additional physicians who are located at the Trenton facility. In addition to the two scanning locations described above, we would like the license amended to allow storage of the ND1100 at our new office location at 323 Godwin Ave., Midland Park, NJ. We will not perform scans at this location, but will store the ND1100 instrument in this facility between use at the other scanning sites. Our new office is a free-standing building (diagram enclosed) that will be locked when it is not occupied by Osteoscan personnel.

We would also like to amend that portion of our license that pertains to leak testing for new sources. As allowed by NRC regulations, we would like the option of not leak testing a new source, and returning it to AECL within six months of the AECL leak test. Enclosed is a check in the amount of \$120.00 to cover the license amendment fee.

Thank you for your consideration of this amendment request.

Sincerely,  
  
Gerald A. Bruno, Ph.D.  
President

Enclosure

Log	SEP 30
Remitter	BMA
Check No.	011665
Amount	\$120
Fee Category	7C
Type of Fee	Amendment
Date Check Rec'd.	9/3/86
Date Completed	9/4/86
Sgt.	A. Centraly

RECEIVED  
'86 SEP -3 A9:39  
U.S. NRC  
LIC. FEE MGMT. BRANCH

1986 AUG 28 PM 2 18  
RECEIVED-REGION I

8612010111 860924  
REG1 LIC30  
29-27856-01 PDR

**'OFFICIAL RECORD COPY'**

106070

ML18