



United Hospital Center

Clarksburg, West Virginia

September 16, 1986

Earl Wright
Radiation Specialist
Nuclear Materials Safety Section
US Nuclear Regulatory Commission -- Region II
101 Marietta Street, Suite 3100
Atlanta, Georgia 30303

RE: Radioactive Materials License No. 47-01458-01
United Hospital Center, Clarksburg, West Virginia

06 SEP 24 A10:35

Dear Mr. Wright,

Thank you for your prompt review of our application to amend the above referenced license.

Amendment 15 of our license permits the use of Xenon 133. We are not currently using Xenon for any studies and will not use Xenon during the temporary relocation of the two nuclear medicine imaging rooms. When the Nuclear Medicine Department is moved to its permanent location on the first floor, in 6 to 9 months, we will again apply to amend our license for the move and for the use of Xenon.

As we discussed today, our renovation schedule is as follows:

September 22nd -- Move both GE Nuclear cameras to renovated area.
September 24th -- A close out survey will be performed. A copy of this survey
or will be sent to you. The doors to the hot lab and to
25th cardiac imaging will be closed and sealed from the new
construction area. The other door to the hot lab will be
opened.
September 29th -- Construction workers will be given access to the closed out
imaging rooms to continue with the construction project.

Thank you for your verbal approval to amend our license and proceed with the schedule detailed above.

Sincerely,

Harry A Bishop, MD
Chief Radiologist
(Licensee -- ALL groups of
materials)

Robert F. O'Connor RT PA
Radiation Oncology
United Hospital Center
(304) 624-2995

cc: Mark Clement, Associate Administrator
Ron Malin, Radiology Manager
P.O. Box 1680, Clarksburg, W. Va. 26301

8612030657 860926
REG2 LIC30
47-01458-01 PDR

Official Copy

Earl Wright

CONVERSATION RECORD

TIME

1152

DATE

9/16/86

TYPE

☐ VISIT

☐ CONFERENCE

☒ TELEPHONE

☐ INCOMING

☐ OUTGOING

ROUTING

NAME/SYMBOL

INT

Location of Visit/Conference:

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU

Bob O'Connor

ORGANIZATION (Office, dept., bureau, etc.)

United Hosp Center

TELEPHONE NO.

304-624-2995

SUBJECT

control 51210 Amend Lic # 47-01458-01
to renovate Nuclear Med Area.

SUMMARY

Bob Asked for verbal ok to move Camera. Sep 22, 1986

Cut door ways Sep 29, 1986

H.P. Services do surveys # 25 and 26
Sep 86.

OK on the above but

Told him, we will need more specific
info on Ventillation system before
we can approve Xenon usage in
the renovated area.

Earl Wright

ACTION REQUIRED

Refer John Potter

Any problem with this. → No - 9/17

NAME OF PERSON DOCUMENTING CONVERSATION

Earl Wright

SIGNATURE

Earl Wright

DATE

9/16/86

ACTION TAKEN

SIGNATURE

TITLE

DATE

CONVERSATION RECORD

TIME

1152

DATE

9/16/86

TYPE

☐ VISIT

☐ CONFERENCE

☒ TELEPHONE

☐ INCOMING

☐ OUTGOING

ROUTING

NAME/SYMBOL INT

Location of Visit/Conference:

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU

Bob O'Connor

ORGANIZATION (Office, dept., bureau, etc.)

United Hosp Center

TELEPHONE NO.

304-624-2995

SUBJECT

Contract 51210 Amend Lic # 47-01458-01
to renovate Nuclear Med Area

SUMMARY

Bob Asked for verbal OK to move Camera. Sep 22, 1986

Cut door ways Sep 29, 1986

HP Services do surveys Sep 25 and 26
Sep 86.

OK on the drive but

Told Jim, we will need more specific info on Ventillation system before we can approve Xenon usage in the renovated area.

Earl Wright

Refer John Potter

ACTION REQUIRED

Any problem with this

NAME OF PERSON DOCUMENTING CONVERSATION

Earl Wright

SIGNATURE

Earl Wright

DATE

9/16/86

ACTION TAKEN

SIGNATURE

TITLE

DATE