



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
631 PARK AVENUE
KING OF PRUSSIA, PENNSYLVANIA 19406

11 DEC 1986

National Medical Services, Inc.
ATTN: Paul Delaney
2300 Stratford Avenue
P.O. Box 433A
Jillow Grove, Pennsylvania 19090

Docket No. 030-13188
License No. 37-13923-02
Control No. 106451

SUBJECT: LICENSE RENEWAL APPLICATION

Gentlemen:

This is to acknowledge receipt of your application for renewal of the material(s) license identified above. Your application is deemed timely filed, and accordingly, the license will not expire until final action has been taken by this office.

Any correspondence regarding the renewal application should reference the control number specified and your license number.

Sincerely,

Original Signed By:
Doris J. Foster

John E. Glenn, Ph.D., Chief
Nuclear Materials Safety Section B
Division of Radiation Safety and
Safeguards

8706040306 970212
REG1 LIC30
37-13923-02 PDR

"OFFICIAL RECORD COPY"

ML10

CAUTION
RADIOACTIVE MATERIAL

NICKEL 63
16.5 MILLICURIES

THE RECEIPT, POSSESSION, USE, AND TRANSFER OF THE NICKEL 63 DETECTOR CONTAINED IN THIS INSTRUMENT IS SUBJECT TO A GENERAL LICENSE OR THE EQUIVALENT AND REGULATIONS OF THE U. S. NUCLEAR REGULATORY COMMISSION OR A STATE WITH WHICH THE U. S. NUCLEAR REGULATORY COMMISSION HAS ENTERED INTO AN AGREEMENT FOR THE EXERCISE OF REGULATORY AUTHORITY. THIS LABEL SHALL BE MAINTAINED ON THE DEVICE IN A LEGIBLE CONDITION. REMOVAL OF THIS LABEL IS PROHIBITED.

THIS DEVICE SHALL BE LEAK TESTED FOR LEAKAGE OF RADIATION AT INTERVALS NOT TO EXCEED THIRTY SIX (36) MONTHS. LEAK TESTS MUST BE MADE AND EVALUATED BY PERSONS HOLDING SPECIFIC LICENSES ISSUED BY THE U.S. REGULATORY COMMISSION OR AN AGREEMENT STATE.

THIS INSTRUMENT MUST NOT BE ABANDONED OR SCRAPPED UNLESS THE NICKEL 63 DETECTOR HAS BEEN REMOVED AND RETURNED TO THE MANUFACTURER.

CAUTION
RADIOACTIVE MATERIAL

MANUFACTURER: TRACOR INC. AUSTIN, TEXAS
TX LIC. 6-1186 G.L.

18791