

JUL 22 1986

Memorial Medical Center, Inc.  
ATTN: Marjorie Kacvinsky  
Radiology Director  
1615 Maple Lane  
Ashland, WI 54806

Dear Ms. Kacvinsky:

Enclosed is Amendment No. 07 renewing your NRC License No. 48-16593-01 in accordance with your request.

In accordance with our telephone conversation on June 23, 1986, License Condition No. 13. has been added to address the availability of a high-level survey meter capable of reading up to 1 Roentgen per hour in order to measure radiation dose rates that may exist in the vicinity of technetium-99m generators.

Please note that we have modified Condition No. 14. of your license to incorporate the Model ALARA Program as contained in Appendix O of Regulatory Guide 10.8, October, 1980 (enclosed). If you wish to develop alternative procedures that are at least equivalent to the ones stated in the model program, your license can be amended accordingly.

If you have any questions or require clarification on any of the information stated above, you may contact us at (312) 790-5625.

Sincerely,

Kathy J. Grader,  
Materials Licensing Section

Enclosures:

1. Amendment No. 07
2. Regulatory Guide 10.8
3. Pages 10.8-25 and 10.8-26  
of old Regulatory Guide 10.8

8610020198 860722  
REG3 LIC30  
48-16593-01 PDR

RIII

KJG  
Graden/bs  
7/ 15/86