

EXHIBIT A

FORM NRC-313M (B-7B) 10 CFR 35	U.S. NUCLEAR REGULATORY COMMISSION APPLICATION FOR MATERIALS LICENSE – MEDICAL	Approved: GAO R0557
--------------------------------------	--	------------------------

INSTRUCTIONS - Complete Items 1 through 26 if this is an initial application or an application for renewal of a license. Use supplemental sheets where necessary. Item 26 must be completed on all applications and signed. Retain one copy. Submit original and one copy of entire application to: Director, Office of Nuclear Materials Safety and Safeguards, U.S. Nuclear Regulatory Commission, Washington, D.C. 20555. Upon approval of this application, the applicant will receive a Materials License. An NRC Materials License is issued in accordance with the general requirements contained in Title 10, Code of Federal Regulations, Part 30, and the Licensee is subject to Title 10, Code of Federal Regulations, Parts 19, 20 and 35 and the license fee provision of Title 10, Code of Federal Regulations, Part 170. The license fee category should be stated in Item 26 and the appropriate fee enclosed.

1.a. NAME AND MAILING ADDRESS OF APPLICANT (institution, firm, clinic, physician, etc.) INCLUDE ZIP CODE Columbus-Cuneo-Cabrini Medical center 2520N. Lakeview Ave Chicago, Illinois 60614 TELEPHONE NO.: AREA CODE() _____	1.b. STREET ADDRESS(ES) AT WHICH RADIOACTIVE MATERIAL WILL BE USED (If different from 1.a.) INCLUDE ZIP CODE Same as 1a 2. Frank Cuneo Hospital 750 W. Montrose, Chicago, 60613 3. St. Francis Cabrini Hospital 811 S. Lytle St. Chicago 60607
--	---

2. PERSON TO CONTACT REGARDING THIS APPLICATION Bashir Pothiawala TELEPHONE NO.: AREA CODE (312) 883 6519	3. THIS IS AN APPLICATION FOR: (Check appropriate item) a. <input type="checkbox"/> NEW LICENSE b. <input type="checkbox"/> AMENDMENT TO LICENSE NO. <u>12-12074-01</u> c. <input type="checkbox"/> RENEWAL OF LICENSE NO. _____
---	---

4. INDIVIDUAL USERS (Name individuals who will use or directly supervise use of radioactive material. Complete Supplements A and B for each individual.) Attached	5. RADIATION SAFETY OFFICER (RSO) (Name of person designated as radiation safety officer. If other than individual user, complete resume of training and experience as in Supplement A.) Bashir Pothiawala
--	---

6.a. RADIOACTIVE MATERIAL FOR MEDICAL USE			ADDITIONAL ITEMS:	MARK ITEMS DESIRED	MAXIMUM POSSESSION LIMITS
RADIOACTIVE MATERIAL LISTED IN:	ITEMS DESIRED "X"	MAXIMUM POSSESSION LIMITS (In millicuries)		"X"	(In millicuries)
10 CFR 31.11 FOR IN VITRO STUDIES	X	3	IODINE-131 AS IODIDE FOR TREATMENT OF HYPERTHYROIDISM	X	90
10 CFR 35.100, SCHEDULE A, GROUP I	X	AS NEEDED	PHOSPHORUS-32 AS SOLUBLE PHOSPHATE FOR TREATMENT OF POLYCYTHEMIA VERA, LEUKEMIA AND BONE METASTASES	X	30
10 CFR 35.100, SCHEDULE A, GROUP II	X	AS NEEDED	PHOSPHORUS-32 AS COLLOIDAL CHROMIC PHOSPHATE FOR INTRACAVITARY TREATMENT OF MALIGNANT EFFUSIONS.	X	60
10 CFR 35.100, SCHEDULE A, GROUP III	X	700	GOLD-198 AS COLLOID FOR INTRACAVITARY TREATMENT OF MALIGNANT EFFUSIONS.	X	200
10 CFR 35.100, SCHEDULE A, GROUP IV	X	AS NEEDED	IODINE-131 AS IODIDE FOR TREATMENT OF THYROID CARCINOMA	X	400
10 CFR 35.100, SCHEDULE A, GROUP V	X	AS NEEDED	XENON-133 AS GAS OR GAS IN SALINE FOR BLOOD FLOW STUDIES AND PULMONARY FUNCTION STUDIES.	X	400
10 CFR 35.100, SCHEDULE A, GROUP VI	X	700			

6.b. RADIOACTIVE MATERIAL FOR USES NOT LISTED IN ITEM 6.a. (Sealed sources up to 3 mCi used for calibration and reference standards are authorized under Section 35.14(d), 10 CFR Part 35, and NEED NOT BE LISTED.)			
ELEMENT AND MASS NUMBER	CHEMICAL AND/OR PHYSICAL FORM	MAXIMUM NUMBER OF MILLICURIES OF EACH FORM	DESCRIBE PURPOSE OF USE
8612150068 REG3 LIC30 12-12074-01	861017 PDR		

INFORMATION REQUIRED FOR ITEMS 7 THROUGH 23

For Items 7 through 23, check the appropriate box(es) and submit a detailed description of all the requested information. Begin each item on a separate sheet. Identify the item number and the date of the application in the lower right corner of each page. If you indicate that an appendix to the medical licensing guide will be followed, do not submit the pages, but specify the revision number and date of the referenced guide: Regulatory Guide 10.8, Rev. _____ Date: _____

7. MEDICAL ISOTOPES COMMITTEE		15. GENERAL RULES FOR THE SAFE USE OF RADIOACTIVE MATERIAL (Check One)	
<input checked="" type="checkbox"/>	Names and Specialties Attached; and	<input checked="" type="checkbox"/>	Appendix G Rules Followed; or
<input checked="" type="checkbox"/>	Duties as in Appendix B; or _____ (Check One)		Equivalent Rules Attached
	Equivalent Duties Attached	16. EMERGENCY PROCEDURES (Check One)	
8. TRAINING AND EXPERIENCE		<input checked="" type="checkbox"/>	Appendix H Procedures Followed; or
<input checked="" type="checkbox"/>	Supplements A & B Attached for Each Individual User; and		Equivalent Procedures Attached
<input checked="" type="checkbox"/>	Supplement A Attached for RSO.	17. AREA SURVEY PROCEDURES (Check One)	
9. INSTRUMENTATION (Check One)		<input checked="" type="checkbox"/>	Appendix I Procedures Followed; or
<input checked="" type="checkbox"/>	Appendix C Form Attached; or		Equivalent Procedures Attached
<input checked="" type="checkbox"/>	List by Name and Model Number	18. WASTE DISPOSAL (Check One)	
10. CALIBRATION OF INSTRUMENTS			Appendix J Form Attached; or
	Appendix D Procedures Followed for Survey Instruments; or _____ (Check One)	<input checked="" type="checkbox"/>	Equivalent Information Attached
<input checked="" type="checkbox"/>	Equivalent Procedures Attached; and	19. THERAPEUTIC USE OF RADIOPHARMACEUTICALS (Check One)	
<input checked="" type="checkbox"/>	Appendix D Procedures Followed for Dose Calibrator; or _____ (Check One)	<input checked="" type="checkbox"/>	Appendix K Procedures Followed; or
	Equivalent Procedures Attached		Equivalent Procedures Attached
11. FACILITIES AND EQUIPMENT		20. THERAPEUTIC USE OF SEALED SOURCES	
<input checked="" type="checkbox"/>	Description and Diagram Attached	<input checked="" type="checkbox"/>	Detailed Information Attached; and
12. PERSONNEL TRAINING PROGRAM		<input checked="" type="checkbox"/>	Appendix L Procedures Followed; or _____ (Check One)
<input checked="" type="checkbox"/>	Description of Training Attached		Equivalent Procedures Attached
13. PROCEDURES FOR ORDERING AND RECEIVING RADIOACTIVE MATERIAL		21. PROCEDURES AND PRECAUTIONS FOR USE OF RADIOACTIVE GASES (e.g., Xenon - 133)	
<input checked="" type="checkbox"/>	Detailed Information Attached	<input checked="" type="checkbox"/>	Detailed Information Attached
14. PROCEDURES FOR SAFELY OPENING PACKAGES CONTAINING RADIOACTIVE MATERIALS (Check One)		22. PROCEDURES AND PRECAUTIONS FOR USE OF RADIOACTIVE MATERIAL IN ANIMALS	
<input checked="" type="checkbox"/>	Appendix F Procedures Followed; or		Detailed Information Attached
	Equivalent Procedures Attached	23. PROCEDURES AND PRECAUTIONS FOR USE OF RADIOACTIVE MATERIAL SPECIFIED IN ITEM 8.b	
			Detailed Information Attached

24. PERSONNEL MONITORING DEVICES

TYPE <i>(Check appropriate box)</i>		SUPPLIER	EXCHANGE FREQUENCY
a. WHOLE BODY	<input checked="" type="checkbox"/> FILM	Siemens Gammasonic, Inc.	Monthly
	<input type="checkbox"/> TLD		
	<input type="checkbox"/> OTHER <i>(Specify)</i>		
b. FINGER	<input type="checkbox"/> FILM		
	<input checked="" type="checkbox"/> TLD	Siemens Gammasonic, Inc.	Monthly
	<input type="checkbox"/> OTHER <i>(Specify)</i>		
c. WRIST	<input type="checkbox"/> FILM		
	<input type="checkbox"/> TLD		
	<input type="checkbox"/> OTHER <i>(Specify)</i>		

d. OTHER *(Specify)*


25. FOR PRIVATE PRACTICE APPLICANTS ONLY

a. HOSPITAL AGREEING TO ACCEPT PATIENTS CONTAINING RADIOACTIVE MATERIAL			
NAME OF HOSPITAL		b. ATTACH A COPY OF THE AGREEMENT LETTER SIGNED BY THE HOSPITAL ADMINISTRATOR. c. WHEN REQUESTING THERAPY PROCEDURES, ATTACH A COPY OF RADIATION SAFETY PRECAUTIONS TO BE TAKEN AND LIST AVAILABLE RADIATION DETECTION INSTRUMENTS.	
MAILING ADDRESS			
CITY	STATE ZIP CODE		

26. CERTIFICATE

(This item must be completed by applicant)

The applicant and any official executing this certificate on behalf of the applicant named in Item 1a certify that this application is prepared in conformity with Title 10, Code of Federal Regulations, Parts 30 and 35, and that all information contained herein, including any supplements attached hereto, is true and correct to the best of our knowledge and belief.

a. LICENSE FEE REQUIRED <i>(See Section 170.31, 10 CFR 170)</i>	b. APPLICANT OR CERTIFYING OFFICIAL <i>(Signature)</i> 
	(1) NAME <i>(Type of Print)</i> Bashir Pothiwala
(1) LICENSE FEE CATEGORY	(2) TITLE Radiation Safety Officer
(2) LICENSE FEE ENCLOSED: \$ _____	c. DATE July 18, 1884

PRIVACY ACT STATEMENT

Pursuant to 5 U.S.C. 552a(e)(3), enacted into law by section 3 of the Privacy Act of 1974 (Public Law 93-579), the following statement is furnished to individuals who supply information to the Nuclear Regulatory Commission on Form NRC-313M. This information is maintained in a system of records designated as NRC-3 and described at 40 Federal Register 45334 (October 1, 1975).

1. **AUTHORITY** Sections 81 and 161(b) of the Atomic Energy Act of 1954, as amended (42 U.S.C. 2111 and 2201(b)).
2. **PRINCIPAL PURPOSE(S)** The information is evaluated by the NRC staff pursuant to the criteria set forth in 10 CFR Parts 30-36 to determine whether the application meets the requirements of the Atomic Energy Act of 1954, as amended, and the Commission's regulations, for the issuance of a radioactive material license or amendment thereof.
3. **ROUTINE USES** The information may be used: (a) to provide records to State health departments for their information and use; and (b) to provide information to Federal, State, and local health officials and other persons in the event of incident or exposure, for their information, investigation, and protection of the public health and safety. The information may also be disclosed to appropriate Federal, State, and local agencies in the event that the information indicates a violation or potential violation of law and in the course of an administrative or judicial proceeding. In addition, this information may be transferred to an appropriate Federal, State, or local agency to the extent relevant and necessary for a NRC decision or to an appropriate Federal agency to the extent relevant and necessary for that agency's decision about you. A copy of the license issued will routinely be placed in the NRC's Public Document Room, 1717 H Street, N.W., Washington, D.C.
4. **WHETHER DISCLOSURE IS MANDATORY OR VOLUNTARY AND EFFECT ON INDIVIDUAL OF NOT PROVIDING INFORMATION** Disclosure of the requested information is voluntary. If the requested information is not furnished, however, the application for radioactive material license, or amendment thereof, will not be processed.
5. **SYSTEM MANAGER(S) AND ADDRESS** Director, Division of Fuel Cycle and Material Safety, Office of Nuclear Material Safety and Safeguards, U.S. Nuclear Regulatory Commission, Washington, D.C. 20555.

July 18, 1984

Control # 76350

Please refer to our application dated February 15, 1984
for following Items.

- Item 4
- Item 5
- Item 7
- Item 8
- Item 10
- Item 11
- Item 12
- Item 13
- Item 14
- Item 15
- Item 16
- Item 17
- Item 18
- Item 19
- Item 20
- Item 21

Control No. 77196

APPENDIX C
INSTRUMENTATION

Cabrin Hospital

1. Survey meters

- a. Manufacturer's name: Victoreen
 Manufacturer's model number: 740B
 Number of instruments available: 1
 Minimum range: 1 mR/hr to 25 mR/hr
 Maximum range: 100 mR/hr to 2500 mR/hr
- b. Manufacturer's name: Victoreen
 Manufacturer's model number: 6A
 Number of instruments available: 1
 Minimum range: .01 mR/hr to .5 mR/hr
 Maximum range: 1 mR/hr to 50 mR/hr



2. Dose calibrator

- Manufacturer's name: Squibb Inc.
 Manufacturer's model number: CRC-16
 Number of instruments available: 1

3. Instruments used for diagnostic procedures

Type of Instrument	Manufacturer's Name	Model No.
Radioisotope dual Probe and well counter	Picker	2852
Gamma Camera	Searle/Nuclear Chicago	30-799
Gamma camera Pulmonex	" Atomic Products	LFOV 130-500
Other (e.g., liquid scintillation counter, area monitor, velometer)		
Atomic Product Primalert- 50 Continuous Geiger monitor		

Cabrini Nuclear Medicine Area -



Hot Lab

- a. Dose calibrator
- b. "L" shield
- c. lead bricks, storage
- d. lead bricks, storage
- e. sink

Scale: 1" = 8'

= LOCATION OF EXHAUST

COLUMBUS-CUNEO-CABRINI MEDICAL CENTER

Cabrini Hospital

Item 17; Page 5

Weekly Survey Check

Date: _____

Performer: _____

Well Setting: HV = 5.09
 LL = 0.50
 UL = 0.5
 Window - OUT


Well Constancy Check: Source ¹³⁷Cs

Readings: _____
 (Ct/min): _____

Operational Check Courses: _____ mR/hr
 _____ mR/hr

Bkgd: _____ ct/min
 MDA: _____ ct/min
 Total: _____ ct/min

G. M. Survey			Wipe Tests - ct/min				
No.	Location	mR/hr	No.	Location	Reading	Bkgd	Net
1	Pt waiting area		3-A	Floor			
2	Ultrasound room		5-A	Shld Disp Area			
3	West Camera Room		6-A	Counter			
4	Floor		7-A	"L" Block			
5	Shielded Disp Area		8-A	Sink			
6	Counter		9-A	Floor			
7	"L" Block		10-A	Floor			
8	Sink						
9	Floor						
10	East Camera Room						



ACTION:

Control No. 77196

XENON-133 PROGRAM

a. Quantities to be used:

1. Maximum number of patients=20 patients per week at an average activity of 10mCi per patient.
2. Desired possession limits= 200 mCi

b. Use and Storage Areas:

1. All Xenon-133 gas is courier carried and delivered to us in prepackaged dose vials from a supplier. These calibrated vials are shielded in lead tubes (8 HVL's) and stored in our radiopharmacy until needed. The Xenon-133 will be used in our east camera room which has two gamma cameras which can be used for the study. (see diagrams-item 21:)
2. All vents were measured using an Alnor Anemometer (serial no. ECI-26). Three measurements were taken and averaged over the face of the vent. All rooms in the Nuclear Medicine Department have a 0% recirculation rate and this air is exhausted directly to the roof which is exhausted far from windows and intake vents. The east camera room has a measured air exhaust rate of 600 (ft³/min) cfm and the radiopharmacy has 315 cfm. Our system will maintain Xenon-133 concentrations well below maximum permissible levels.
3. Air flow rates will be monitored and quantitatively checked on an annual schedule. If unusual Xenon-133 background are detected during routine surveys, then the cause will be determined.

c. Procedures for Routine Use:

- 1) Xenon-133 is purchased in the prepackaged shielded vials which are loaded into a New England Nuclear NRP-186 Calidose dispenser. The radioactive gas is injected into an Atomic Products Corporation (see enclosed information Itam 21: pages 7-11) Pulmonex Xenon System (Model no. 130-500) with a charcoal filter used for trapping most of the Xenon-133 gas. The filter exhaust is being monitored semi-annually to determine at what time the filter should be replaced. The semi-annual check of the system will also evaluate if any leaks are occurring within the system itself.

Monitoring of the charcoal filter is being measured by comparing a known activity and volume of Xenon-133 to that collected from the filter exhaust and other places as needed. The collection bag-valve system collects air samples which are then evaluated by placing the collection system on the uncollimated face of our mobile gamma camera. Background levels are subtracted to determine the net activity per millimeter of air. When the Charcoal filter approaches the maximum permissible exhaust concentration (1×10^{-5} mCi/ml), it will then be replaced. Spent filters will be sealed and stored for decay and all filters are sufficiently lead shielded for normal handling.

- 2) Xenon-133 Dispensing System:

Pulmonex Xenon Delivery System-Atomic Products Corporation;
Model no. 130-500.

- 3) In addition to the surveys listed in (c.1) above, most removable joints have clamps attached to prevent leakage. Every reasonable attempt will be made through observation and instruction to the patient to ensure that the gas delivery mask is making an airtight seal before commencing with the Xenon-133 procedure.

d. Emergency Procedures:

See item 21: page 12 for emergency procedures of an accidental release of Xenon-133.

e) Air concentration of Xenon-133 in restricted area:

a) Maximum amount of activity to be used per week

$$A = (10 \text{ mCi/pt}) (10 \text{ pt/wk}) (10^3 \text{ uCi/mCi}) = 1 \times 10^5 \text{ uCi/wk}$$

b) Assume the loss factor during use and storage is

$$f = 20\%$$

c) Minimum required exhaust rate (v) in the area to maintain the Xe-133 concentration below maximum permissible concentration 1×10^{-5} uCi/ml is

$$V = \frac{Axf}{\text{MPC}} = \frac{1 \times 10^5 \times .2}{1 \times 10^{-5}} = 2 \times 10^9 \text{ ml/wk}$$

Conversion factor 1 cfm = 40 hr/wk \times 1.7×10^6 ml/hr-cfm

$$= 6.8 \times 10^7 \text{ ml/wk}$$

$$\text{Therefore } V = \frac{2 \times 10^9}{40 \times 1.7 \times 10^6} = 29.4 \text{ cfm}$$

The room is purged at a measured rate of 600 cfm in east camera room and total air flow exhaust rate is 1715 cfm which is well above the required value.

f) Air concentration of Xe-133 in unrestricted area:

1) As mentioned above, the only Xe-133 gas expelled to the unrestricted area is the residue gas of xenon filter through the exhaust vent into the atmosphere. this amount is expected to be minimum.

2) Assume the maximum factor (f) for Xe-133 released during the procedure and storage is

$$A = 20\%$$

$$\begin{aligned} \text{Amount released during a year (A)} &= 2 \times 10^4 \text{ uCi/wk} \times 52 \text{ wk/yr} \\ &= 1.04 \times 10^6 \text{ uCi/yr} \end{aligned}$$

$$\text{MPC in unrestricted area} = 3 \times 10^{-7} \text{ uCi/ml}$$

$$\begin{aligned} \text{Required minimum ventillation rate (V)} &= \frac{1.04 \times 10^6 \text{ uCi/yr}}{3 \times 10^{-7} \text{ uCi/ml}} \\ &= 3.46 \times 10^{12} \text{ ml/yr} \end{aligned}$$

$$\text{Conversion factor 1 cfm} = \frac{1.484 \times 10^{10} \text{ ml/yr}}{3.46 \times 10^{12} \text{ ml/yr}}$$

$$\text{Therefore } V = \frac{3.46 \times 10^{12} \text{ ml/yr}}{1.484 \times 10^{10} \text{ ml/yr-cfm}} = 233 \text{ cfm}$$

The total flow rate of exhaust system of the department is 1715 cfm which is way above the required flow rate of 233 cfm to maintain MPC of 3×10^{-7} uCi/ml.

$$V = 1715 \text{ cfm} \times 1.484 \times 10^{10} \text{ ml/yr} = 2.545 \times 10^{13} \text{ ml/yr}$$

$$C = \frac{1.04 \times 10^6 \text{ uCi/Yr}}{2.545 \times 10^{13} \text{ ml/yr}} = 4.08 \times 10^{-8} \text{ uCi/ml}$$

4.08×10^{-8} uCi/ml is well below the concentration limit of 3×10^{-7} uCi/ml

Calculation of Emergency Procedure for Xenon - 133

Approximate room size = 15' x 15' x 9' = 2025 ft³

$$1 \text{ ft} = 2.832 \times 10^4 \text{ ml}$$

Total ml in the room = $2.832 \times 10^4 \times 2025 = 5.7 \times 10^7 \text{ ml}$

If one vial accidentally ruptures or breaks, this would release about 10 mCi

$$\begin{aligned} \text{Air concentration (Xe-133)} &= \frac{10 \text{ mCi} \times 10^3 \text{ uCi/mCi}}{5.7 \times 10^7 \text{ ml}} \\ &= 1.75 \times 10^{-4} \text{ uCi/ml} \end{aligned}$$

Room purge rate of 600 cfm:

$$\text{For one room air turnover} = \frac{2025 \text{ ft}^3}{600 \text{ ft}^3/\text{min}} = \underline{3.4 \text{ min}}$$

Using the assumption that 1 complete room purge removes $\frac{1}{2}$ of the activity this would mean that after 4 complete room purges, one would expect 1/16 of the initial concentration to be present. After this period of time, the concentration should be well below 1×10^{-5} uCi/ml level.

$$\text{Evacuation time: } 4 \times 3.4 \text{ min} = 13.6 \text{ min}$$

the background count will also be performed using the detection equipment in the room.



COLUMBUS - CUNED - CABRINI
MEDICAL CENTER

DETAIL PROCEDURES

SUBJECT/DEPARTMENT:

NUCLEAR MEDICINE

COPY:

DATE ISSUED:

REVISION DATE:

PG

OF

APPROVED BY: B. Pothiwala, Medical Physicist

TITLE: EMERGENCY PROCEDURES FOR AN
ACCIDENTAL RELEASE OF XENON-133

NO:

RESPONSIBILITY

ACTION

If an accidental release of Xenon-133 occurs in East Camera Room, then the following steps should be taken:

- 1) Clear gamma camera room of all patients and personnel
- 2) Close all entrances, sealing off area.
- 3) Call the radiation Safety Office (ext. 6517 - 7350)
- 4) The room should be safe to enter, complying with Section 20.103, 10 CFR. Part 20 (1×10^{-5} $\mu\text{Ci/ml}$) after twenty minutes if only one 10 mCi vial ruptures. If more than 10 mCi are released, contact the Radiation Safety Office. A background level set on a Xenon-133 detection technique can be performed to qualitatively evaluate Xenon-133 concentrations.
- 5) If a release of Xenon-133 occurs in any other area, close off access and doors to that area and contact the Radiation Safety Office.

Control No. 77196