

APPLICATION FOR MATERIAL LICENSE

U.S. NUCLEAR REGULATORY COMMISSION
APPROVED BY OMB
3150-0120
Expires: 5-31-87

INSTRUCTIONS: SEE THE APPROPRIATE LICENSE APPLICATION GUIDE FOR DETAILED INSTRUCTIONS FOR COMPLETING APPLICATION. SEND TWO COPIES OF THE ENTIRE COMPLETED APPLICATION TO THE NRC OFFICE SPECIFIED BELOW.

FEDERAL AGENCIES FILE APPLICATIONS WITH:

U.S. NUCLEAR REGULATORY COMMISSION
DIVISION OF FUEL CYCLE AND MATERIAL SAFETY, NMSS
WASHINGTON, DC 20555

ALL OTHER PERSONS FILE APPLICATIONS AS FOLLOWS, IF YOU ARE LOCATED IN:

CONNECTICUT, DELAWARE, DISTRICT OF COLUMBIA, MAINE, MARYLAND, MASSACHUSETTS, NEW JERSEY, NEW YORK, PENNSYLVANIA, RHODE ISLAND, OR VERMONT, SEND APPLICATIONS TO:

U.S. NUCLEAR REGULATORY COMMISSION, REGION I
NUCLEAR MATERIAL SECTION B
631 PARK AVENUE
KING OF PRUSSIA, PA 19406

ALABAMA, FLORIDA, GEORGIA, KENTUCKY, MISSISSIPPI, NORTH CAROLINA, PUERTO RICO, SOUTH CAROLINA, TENNESSEE, VIRGINIA, VIRGIN ISLANDS, OR WEST VIRGINIA, SEND APPLICATIONS TO:

U.S. NUCLEAR REGULATORY COMMISSION, REGION II
MATERIAL RADIATION PROTECTION SECTION
101 MARIETTA STREET, SUITE 2900
ATLANTA, GA 30323

IF YOU ARE LOCATED IN:

ILLINOIS, INDIANA, IOWA, MICHIGAN, MINNESOTA, MISSOURI, OHIO, OR WISCONSIN, SEND APPLICATIONS TO:

U.S. NUCLEAR REGULATORY COMMISSION, REGION III
MATERIALS LICENSING SECTION
799 ROOSEVELT ROAD
GLEN ELLYN, IL 60137

ARKANSAS, COLORADO, IDAHO, KANSAS, LOUISIANA, MONTANA, NEBRASKA, NEW MEXICO, NORTH DAKOTA, OKLAHOMA, SOUTH DAKOTA, TEXAS, UTAH, OR WYOMING, SEND APPLICATIONS TO:

U.S. NUCLEAR REGULATORY COMMISSION, REGION IV
MATERIAL RADIATION PROTECTION SECTION
611 RYAN PLAZA DRIVE, SUITE 1000
ARLINGTON, TX 76011

ALASKA, ARIZONA, CALIFORNIA, HAWAII, NEVADA, OREGON, WASHINGTON, AND U.S. TERRITORIES AND POSSESSIONS IN THE PACIFIC, SEND APPLICATIONS TO:

U.S. NUCLEAR REGULATORY COMMISSION, REGION V
MATERIAL RADIATION PROTECTION SECTION
1450 MARIA LANE, SUITE 210
WALNUT CREEK, CA 94596

PERSONS LOCATED IN AGREEMENT STATES SEND APPLICATIONS TO THE U.S. NUCLEAR REGULATORY COMMISSION ONLY IF THEY WISH TO POSSESS AND USE LICENSED MATERIAL IN STATES SUBJECT TO U.S. NUCLEAR REGULATORY COMMISSION JURISDICTION.

1. THIS IS AN APPLICATION FOR (Check appropriate item)

- ☒ A. NEW LICENSE
☐ B. AMENDMENT TO LICENSE NUMBER _____
☐ C. RENEWAL OF LICENSE NUMBER _____

2. NAME AND MAILING ADDRESS OF APPLICANT (Include Zip Code)

Wisconsin Wires, Inc.
6480 West College Avenue
Appleton, WI 54915

3. ADDRESS(ES) WHERE LICENSED MATERIAL WILL BE USED OR POSSESSED.

Wisconsin Wires, Inc.
6480 West College Avenue
Appleton, WI 54915

Two miles west of Highway 41 on
College Avenue

4. NAME OF PERSON TO BE CONTACTED ABOUT THIS APPLICATION

Richard Watry

8610160138 860904
REG3 LIC30
48-24750-01 PDR

TELEPHONE NUMBER
(414) 734-2607

SUBMIT ITEMS 5 THROUGH 11 ON 8 1/2 x 11" PAPER. THE TYPE AND SCOPE OF INFORMATION TO BE PROVIDED IS DESCRIBED IN THE LICENSE APPLICATION GUIDE.

5. RADIOACTIVE MATERIAL

a. Element and mass number, b. chemical and/or physical form, and c. maximum amount which will be possessed at any one time.

6. PURPOSE(S) FOR WHICH LICENSED MATERIAL WILL BE USED.

7. INDIVIDUAL(S) RESPONSIBLE FOR RADIATION SAFETY PROGRAM AND THEIR TRAINING AND EXPERIENCE.

8. TRAINING FOR INDIVIDUALS WORKING IN OR FREQUENTING RESTRICTED AREAS.

9. FACILITIES AND EQUIPMENT.

10. RADIATION SAFETY PROGRAM.

11. WASTE MANAGEMENT.

12. LICENSEE FEES (See 10 CFR 170 and Section 170.31)

FEE CATEGORY 3P AMOUNT ENCLOSED \$ 230.00

13. CERTIFICATION. (Must be completed by applicant) THE APPLICANT UNDERSTANDS THAT ALL STATEMENTS AND REPRESENTATIONS MADE IN THIS APPLICATION ARE BINDING UPON THE APPLICANT.

THE APPLICANT AND ANY OFFICIAL EXECUTING THIS CERTIFICATION ON BEHALF OF THE APPLICANT, NAMED IN ITEM 2, CERTIFY THAT THIS APPLICATION IS PREPARED IN CONFORMITY WITH TITLE 10, CODE OF FEDERAL REGULATIONS, PARTS 30, 32, 33, 34, 35, AND 40 AND THAT ALL INFORMATION CONTAINED HEREIN, IS TRUE AND CORRECT TO THE BEST OF THEIR KNOWLEDGE AND BELIEF.

WARNING: 18 U.S.C. SECTION 1001 ACT OF JUNE 25, 1948, 62 STAT. 749 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

SIGNATURE—CERTIFYING OFFICER

TYPED/PRINTED NAME

TITLE

DATE

R. L. McDonald

R. L. McDonald

Vice President

6/25/86

14. VOLUNTARY ECONOMIC DATA

ANNUAL RECEIPTS	
<\$250K	\$1M-3.5M
\$250K-500K	\$3.5M-7M
\$500K-750K	\$7M-10M
\$750K-1M	>\$10M

b. NUMBER OF EMPLOYEES (Total for entire facility excluding outside contractors)

c. NUMBER OF BEDS

d. WOULD YOU BE WILLING TO FURNISH COST INFORMATION (Dollar and/or staff hours) ON THE ECONOMIC IMPACT OF CURRENT NRC REGULATIONS OR ANY FUTURE PROPOSED NRC REGULATIONS THAT MAY AFFECT YOU? (NRC regulations permit it to protect confidential commercial or financial—proprietary—information furnished to the agency in confidence)

☐ YES

☐ NO

FOR NRC USE ONLY

TYPE OF FEE

FEE LOG

FEE CATEGORY

COMMENTS

RECEIVED

APPROVED BY

AMOUNT RECEIVED

CHECK NUMBER

JUN 30 1986

DATE

REGION III

PRIVACY ACT STATEMENT

Pursuant to 5 U.S.C. 552a(e)(3), enacted into law by section 3 of the Privacy Act of 1974 (Public Law 93-579), the following statement is furnished to individuals who supply information to the Nuclear Regulatory Commission on NRC Form 313. This information is maintained in a system of records designated as NRC-3 and described at 40 Federal Register 45334 (October 1, 1975).

1. **AUTHORITY:** Sections 81 and 161(b) of the Atomic Energy Act of 1954, as amended (42 U.S.C. 2111 and 2201(b)).
2. **PRINCIPAL PURPOSE(S):** The information is evaluated by the NRC staff pursuant to the criteria set forth in 10 CFR Parts 30, 32, 33, 34, 35 and 40 to determine whether the application meets the requirements of the Atomic Energy Act of 1954, as amended, and the Commission's regulations, for the issuance of a radioactive material license or amendment thereof.
3. **ROUTINE USES:** The information may be (a) provided to State health departments for their information and use; and (b) provided to Federal, State, and local health officials and other persons in the event of incident or exposure, for their information, investigation, and protection of the public health and safety. The information may also be disclosed to appropriate Federal, State, and local agencies in the event that the information indicates a violation or potential violation of law and in the course of an administrative or judicial proceeding. In addition, this information may be transferred to an appropriate Federal, State, or local agency to the extent relevant and necessary for an NRC decision or to an appropriate Federal agency to the extent relevant and necessary for that agency's decision about you.
4. **WHETHER DISCLOSURE IS MANDATORY OR VOLUNTARY AND EFFECT ON INDIVIDUAL OF NOT PROVIDING INFORMATION:** Disclosure of the requested information is voluntary. If the requested information is not furnished, however, the application for radioactive material license, or amendment thereof, will not be processed. A request that information be held from public inspection must be in accordance with the provisions of 10 CFR 2.790. Withholding from public inspection shall not affect the right, if any, of persons properly and directly concerned need to inspect the document.
5. **SYSTEM MANAGER(S) AND ADDRESS:** U.S. Nuclear Regulatory Commission
Director, Division of Fuel Cycle and Material Safety
Office of Nuclear Material Safety and Safeguards
Washington, D.C. 20555

Application For Material License
Attachment To NRC Form 313
Wisconsin Wires, Inc.
6480 West College Avenue
Appleton, WI 54915

ITEM NO. 5

- A. ELEMENT AND MASS NUMBER: AMERICIUM 241 (AM-241)
- B. FORM: AMERSHAM AMCPI
- C. MAXIMUM AMOUNT: UP TO 150 MCI
- D. INSTRUMENT NOMENCLATURE: NDC SYSTEMS MODEL 104

Application For Material License
Attachment To NRC Form 313
Wisconsin Wires, Inc.
6480 West College Avenue
Appleton, WI 54915

ITEM NO. 6

The portable gamma backscatter mass gauge manufactured under a State of California specific license 1451-70 and sold under a State of California general license GL 1933-70.

This gauge is used in paper mills to measure drainage profiles under the wire of a Fourdrinier papermaking machine.

Application For Material License
Attachment To NRC Form 313
Wisconsin Wires, Inc.
6480 West College Avenue
Appleton, WI 54915

ITEM NO. 7

Radiation Safety Officer: Richard Watry

The Radiation Safety Officer, Richard Watry, will complete the radiation safety training conducted by NDC Systems' Radiation Safety Officer. Dr. J. Bert Fishman will be holding this training session.

Application For Material License
Attachment To NRC Form 313
Wisconsin Wires, Inc.
6480 West College Avenue
Appleton, WI 54915

ITEM NO. 8

TRAINING PROVIDED TO OTHER USERS

- A. Wisconsin Wires, Inc. will ensure that each employee who will use the gauge will complete NDC's training program prior to operating the unit.
- B. The Radiation Safety Officer will keep a record to document the training of each employee. This record will include an outline of the topics covered in the training and will be on file for a minimum of five years.


Application For Material License
Attachment To NRC Form 313
Wisconsin Wires, Inc.
6480 West College Avenue
Appleton, WI 54915

ITEM NO. 9

FACILITIES & EQUIPMENT

The NDC Huyfiler II will be stored in a locked cabinet when not in use. The Radiation Safety Officer will maintain a key to this cabinet.

When the gauge is transported, by vehicle, it will be locked in the trunk. Unauthorized persons will not have access to this device.



Application For Material License
Attachment To NRC Form 313
Wisconsin Wires, Inc.
6480 West College Avenue
Appleton, WI 54915

ITEM NO. 10

RADIATION SAFETY PROGRAM

The following procedures will be followed in the use of the Model 104.

1. The Radiation Safety Officer (RSO) will have overall responsibility for the gauge. The gauge, when not in use, shall be in the RSO's possession and stored in a secured area. The RSO shall have the only key to the secured area.
2. Requests by customers for surveys or demonstrations using the gauge will be recorded by the RSO who will assign dates and engineering personnel for each field trip.
3. The RSO will make all necessary arrangements for transportation. During transportation by car, the radiation device, in its locked case, shall be securely locked into the car trunk. All applicable State and Federal transportation regulations regarding securing of the unit and identification shall be observed.
4. The RSO will send necessary notification to Agreement States and NRC offices at least three days prior to performing field work in those states, giving estimated dates that the gauge will be at that particular plant.
5. The field engineer performing the survey is not to leave the gauge unattended in unsecure areas. If it is necessary to leave the plant survey area, the radiation device shall be repacked in its case, the case should be locked, and then stored in a location where it can be either kept under lock and key or be under the constant control of a senior plant employee.
6. Prior to assigning the gauge to a field engineer, the RSO will check the gauge for proper operation, and make sure that the shutter is working properly. The RSO will check that detailed operating instructions for use of the gauge are packed in a prominent location with the gauge.

Application For Material License
Attachment To NRC Form 313
Wisconsin Wires, Inc.
6480 West College Avenue
Appleton, WI 54915

7. When the gauge is returned, the above shall be repeated.
8. The RSO will be responsible for ensuring that the six-month leak tests and shutter tests are performed, and that all records be kept on file.
9. At no time will company personnel attempt to perform maintenance or repairs on the radiation device. In case of damage to the radiation device, the following procedures shall be followed:

Procedures in Case of Suspected Damage

This procedure is designed to prevent the possible spread of radioactive material. It should be carried out preferably by one person, conducted in as limited a physical area as possible. Because of its physical location and its construction (helium arc welded stainless steel), it is highly unlikely that any damage will occur to the source. However, in the case of actual or suspected damage, the following procedure should be followed:

1. Place the entire probe in a plastics bag and seal, or in any container with a tight lid.
2. Notify the RSO and the nearest State Department of Public Health or the nearest NRC office. This information will be given to the field engineer at the time of his receiving the gauge.
3. Isolate the sealed unit in an unused and secure area.
4. All personnel involved should thoroughly wash their hands.
5. If possible, rope off the area involved until such time as the proper authorities have arrived to investigate.
6. Call NDC Systems collect during normal west coast working hours at (818) 358-1871.
7. Call (818) 357-9636 on weekends or after normal working hours and leave a message.

Application For Material License
Attachment To NRC Form 313
Wisconsin Wires, Inc.
6480 West College Avenue
Appleton, WI 54915

Loss of Gauge

In the event that the gauge is lost, the following procedure should be followed:

1. If lost by a common carrier, the RSO should be notified. It should be emphasized to the common carrier personnel that the lost article is a radioactive device.
2. If the gauge is not traced within 24 hours, notify the responsible State or NRC Radiological Health Office.
3. If stolen from car or if car is stolen with unit in the trunk, notify RSO and all relevant State Police, Sheriff, and/or Highway Patrol offices. If not recovered within 24 hours, notify the responsible State or NRC Radiological Health office.

10.1 and 10.2

No personal monitoring devices will be required. This device has been in use for over 15 years as a portable mass measuring instrument and has an unblemished safety record.

The energy output is soft 60 KEV radiation. The maximum exposure at 1 foot with the shutter open is 3.5 mrem. The operator is never positioned in the radiation beam and is normally at least 3 feet away from the active end of the device. The use of the device is intermittent and would average less than 10 hours per week.

Because of the above, and based on years of experience by other users performing the same measurements with this device, it is extremely unlikely that the user or plant personnel would receive any detectable radiation from the use of this device.

- 10.3 Leak Tests will be performed at six-month intervals by NDC Systems or by persons specifically licensed to do so.

Application For Material License
Attachment To NRC Form 313
Wisconsin Wires, Inc.
6480 West College Avenue
Appleton, WI 54915

- 10.4 The radioactive device will be returned to NDC Systems for all repairs. No maintenance (other than wipe tests) will be required.
- 10.5 The RSO will make all necessary arrangements for transportation of the gauge. All applicable State and Federal transportation regulations shall be observed.
- 10.6 Operating and emergency procedures as described above will be provided to each field engineer using the device.
- 11 No waste is generated. Device to be returned to NDC Systems in the event of cessation of use.

Revised December 1985

Application For Material License
Attachment To NRC Form 313
Wisconsin Wires, Inc.
6480 West College Avenue
Appleton, WI 54915

ITEM NO. 11

WASTE MANAGEMENT

No waste is generated. Device to be returned to NDC Systems
in the event of cessation of use.