

DEC 2 1986

Christ Hospital
ATTN: Jack Cook
President
2139 Auburn Avenue
Cincinnati, OH 45219

Gentlemen:

Enclosed is Amendment No. 14 to your NRC License No. 34-03831-03 in accordance with your request.

We remind you that Conditions 16. and 17. of your license require you to perform a radiation survey and to submit a survey report:

- a. Each time your teletherapy source is replaced.
- b. Whenever you make any changes in the shielding, location of use of your teletherapy installation that could affect radiation levels in surrounding areas.

A report of the results of the required survey should be sent to the U.S. Nuclear Regulatory Commission, Region III, 799 Roosevelt Road, Glen Ellyn, Illinois 60137, ATTN: Chief, Materials Licensing Section not more than 30 days after each installation of a teletherapy source.

Appendix F of the enclosed "Guide for the Preparation of Applications for Licenses for Medical Teletherapy Programs" dated December 1985, contains information that should be addressed in order to adequately satisfy the license requirements. You may find it helpful to refer to Appendix F when preparing your survey report.

If you have any questions or require clarification on any of the information stated above, you may contact us at (312) 790-5625.

Sincerely,

8704170273 861202
REG3 LIC30
34-03831-03 PDR

Original Signed By
Patricia J. Whiston
Materials Licensing Section

Enclosures:

1. Amendment No. 14
2. Guide for Teletherapy Programs

RJII
pdw
Whiston/pd
12/1/86