

L&R:IB:WFN

OCT 30 1961

Dr. J. H. Buck  
The Budd Company  
P. O. Box 245  
Phoenixville, Pa.

Dear Doctor Buck:

Reference is made to the Budd Company leak test kit No. LT-100. Item 7 of the instructions which are provided the user requires that an ICC label be filled in with a statement of the type of isotope, the activity, and the number of radiation units. Experience has indicated the need for requiring the user of such kits to make a physical radiation survey of the wipe in order to determine the possibility of gross contamination.

It is requested that the Budd Company revise the instructions provided to users of the LT-100 kit to include specific instructions requiring a survey of the wipe.

A copy of the revised instructions should be submitted to this office.

Very truly yours,

William O. Miller  
Senior Licensing Reviewer  
Isotopes Branch  
Division of Licensing and  
Regulation

37-611-4

General Ledger

279-340