

Bozeman Deaconess Hospital
Radiation Therapy Department
ATTN: Gary Kenner
Administrator
22 West Beall
Bozeman, Montana 59715

FEB 19 1987

Gentlemen:

Please find enclosed Amendment No. 01 to your NRC material license. You should review this amendment carefully and be sure that you understand all conditions. If you have any questions, you may contact the reviewer who signed your license amendment at 817/860-8100.

At your request we have amended License Number 25-10994-03 to authorize the Picker Corporation Model 6235 (4M/60) teletherapy unit to be used for storage only. Condition 15 has been added to your license to require your institution to perform the following:

1. Lock the teletherapy control panel in the "off" position.
2. At intervals of not less than 6 months verify that the source and control panel are locked.
3. Maintain records of this verification for inspection by the Commission. Verification records should include the date of inspection, control panel status, teletherapy head status, survey meter used, and individual conducting the inspection/verification.

Prior to reestablishing patient treatment at the existing facility or relocating the teletherapy unit to another facility your license must be amended.

Please be advised that you must conduct your program involving radioactive materials in accordance with the conditions of your NRC license, representations made in your license application, and NRC regulations. In particular, note that you must:

1. Operate in accordance with NRC regulations 10 CFR Part 19, "Notices, Instructions and Reports to Workers; Inspections," 10 CFR Part 20, "Standards for Protection Against Radiation," and other applicable regulations.

RIV:NMSS
JEWitten;df
2/19/87

C:NMSS
RJEverett
2/12/87

B706300659 B70219
REG4 LIC30
25-10994-03 PDR

2. Possess radioactive material only in the quantity and form indicated in your license.
3. Use radioactive material only for the purpose(s) indicated in your license.
4. Notify NRC in writing of any change in mailing address (no fee required if the location of radioactive material remains the same).
5. Request and obtain appropriate amendments if you plan to change ownership of your organization, change locations of radioactive material, or make any other changes in your facility or program which are contrary to your license conditions or representations made in your license application and any supplemental correspondence with NRC. A license fee may be charged for the amendments if you are not in a fee-exempt category.
6. Submit a complete renewal application with proper fee, or termination request at least 30 days before the expiration date on your license. You will receive a reminder notice approximately 90 days before the expiration date. Possession of radioactive material after your license expires is a violation of NRC regulations.
7. Request termination of your license if you plan to permanently discontinue activities involving radioactive material.

You will be periodically inspected by NRC. Failure to conduct your program in accordance with NRC regulations, license conditions, and representations made in your license application and supplemental correspondence with NRC will result in enforcement action against you. This could include issuance of a notice of violation; imposition of a civil penalty; or an order suspending, modifying, or revoking your license as specified in the General Policy and Procedures for NRC Enforcement Action, 10 CFR Part 2, Appendix C. Since serious consequences to employees and the public can result from failure to comply with NRC requirements, prompt and vigorous enforcement action will be taken when dealing with licensees who do not achieve the necessary meticulous attention to detail and the high standard of compliance which the NRC expects of its licensees.

Thank you for your cooperation.

Sincerely,

Original signed by
JACK E. WHITTEN

R. J. Everett, Chief
Nuclear Materials Safety Section

Enclosure:
As stated