

19 MAY 1987

License No. 37-13251-02  
Docket No. 030-01006  
Control No. 106609

Slippery Rock University  
ATTN: Dr. Cyril E. Sagan  
Department of Chemistry  
Slippery Rock, Pennsylvania 16057

Gentlemen:

This is in reference to your letter dated April 14, 1987, which contains additional information concerning your request to renew License No. 37-13251-02. In order to continue our review, we need the following additional information:

1. Please identify your method of assuring that your possession limits are not exceeded. We recommend that a purchase order be reviewed by the Radiation Safety Officer.
2. Confirm that Room 313 VSH will be locked when authorized personnel are not present.
3. Please confirm that protective gloves will be worn while examining and opening incoming packages containing licensed material.
4. Confirm whether radioactive waste containers are labeled as such. A method for identifying these containers should be established.
5. The most efficient method for detecting low energy beta emitters is through the use of liquid scintillation counting. Please confirm that this method of detection will be used.

We will continue our review upon receipt of this information. Please reply in duplicate to my attention at the Region I office and refer to Mail Control No. 106609.

BB010504B2 B70623  
REG1 LIC30  
37-13251-02 PDR

OFFICIAL RECORD COPY

ML 007 JOUSTRA - 0001.0.0  
05/14/87

ML1B

19 MAY 1987

In order to continue prompt review of your application, we request that you submit your response to this letter within 30 calendar days from the date of this letter.

Sincerely,

Original Signed By:  
Judith A. Joustra

Judith A. Joustra  
Nuclear Materials Safety Section B  
Division of Radiation Safety  
and Safeguards

RI:DRSS (AS)  
Joustra/bc

05/18/87

OFFICIAL RECORD COPY

ML 007 JOUSTRA - 0002.0.0  
05/14/87