

AUG 28 1987

Hospital Metropolitano
Radiotherapy Institute
ATTN: Hiram Soler Favale, M.D.
Medical Director
P. O. Box EH, Caparra Heights Sta.
San Juan, PR 00922

Gentlemen:

SUBJECT: REQUEST FOR ADDITIONAL INFORMATION CONCERNING A MATERIAL LICENSE
APPLICATION (REFERENCE: 251712, 030-17482)

This refers to your letter dated July 15, 1987, for renewal of License Number 52-16033-02. The following additional information is needed in support of your request.

1. In Item 10.5 (Operating Procedures) of your letter, it states that a copy is already on file with the NRC. I was unable to locate these procedures; therefore, make a clear and specific reference to the document by title and date or submit your procedures as described in 10.5.2 of TASK 414-4.
2. In Item 10.6 (Emergency Procedures) of your letter, it states that a copy is already on file with the NRC. I was also unable to locate these procedures; therefore, make a clear and specific reference to this document or submit your procedures as described in 10.5.3 of TASK 414-4.
3. Item e. of your letter states that all conditions are as described in correspondence dated April 17, 1986. Only head leakage radiation levels were submitted with this letter which did not satisfy the requirements of Condition 18 of your license. Please submit the remainder of the information required following your last source replacement.
4. Enclosed are copies of the latest information submitted for Hospital Metropolitano's ALARA program and Radiation Safety Committee. Please confirm that these are still applicable. Note that the committee membership included a nurse, as required, which was not in your July 15, 1987 letter.

Our review of your application will continue upon receipt of the above information. Please provide two copies of your reply and reference Mail Control Number 251712.


In order to continue prompt review of your application, we request that you submit your response to this letter within thirty (30) calendar days from the date of this letter.

8902160072 880126
REG2 LIC30
52-16033-02 PNU

AUG 28 1987

If you should have any questions, do not hesitate to contact me at (404) 331-2675.

Sincerely,



Carol A. Connell
Radiation Specialist
Nuclear Materials Safety Section

RII *CAC*
CConnell:tj
8/28/87

RII *SP*
~~SP~~ Potter
8/28/87