

FEB 19 1987

Doctors King, Ram & Associates, Inc.
6789 Ridge Road
Parma, OH 44129

Docket No. 030-29710

Gentlemen:

We have reviewed your January 14, 1987 application for a new byproduct material license and find we will need the following information:

1. Please confirm that the bone densitometer you will purchase is a Norland Corporation unit. Your application did not specifically mention the manufacturer's name.
2. Please submit sketches of the two places of use showing where the unit will be used and stored.
3. Please provide a description of your security procedures to prevent unauthorized persons from gaining access to the bone densitometer.

If you have any questions or require clarification on any of the information stated above, you may contact us at (312) 790-5625.

We will continue our review of your application upon receipt of this information. Please reply in duplicate, within 30 days, and refer to Control Number 82781.

Sincerely,

Original Signed By
Patricia M. Vacherlon
Materials Licensing Section

8707100312 870317
REG3 LIC30
34-24583-01 PDR

RLL
PMV
Vacherlon/pd
2/19/87