

FEB 29 1988

University of Wisconsin-Green Bay
ATTN: Jack C. Norman
Associate Professor of
Chemistry
Chair, Radiation Safety
Committee
2420 Nicolet Drive
Green Bay, WI 54301-7001

Gentlemen:

Enclosed is Amendment No. 04 renewing your NRC License No. 48-13818-01 in accordance with your request.

Please be advised that you are required to comply with conditions of Section 33.17(a) of 10 CFR Part 33, "Specific License of Broad Scope for Licensed Material."

Please review the enclosed document carefully and be sure that you understand all conditions. You must conduct your program involving radioactive materials in accordance with the conditions of your NRC license, representations made in your license application, and NRC regulations. In particular, note that you must:

1. Operate in accordance with NRC regulations 10 CFR Part 19, "Notices, Instructions and Reports to Workers; Inspections," 10 CFR Part 20, "Standards for Protection Against Radiation," and other applicable regulations.
2. Possess radioactive material only in the quantity and form indicated in your license.
3. Use radioactive material only for the purpose(s) indicated in your license.
4. Notify NRC in writing of any change in mailing address.
5. Request and obtain appropriate amendment if you plan to change ownership of your organization, change locations of radioactive material, or make any other changes in your facility or program which are contrary to your license conditions or representations made in your license application and any supplemental correspondence with NRC. Any amendment request should be accompanied by the appropriate fee specified in 10 CFR Part 170.

6. Submit a complete renewal application with proper fee or termination request at least 30 days before the expiration date on your license. You will receive a reminder notice approximately 90 days before the expiration date. Possession of radioactive material after your license expires is a violation of NRC regulations.
7. Request termination of your license if you plan to permanently discontinue activities involving radioactive material prior to your expiration date.

You will be periodically inspected by NRC. Failure to conduct your program in accordance with NRC regulations, license conditions and representations in your license application will result in enforcement action against you in accordance with the General Policy and Procedures for NRC Enforcement Actions, 10 CFR Part 2, Appendix C.

If you have any questions or require clarification of any of the above stated information, contact us at (312) 790-5625.

Sincerely,

Original Signed By
Kathy J. Graden
Materials Licensing Section

Enclosure: Amendment No. 04

RIII

Graden/pd
2/20/88

CONVERSATION RECORD

TIME 10:19 AM

DATE 2/10/88

TYPE

☐ VISIT☐ CONFERENCE☒ TELEPHONE☐ INCOMING☒ OUTGOING

ROUTING

NAME/SYMBOL

INT

Location of Visit/Conference:

NAME OF PERSON(S) CONTACTED OR IN CONTACT
WITH YOU

Jack Norman

ORGANIZATION (Office, dept., bureau,
etc.)

Univ. of WI Green

TELEPHONE NO.

(414) 465-2371

SUBJECT

Control No. 80876 Bay

Response to defusing letter

SUMMARY

Inform licensee that survey meter calibrations are not adequate, procedures are not adequate, therefore, NRC will condition licensee to have survey instruments calibrated by the device manufacturer, or another person specifically authorized by the NRC or an Agreement State to perform such services.

Mr. Norman was agreeable to the above condition being added to the license.

ACTION REQUIRED

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

issue renewal

Kathy J. Graden 2/10/88

ACTION TAKEN

SIGNATURE

TITLE

DATE