



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
631 PARK AVENUE
KING OF PRUSSIA, PENNSYLVANIA 19406

DEC 30 1986

License No. 29-17005-01
Docket No. 030-12043
Control No. 105979

New Jersey Steel Corporation
ATTN: Richard Urbanowski
Maintenance Superintendent
P.O. Box 11
North Crossman Road
Sayreville, NJ 08872

Gentlemen:

This is in reference to your letter dated October 10, 1986, in response to questions raised in our letter dated September 20, 1986. In order to continue our review, we need the following questions answered regarding your reply to question 4:

1. You did not provide a complete response to our question on your procedures for removing and installing the Ronan gauges. Submit the step-by-step procedures, including the radiation safety procedures to be followed when performing these functions.
2. Your requested exemption from the requirements of 10 CFR 20.202 (enclosed) can only be adequately reviewed upon analysis of exposure data obtained from past dosimetry records on employees who have been performing the installations and removals. Therefore, you will need to initially supply such personnel with the necessary dosimetry to establish a history of exposure from which to base your request for exemption From Section 20.202. Specify the name of the supplier, type of dosimetry system to be used and the frequency of exchange of these devices.

We will continue our review of your application upon receipt of the above information. Please reply in duplicate, referencing Control No. 105979. In order to continue prompt review of your application, we request that you submit your response to this letter within 30 calendar days from the date of this letter.

8707210232 870320
REG1 LIC30
29-17005-01 PDR

"OFFICIAL RECORD COPY" ML10

Sincerely,

Thomas L. Thompson

for John E. Glenn, Ph.D., Chief
Nuclear Materials Safety Section B
Division of Radiation Safety and
Safeguards

Enclosure: 10 CFR 20

RI:DRSS
Varela

12/ /86

for *PT*
RI:DRSS
Glenn

12/30/86

OFFICIAL RECORD COPY