

APR 21 1987

Cleveland Museum of Art  
ATTN: Jane Glaubinger  
Associate Curator  
Department of Prints and  
Drawings  
11150 East Blvd.  
Cleveland, OH 44106

License No.: 34-24631-01  
Docket No.: 030-18725  
Control No.: 381203

We have reviewed your amendment request dated April 28, 1986 requesting relief from your leak test condition, and find that we will need additional information as follows:

We have received a response from our Sealed Source & Device Certification Branch and they declined to grant complete relief from the leak test procedure based on experiences where the beta plates have cracked or broken causing the C-14 to "leak". However, they would concur with an amendment to change your leak test procedure to read "at intervals not to exceed six months or after each use whichever is greater," contingent on the maintenance of a signed use log by your authorized users.

Please inform me if amending the leak-test condition as outlined above would be satisfactory to you. If so, please send a copy of the use log that would be maintained. It should include the date the beta plate is used and the name of the user.

We will continue our review of your application upon receipt of this information. Please reply in duplicate, within 30 days, and refer to Control Number 381203.

Sincerely,

Original Signed By  
William J. Adam, Ph.D.  
Materials Licensing Section

8710080260 870522  
REG3 LIC30  
34-24631-01 PDR

RIII  
WJG  
Adam/pd  
4/21/87