



UNITED STATES
NUCLEAR REGULATORY COMMISSION

REGION IV
611 RYAN PLAZA DRIVE, SUITE 1000
ARLINGTON, TEXAS 76011

NMLS:CLC
Control No. 462387

February 6, 1989

Erlanger Tubular Corp.
ATTN: Carl Beck, Director
Technical Services
5610 Bird Creek Ave.
Catoosa, Oklahoma 74015

Gentlemen:

This is in reference to your application dated January 30, 1989, requesting a byproduct material license for use of sealed sources and devices in portable gauging devices. We have completed review of your application and have the following comments and need for additional information. These were discussed telephonically with Mr. Beck by the undersigned on February 6, 1989.

1. We understand from our telephone discussion that you wish to replace Mr. Richard Cope as your candidate for Radiation Safety Officer. You should provide information per our telephone discussion regarding the new candidate's training and experience.
2. You should confirm that all users of the gauge will successfully complete the training course presented by International Pipe Inspectors Association before using the device. Use of other training courses must be specifically authorized by NRC through license amendment.
3. Confirm that records of the training identified in the above Items will be maintained for inspection by NRC for two years following completion of training. Also you should confirm that you will maintain a listing of authorized users.
4. Confirm that the gauge will be maintained in a locked enclosure that will prevent access by unauthorized persons, including persons employed by Erlanger who have not completed the required training.
5. Item 3. of your application indicates that the gauge will be used only at your facility at 5610 Bird Creek Ave., Catoosa, Oklahoma. Please confirm that the gauge will not be transported to temporary job sites.
6. Your application lists your address as on Birdcreek Avenue; whereas, your letterhead indicates the street name as Bird Creek. Please identify which is the correct street name.

8903130844 890213
REG4 LIC30
35-26975-01 PNU

copy

Erlanger Tubular Corp.

-2-

If we do not receive a reply from you within 30 calendar days from the date of this letter, we shall assume that you do not wish to pursue your application. Please reply in duplicate and refer to Control No. 462387.

Sincerely,

Original signed by
C. L. Cain

Charles L. Cain, Chief
Nuclear Materials Licensing Section

(FOR LFMS USE)
INFORMATION FROM LTS

PROGRAM CODE: _____
STATUS CODE: 3
FEE CATEGORY: _____
EXP. DATE: 0
FEE COMMENTS: _____

BETWEEN:

LICENSE FEE MANAGEMENT BRANCH, ARM
AND
REGIONAL LICENSING SECTIONS

LICENSE FEE TRANSMITTAL

A. REGION

1. APPLICATION ATTACHED
APPLICANT/LICENSEE: ERLANGER TUBULAR CORPORATION
RECEIVED DATE: 890130
DOCKET NO.: 3030975
CONTROL NO.: 462387
LICENSE NO.:
ACTION TYPE: NEW LICENSEE

2. FEE ATTACHED \$ 299.00
AMOUNT: 299.00
CHECK NO.: 14517

3. COMMENTS

SIGNED
DATE

Billig Grynyski
1/30/89

B. LICENSE FEE MANAGEMENT BRANCH (CHECK WHEN MILESTONE 03 IS ENTERED 1/✓)

1. FEE CATEGORY AND AMOUNT: 38 (6230)

2. CORRECT FEE PAID. APPLICATION MAY BE PROCESSED FOR:
AMENDMENT
RENEWAL
LICENSE

3. OTHER

SIGNED
DATE

Mr. Mesari
7/2/89

RECEIVED
FEB - 7 1989