

L...
DMB COPY

JUL 21 1987

03000537

St. Catherine's Hospital
ATTN: Michael Muenzberg
President
3556 Seventh Avenue
Kenosha, WI 53141-1490

License No. 48-01670-02

Gentlemen:

Thank you for your letter received July 13, 1987, informing us of the steps you have taken to correct the violation identified in our letter dated June 26, 1987. As a result of a telephone conversation on July 14, 1987, with Mr. Marc Wojciechowski, we understand that you will request an amendment to your license for authorization to replace the treatment timer, a component of your Eagle Signal teletherapy timer. It should be understood that no activity is in accord with NRC regulatory requirements unless and until it is specifically authorized in a NRC license. Steps should be taken to ensure that NRC licensed materials are hereafter possessed and used only as authorized by your license. We will examine this matter during a future inspection.

Sincerely,

DJS

D. J. Sreniawski, Chief
Nuclear Materials Safety
Section 2

cc w/ltr rcd July 13, 1987:
DCS/RSB (RIDS)

B707290151 B70721
REG3 LIC30
48-01670-02 PDR

RIII

SPH
for Pankratz/abl

7/17

RIII

YES

SPH
Lasuk

7/17/87

RIII

YES
SPH
Sreniawski

7/21/87

IE07

||