

VOID SHEET

TO: License Fee Management Branch

FROM: RIII - COLLEEN C. CASEY

SUBJECT: VOIDED APPLICATION

Control Number:

304143

Applicant:

ST. JOSEPH MERCY HEALTH SYSTEM

License Number:

21-00943-03

Docket Number:

030-01997

Date Voided:

7/13/98

Reason for Void:

*The licensee sent in letter dated 6/30/98 as a response to an old control no. <sup>(302896)</sup> 304143. It was in response to 304089. The 6/30/98 ltr. is being combined with 304089 as it should be.*

Colleen C. Casey  
Signature

7/13/98  
Date

Attachment:  
Official Record Copy of  
Voided Action

FOR LFMB USE ONLY

- ☐ Refund Authorized and processed
- ☐ No Refund Due
- ☒ Fee Exempt or Fee Not Required

Comments: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Log completed ☒

Processed by: SAC 7/21/98

9807300006 980713  
PDR ADOCK 03001997  
C PDR

ML  
30 DH

BETWEEN:

License Fee Management Branch, ARM  
and  
Regional Licensing Sections

(FOR LFMS USE)  
INFORMATION FROM LTS

Program Code: 02120  
Status Code: 0  
Fee Category: 7C 2B  
Exp. Date: 20050831  
Fee Comments: CODE 21  
Decom Fin Assur Req'd: N

LICENSE FEE TRANSMITTAL

A. REGION

1. APPLICATION ATTACHED

Applicant/Licensee: SAINT JOSEPH MERCY HEALTH SYSTEM  
Received Date: 980706  
Docket No: 3001997  
Control No.: 304143  
License No.: 21-00943-03  
Action Type: Amendment

2. FEE ATTACHED

Amount: Ø  
Check No.: Ø

3. COMMENTS

Signed  
Date

B. LICENSE FEE MANAGEMENT BRANCH (Check one if entered / ✓ /)

1. Fee Category and Amount: 7C 2B

**FEE NOT REQUIRED**

2. Correct Fee Paid, / Application may be processed for:

Amendment ✓  
Renewal  
License

3. OTHER

Signed  
Date

Log	<u>File to III</u>
Remitter	<u>SC</u>
Check No.	<u>7113</u>
Amount	<u>7C 2B</u>
Fee Category	<u>7C 2B</u>
Type of Fee	<u>AMD</u>
Date Check Rec'd	<u>7/17/98</u>
Date Completed	<u>7/17/98</u>
By:	<u>SC</u>

SAINT  
JOSEPH  
MERCY  
HEALTH SYSTEM



A Member of Mercy Health Services

ST. JOSEPH MERCY HOSPITAL  
5301 East Huron River Drive  
P.O. Box 995  
Ann Arbor, MI 48106-0995  
Telephone (313) 712-3456

June 30, 1998

Colleen C. Casey, Regional Administrator, Region III  
Nuclear Materials Licensing Branch  
United States Nuclear Regulatory Commission  
801 Warrendale Road  
Lisle, Illinois 60532-4351

RE: Request for Amendment  
License Number: 21-00943-03  
Control No.: 302846

Dear Ms. Casey,

This is to clarify our request for an amendment to our license for our HDR brachytherapy program. We have the Microselection HDR 105.999. The sealed source for this device is Mallinckrodt ID# DRN07736 (Iridium 192) manufactured by Mallinckrodt Medical B.V., Westerduinweg 3, NL-1755 LE Petten.

Per your discussion with Ralph Schooping of Nucletron Corporation, we will not store a 12 curie replacement source for decay to 10 curies at our facility. Rather, Nucletron will store the 12 curie source at a facility in Maryland to be shipped for replacement when it has decayed to 10 curies.

Further, when Nucletron amends its source registry MD-497-D-108-S, our license will be amended at no charge to us.

Thank you for your speedy action on this matter.

If you have further questions you can contact Jim Spicka or Sandy Seehaver at (734) 712-3595.

Sincerely,

Lenore Andres  
Administrative Director  
Radiation Oncology

LA/max

*Control # 302846*  
**FEE NOT REQUIRED**

304143

**RECEIVED**

**JUL 06 1998**

**REGION III**

*pm: 6/30/98* **JUL 06 1998**





UNITED STATES  
NUCLEAR REGULATORY COMMISSION

REGION III  
801 WARRENVILLE ROAD  
LISLE, ILLINOIS 60532-4351

July 9, 1998

James T. Spica, M.S.  
Radiation Safety Officer  
Saint Joseph Mercy Health System  
St. Joseph Mercy Hospital  
P.O. Box 995  
Ann Arbor, MI 48106-0995

SUBJECT: ACKNOWLEDGEMENT OF CORRESPONDENCE  
(Letter Dated June 30, 1998)

Dear Licensee:

In response to your request, we have completed the initial processing, which is an administrative review of your application for a(n):

☐ New License                      ☒ Amendment                      ☐ Renewal  
☐ Termination                      ☐ Auth User (Amendment not required)  
☐ Other \_\_\_\_\_

No administrative deficiencies were identified during this initial review. However, it should be noted that a technical review may identify omissions in the submitted information.

It appears that your request is routine (see 1-3 below, as applicable).

1. New and amendment actions are normally completed within 90 days, unless we find major deficiencies, or policy issues requiring central program office assistance.
2. Renewal actions are normally completed within 180 days, however, under timely filing (before expiration), you may continue to operate under your existing license.
3. Termination actions are normally completed within 90 days, unless confirmatory surveys following decontamination/decommissioning activities are involved.

A copy of your correspondence has been forwarded to our Licensing Fee and Debt Collection Branch (301/415-6097) for approval of the fee category and amount, if required.

We will try to complete your request as soon as practicable. Any correspondence about this request should reference the control number. Please direct any questions concerning your request to the Materials Licensing Branch at (630) 829-9887.

Materials Support Branch

Mail Control No. 304143  
License No. 21-00943-03