

12 AUG 1987

Docket No. 50-289

GPU Nuclear Corporation
ATTN: Mr. H. D. Hukill
Vice President and Director of TMI-1
P. O. Box 480
Middletown, Pennsylvania 17057

Gentlemen:

Subject: Results of Non-Radiological Chemistry Standards Inspection
Activities for All Region I licensees

Routine safety inspections were conducted by this office during the period 1985 to 1986 at USNRC Region I licensees in the area of non-radiological water chemistry laboratory operations. The purpose of this report is to present results of all licensees inspected, as well as results of measurements made by Brookhaven National Laboratory. Each licensee is identified by a code. Your code is 20.

No reply to this letter is required.

Sincerely,

Original Signed By:

Ronald R. Bellamy
Thomas T. Martin, Director
Division of Radiation Safety
and Safeguards

Enclosure: Combined Results of Non-radiological Standards Comparison
Inspection Activities for Region I Licensees

OFFICIAL RECORD COPY

LTR RAD RESULTS - 0051.0.0
06/17/87

8708190269 870812
PDR ADOCK 05000289
Q PDR

1/1 IE06


12 AUG 1987

cc w/encl:

R. J. Toole, Operations and Maintenance Director, TMI-1
C. W. Smyth, Manager, TMI-1 Licensing
R. J. McGoey, Manager, PWR Licensing
E. L. Blake, Jr., Esquire
TMI-1 Hearing Service List
Public Document Room (PDR)
Local Public Document Room (LPDR)
Nuclear Safety Information Center (NSIC)
NRC Resident Inspector
Commonwealth of Pennsylvania

bcc w/encl:


Region I Docket Room (with concurrences)
W. D. Travers, Director, TMI-2 Cleanup Project Directorate
J. Goldberg, OELD: HQ
Management Assistant, DRMA (w/o encl)
DRP Section Chief
John Thoma, PM, NRR
K. Abraham, PAO
Robert J. Bores, DRSS


RI:DRSS
Kottan/gcb/mlb

8/24/87


RI:DRSS
Zibulsky

6/18/87


RI:DRSS
Pasciak

8/24/87


RI:DRSS
Bellamy

8/27/87


RI:DRSS
Martin

6/17/87

OFFICIAL RECORD COPY

LTR RAD RESULTS - 0052.0.0
06/17/87

TMI-1 Hearing Service List

Sheldon J. Wolfe, Chairman
Administrative Judge
Atomic Safety & Licensing Board Panel
U.S. Nuclear Regulatory Commission
Washington, D.C. 20555

Dr. Oscar H. Paris
Administrative Judge
Atomic Safety & Licensing Board Panel
U.S. Nuclear Regulatory Commission
Washington, D.C. 20555

Frederick J. Shon
Administrative Judge
Atomic Safety & Licensing Board Panel
U.S. Nuclear Regulatory Commission
Washington, D.C. 20555

Joanne Doroshow, Esquire
Three Mile Island Alert, Inc.
315 Pepper Street
Harrisburg, PA 17102

Louise Bradford
1011 Green Street
Harrisburg, PA 17102

Thomas Y. Au
Assistant Counsel Commonwealth
of Pennsylvania
Dept. of Environmental Resources
Bureau of Environmental Resources
Room 505, Executive House
P. O. Box 2357
Harrisburg, PA 17120

Bruce W. Churchill, Esquire
Shaw, Pittman, Potts & Trowbridge
2300 N Street, N.W.
Washington, D.C. 20037

Atomic Safety & Licensing Board
Panel
U.S. Nuclear Regulatory Commission
Washington, D.C. 20555

Atomic Safety & Licensing Appeal
Board Panel
U.S. Nuclear Regulatory Commission
Washington, D.C. 20555

Docketing & Service Section
Office of the Secretary
U.S. Nuclear Regulatory Commission
Washington, D.C. 20555

Mary E. Wagner, Esquire
Office of Executive Legal Director
U.S. Nuclear Regulatory Commission
Washington, DC 20555

USNRC
Resident Inspector
Box 311
Middletown, PA 17057

U.S. NUCLEAR REGULATORY COMMISSION
REGION I

Combined Results of Non-Radiological Standards
Comparison Inspection Activities for
Region I Licensees

Inspectors:

Harvey Zibulsky
Harvey Zibulsky, Chemist, Effluents Radiation
Protection Section, EP&RPB, DRSS

J. J. Kottan
James J. Kottan, Radiation Laboratory Specialist,
Effluents Radiation Protection Section, EP&RPB, DRSS

Approved By:

Walter J. Pasciak
Walter J. Pasciak, Chief, Effluents Radiation
Protection Section, EP&RPB, DRSS

8748170369

39pp

Details

1. Introduction

Region I has implemented an inspection program in the area of non-radio-logical water chemistry. As part of this inspection program, licensees are required to analyze test standards prepared by Brookhaven National Laboratory (BNL) for NRC Region I. The data from the first round of this part of the inspection program are now available from all power reactors in Region I and are presented here.

The data as presented permits each Region I nuclear power plant chemistry laboratory to compare its performance with all other Region I nuclear power plant laboratories. The data for the BNL known values are also presented so that comparison of the average from all Region I sites to the known value may be made. The data were taken from inspection reports, but are coded here. The letter transmitting this report identified the individual code(s) for each addressee. Section 3.0 lists the inspection reports data were taken from. The data are presented in Appendix B.

2. Data Analysis

The data were analyzed as follows.

1. The analytical data from all Region I sites were entered into a data base at BNL. The analytical data consisted of each licensee's result for each analyte at each analyzed concentration.
2. At each concentration a grand average and standard deviation were calculated using the data from all sites.
3. A range consisting of the grand average plus or minus two standard deviations was established. Any values outside of this range were considered outliers.
4. A new grand average and standard deviation were calculated excluding the outliers determined in step 3.
5. A range consisting of the grand average plus or minus two standard deviations was again established. Any values outside this range were considered outliers.
6. The above process was repeated until a range was obtained with no outliers. This range was used for the data comparison.

A sample calculation is presented in Appendix A.

It must be recognized that the data used to generate the averages are not of equal weight. For example, some laboratories were permitted to reanalyze the standards if problems could be identified and corrected during the inspection. Also some of the BNL standards were diluted to different final concentrations by the licensees prior to analysis to allow them to analyze within their calibration range. The results were then normalized back to the original dilution in order to enable comparison. In addition, all methods of analysis for a particular analyte were grouped together.

3. References:

<u>Site</u>	<u>Docket No.</u>	<u>Inspection No.</u>
Beaver Valley #1	50-334	85-07
Beaver Valley #2	50-412	86-14
Calvert Cliffs #1	50-317	86-08
Calvert Cliffs #2	50-318	86-08
FitzPatrick	50-333	85-23
Ginna	50-244	85-14
Ct. Yankee	50-213	85-24
Hope Creek	50-354	85-59
Indian Pt. #2	50-247	86-07
Indian Pt. #3	50-286	86-04
Limerick	50-352	86-10
Maine Yankee	50-309	85-33
Millstone #1	50-245	86-04
Millstone #2	50-336	86-04
Millstone #3	50-423	86-13
Nine Mile Pt. #1	50-220	85-10
Oyster Creek	50-219	86-16
Peach Bottom #2	50-277	85-28
Peach Bottom #3	50-278	85-26
Pilgrim	50-293	85-23
Salem #1	50-272	86-03
Salem #2	50-311	86-03
Seabrook	50-443	86-22
Shoreham	50-322	86-11
Susquehanna #1	50-387	85-27
Susquehanna #2	50-388	85-22
Three Mile Island #1	50-289	85-17
Vermont Yankee	50-271	85-38
Yankee Atomic	50-29	85-21

Appendix A

Sample Calculation

The chloride measurements at approximately 30 ppb were chosen for this example.

Step 1

The following data were in the BNL data base for chloride at approximately 30 ppb.

<u>Site No.</u>	<u>Value (ppb)</u>	<u>Site No.</u>	<u>Value (ppb)</u>	<u>Site No.</u>	<u>Value (ppb)</u>
1	30.7	16	27.1	31*	26.7
2	30	17	39.3	32*	71
3	30.2	18	28.8	33	not analyzed
4	24.3	19	28		
5	30.9	20	29.3		
6	30.9	21	23.3		
7	32.7	22	28.9		
8	38.3	23	not analyzed		
9	32	24	30.7		
10	34.2	25	not analyzed		
11	37	26*	41.3		
12	28.7	27	33		
13	not analyzed	28	-		
14	30.7	29	52		
15*	43.35	30	29.7		

*normalized data

Step 2

A grand average and standard deviation were calculated for all 28 values:

$$\text{Grand average} = 33.6804 = \frac{\sum x_i}{n}$$

$$\text{Standard deviation} = 9.5116 = \left[\frac{\sum (x_i - \bar{x})^2}{n-1} \right]^{1/2}$$

Step 3

From step two above, a range of 52.7036 - 14.6572 was established. The value from site 32 was determined to be an outlier.

Step 4

A new grand average and standard deviation were calculated excluding the outlier determined in Step 3.

grand average = 32.2981 (n=27)
standard deviation = 6.1966

Appendix A

Sample Calculation (continued)

Step 5

From step 4 above a range of 44.6914 - 19.9048 was established. The value from site 29 was determined to be an outlier.

Step 6

The above process was completed through seven more iterations as follows:

<u>n</u>	<u>x</u>	<u>s</u>	<u>range</u>	<u>outlier sites</u>
26	31.5404	4.8796	41.2996 - 21.7812	26, 15
24	30.6417	3.8514	38.3444 - 22.9390	17
23	30.2652	3.4571	37.1794 - 23.3510	8, 21
21	30.2143	2.7368	35.6880 - 24.7406	4, 11
19	30.1684	1.9542	34.0768 - 26.2600	10
18	29.9444	1.7419	33.4283 - 26.4605	None

Thus out of a total of 28 values 10 were outliers. The range of 33.4283 - 26.4605, which contained no outliers, was used for final comparison.

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	69.8646	3.6778	24	-	77.2202	71.3 ELEC
FL	144.627	5.18286	15	-	154.992	140 ELEC
FL	303.444	37.8747	18	-	379.194	300
B	1.00262E+06	7309.1	13	-	1.01723E+06	1.004E+06
B	2.98458E+06	37483.4	18	-	3.05955E+06	2.946E+06
B	4.93482E+06	34950.4	17	-	5.00472E+06	4.875E+06
FE	1.27239	.0679298	15	-	1.40825	1.23 TITR
FE	2.4702	.102406	19	-	2.67501	2.4
FE	3.62507	.116941	18	-	3.85895	3.47
CU	1.34168	.068628	17	-	1.47894	1.39 AAGF
CU	2.63135	.0660583	13	-	2.76347	2.73
CU	3.9834	.100391	15	-	4.18418	3.95 AAGF
AIH	1193.62	88.4432	9	-	1370.51	1500 **
AIH	112.915	9.42891	10	-	131.773	157 **
AIH	355.4	28.397	9	-	412.194	443 **
HY	100.592	4.68925	13	-	109.971	96.8 ELEC
HY	20.5538	.916166	13	-	22.3862	19.5
HY	50.55	1.88076	12	-	54.3115	49.9 SPEC

A OUTLIER

RECORD: 5 Code 26

ANALYTE	MEAN	STD DEV	N		RANGE	REPORTED VALUE	
CL	96.4667	3.52271	15		89.4212	103.512	92
CL	698.646	36.778	24		625.09	772.202	647
CL	299.444	17.4195	18		264.606	334.283	413
							**
FL	100.587	6.49237	15		87.6019	113.571	107
FL	144.627	5.18286	15		134.261	154.992	150
FL	303.444	37.8747	18		227.695	379.194	283
							MERT
B	1.00262E+06	7309.1	13		987997	1.01723E+06	1.005E+06
B	2.98458E+06	37483.4	18		2.90962E+06	3.05955E+06	3.011E+06
B	4.93482E+06	34950.4	17		4.86492E+06	5.00472E+06	4.932E+06
							TTR
APH	112915	9428.91	10		94057.2	131773	110000
APH	355400	28397	9		298606	412194	337000
HY	100.592	4.68925	13		91.2138	109.971	95
HY	20.5538	.916166	13		18.7215	22.3862	21.5
HY	50.55	1.88076	12		46.7885	54.3115	56.8
							**
							SPEC
							SPEC

* OUTLIER

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	19.2933	.704543	15	17.8842	20	
CL	139.729	7.35559	24	125.018	128.3	
CL	29.9444	1.74195	18	26.4606	23.3	**
						ELEC
B	1.00262E+06	7309.1	13	987997	1.01723E+06	
B	2.98458E+06	37483.4	18	249096.2E+06	3.05955E+06	
B	4.93462E+06	34950.4	17	4.86492E+06	5.00472E+06	
FE	1272.39	67.9298	15	1136.53	1408.25	
FE	2470.2	102.406	19	2265.38	2675.01	
FE	3625.07	116.941	18	3391.19	3856.95	
						TTR
CU	1341.68	68.628	17	1204.43	1478.94	
CU	2631.35	66.0583	18	2499.23	2763.47	
CU	3983.4	100.391	15	3782.62	4184.18	
						PLAS
NI	1278.97	33.6022	10	1211.76	1346.17	
NI	2576.61	48.6409	11	2479.33	2673.89	
NI	3805.1	187.29	12	3430.53	4179.68	
						PLAS
CR	1238.83	77.0573	10	1084.71	1392.94	
CR	2188.01	201.843	12	1984.32	2791.7	
CR	3698.07	233.211	11	3231.65	4164.49	
						PLAS

** OUTLIER

RECORD: 7 Code 30

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	19.2933	.704543	15	17.8842	20.7024	22.7
CL	69.8646	3.6778	24	62.509	77.2202	68.6
CL	29.9444	1.74195	18	26.4606	33.4283	29.7
						ELEC
B	1.00262E+06	7309.1	13	987997	1.01723E+06	995000
B	2.98458E+06	37483.4	18	2.90962E+06	3.05955E+06	2.953E+06
B	4.93482E+06	34950.4	17	4.86492E+06	5.00472E+06	4.955E+06
FE	848.257	45.2865	15	757.684	938.83	800
FE	1646.8	68.2708	19	1510.26	1783.34	1680
FE	2416.71	77.9603	18	2260.79	2572.63	2500
						TITR
CU	894.455	45.752	17	802.951	985.959	860
CU	1754.23	44.0389	18	1666.16	1842.31	1770
CU	2655.6	66.9275	15	2521.75	2789.46	2580
						AA
HI	852.645	22.4014	10	807.842	897.448	850
HI	1717.74	32.4273	11	1652.89	1782.6	1690
HI	2536.74	124.86	12	2287.02	2786.46	2440
						AA
CR	825.884	51.3716	10	723.141	928.627	750
CR	1592.01	134.562	12	1322.88	1861.13	1540
CR	2465.38	155.474	11	2154.43	2776.33	2430
						AA

RECORD: 8 Code 23

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
FL	20.1173	1.29847	15	17.5204	22.7143	20.7
FL	72.3133	2.59143	15	67.1305	77.4962	68
FL	30.3444	3.78747	18	22.7695	37.9194	30.3
						ELEC
B	1.00262E+06	7309.1	13	987997	1.01723E+06	1.01E+06
B	2.98458E+06	37483.4	18	2.90962E+06	3.05955E+06	2.989E+06
B	4.93482E+06	34950.4	17	4.86492E+06	5.00472E+06	4.929E+06
						TITR
FE	5.08954	.271719	15	4.5461	5.63298	5.05
FE	9.88078	.409625	19	9.06154	10.7	9.65
FE	14.5003	.467762	18	13.5648	15.4358	13.63
						AAGF
CU	5.36673	.274512	17	4.81771	5.91576	4.35
CU	10.5254	.264233	18	9.99694	11.0539	9.02
CU	15.9336	.401565	15	15.1305	16.7367	13.09
						AAGF
						**
						**
						**
W	100.592	4.68925	13	91.2138	109.971	103.7
W	20.5538	.916166	13	18.7215	22.3862	20.6
W	50.55	1.88076	12	46.7885	54.3115	52
						SPEC

OUTLIER

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	19.2933	.704543	15	-	20.7024	**
CL	69.8646	3.6778	24	-	77.2202	**
CL	29.9444	1.74195	18	-	33.4283	
						IC
B	1.00262E+06	7309.1	13	-	1.01723E+06	**
B	2.98458E+06	37483.4	18	-	3.05955E+06	**
B	4.93482E+06	34950.4	17	-	5.00472E+06	**
						TITR
FE	1272.39	67.9298	15	-	1408.25	
FE	2470.2	102.406	19	-	2675.01	
FE	3625.07	116.941	18	-	3858.95	
						AA
CU	1341.68	68.628	17	-	1478.94	
CU	2631.35	66.0583	18	-	2763.47	
CU	3983.4	100.391	15	-	4184.18	
						AA
NI	1278.97	33.6022	10	-	1346.17	**
NI	2576.61	48.6409	11	-	2673.89	**
NI	3805.1	187.29	12	-	4179.68	
						AA
CR	1238.83	77.0573	10	-	1392.94	
CR	2388.01	201.843	12	-	2791.7	
CR	3698.07	233.211	11	-	4164.49	
						AA

RECORD: 11 Code 9

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	9.64667	.352271	15	8.94212	10	IC
CL	69.8646	3.6778	24	62.509	65.3	
CL	29.9444	1.74195	18	26.4606	32	
FL	100.587	6.49235	35	87.6019	100	**
FL	144.627	5.18286	15	134.261	130	
FL	303.444	37.8747	18	227.695	340	
						ELEC
B	1.00262E+06	7309.1	13	987997	1.002E+06	
B	2.98458E+06	37483.4	18	2.90962E+06	2.986E+06	
B	4.93482E+06	34950.4	17	4.86492E+06	4.806E+06	**
FE	5.08954	.271719	15	4.5461	9.2	**
FE	9.88078	.409625	19	9.06154	16.9	**
FE	14.5003	.467762	18	13.5648	20.3	**
						TITR
CU	5.36673	.274512	17	4.81771	5	AAGF
CU	10.5254	.264233	18	9.99694	10.2	
CU	15.9336	.401565	15	15.1305	16	AAGF
AHH	1193.62	88.4432	9	1016.74	1200	
AHH	112.915	9.42891	10	94.0572	100	
AHH	355.4	28.397	9	298.606	360	
HY	100.592	4.68925	13	91.2138	101.3	SPEC
HY	20.5538	.916166	13	18.7215	20	
HY	50.55	1.88076	12	46.7885	51.3	SPEC

RECORD: 12 Code 19

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	69.8646	3.6778	24	-	77.2202	ELEC
CL	29.9444	1.74195	18	-	33.4283	
FE	50.8954	2.71719	15	-	56.3298	**
FE	9.88078	.409625	19	-	10.7	
FE	14.5003	.467762	18	-	15.4358	
CU	53.6673	2.74512	17	-	59.1576	PLAS
CU	10.5254	.264233	18	-	11.0539	**
CU	15.9336	.401565	15	-	16.7367	**
HI	51.1587	1.34409	10	-	53.8469	PLAS
HI	10.3064	.194564	11	-	10.6956	**
HI	15.2204	.749159	12	-	16.7187	**
CR	49.553	3.08229	10	-	55.7176	PLAS
CR	9.55205	.807373	12	-	11.1668	**
CR	14.7923	.932843	11	-	16.658	**

** OUTLIER

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	9.64667	.352271	15	8.94212	10.3512	9.6
CL	69.8646	3.6778	24	62.509	77.2202	83.2
CL	29.9444	1.74195	18	26.4606	33.4283	34.2
						IC
FL	100.587	6.49237	15	87.6019	113.571	109
FL	72.3133	2.59143	15	67.1305	77.4962	74.7
FL	30.3444	3.78747	18	22.7695	37.9194	37.7
						ELEC
B	1.00262E+06	7309.1	13	987997	1.01723E+06	995000
B	2.98458E+06	37483.4	18	2.90962E+06	3.05955E+06	3.013E+06
B	4.93482E+06	34950.4	17	4.86492E+06	5.00472E+06	4.962E+06
						TITR
FE	101.791	5.43438	15	90.9224	112.66	98
FE	98.8078	4.09625	19	90.6154	107	102.3
FE	145.003	4.67762	18	135.648	154.358	146.5
						ICPL
CU	107.335	5.49024	17	96.3541	118.315	111.7
CU	105.254	2.64233	18	99.9694	110.539	103
CU	159.336	4.01565	15	151.305	167.367	163.2
						ICPL
HY	100.592	4.68925	13	91.2138	109.971	106
HY	20.5538	.916166	13	18.7215	22.3862	21.7
HY	50.55	1.88076	12	46.7885	54.3115	53.7
						SPEC

** OUTLIER

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	9.64667	.352271	15	-	10.3512	10
CL	69.8646	3.6778	24	-	77.2202	67.5
CL	29.9444	1.74195	18	-	33.4283	30
						SPEC
B	1.00262E+06	7309.1	13	-	1.01723E+06	1.081E+06
B	2.98458E+06	37483.4	18	-	3.05955E+06	2.919E+06
B	4.93482E+06	34950.4	17	-	5.00472E+06	4.919E+06
						TITR
FE	1272.39	67.9298	15	-	1408.25	1350
FE	2470.2	102.406	19	-	2675.01	2640
FE	3625.07	116.941	18	-	3858.95	3760
						AA
CU	1341.60	68.628	17	-	1478.94	1330
CU	2631.35	66.0583	18	-	2763.47	2720
CU	3983.4	100.391	15	-	4184.18	4060
						AA
NI	1278.97	33.6022	10	-	1346.17	1300
NI	2576.61	48.6409	11	-	2673.89	2550
NI	3805.1	187.29	12	-	4179.68	3640
						AA
CR	1238.83	77.0573	10	-	1392.94	1350
CR	2388.01	201.843	12	-	2791.7	2540
CR	3698.07	233.211	11	-	4164.49	3310
						AA

AA OUTLIER

RECORD: 15 Code 31

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	SPEC
CL	96.4667	3.52271	15	89.4212	103.512	72
CL	698.646	36.778	24	625.09	772.202	663
CL	299.444	17.4195	18	264.606	334.283	267
						**
B	1.00262E+06	7309.1	13	987997	1.01723E+06	994000
B	2.98458E+06	37483.4	18	2.90962E+06	3.05955E+06	2.927E+06
B	4.93482E+06	34950.4	17	4.86492E+06	5.00472E+06	4.885E+06
FE	1272.39	67.9298	15	1136.53	1408.25	1040
FE	2470.2	102.406	19	2265.38	2675.01	2460
FE	3625.07	116.941	18	3391.19	3858.95	3720
						**
CU	1341.68	68.628	17	1204.43	1478.94	1350
CU	2631.35	66.0583	18	2499.23	2763.47	2710
CU	3983.4	100.391	15	3782.62	4184.18	4120
NI	1278.97	33.6022	10	1211.76	1346.17	1270
NI	2576.61	48.6409	11	2479.33	2673.89	2640
NI	3805.1	187.29	12	3430.53	4179.68	3980
CR	1238.83	77.0573	10	1084.71	1392.94	1280
CR	2388.01	201.843	12	1984.32	2791.7	2580
CR	3698.07	233.211	11	3231.65	4164.49	3880
						ICPL

RECORD: 16 Code 27

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	9.64667	.352271	15	-	10.3512	**
CL	69.8646	3.6778	24	-	77.2202	
CL	29.9444	1.74195	18	-	33.4283	
FL	100.587	6.49237	15	-	113.571	ELEC
FL	144.627	5.18286	15	-	154.992	
FL	303.444	37.8747	18	-	379.194	
B	1.00262E+06	7309.1	13	-	1.01723E+06	ELEC
B	2.98458E+06	37483.4	18	-	3.05955E+06	
B	4.93482E+06	34950.4	17	-	5.00472E+06	
FE	5.08954	.271719	15	-	5.63298	TITR
FE	9.88078	.409625	19	-	10.7	**
FE	14.5003	.467762	18	-	15.4358	
CU	5.36673	.274512	17	-	5.91576	AAGF
CU	10.5254	.264233	18	-	11.0539	
CU	15.9336	.401565	15	-	16.7367	**
HY	100.592	4.68925	13	-	109.971	
HY	20.5538	.916166	13	-	22.3862	105.3
HY	50.55	1.88076	12	-	54.3115	20
						51
						SPEC

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	9.64667	.352271	15	-	10.3512	**
CL	69.8646	3.6778	24	-	77.2202	**
FL	40.2347	2.59695	15	-	45.4286	**
FL	72.3133	2.59143	15	-	77.4962	**
FL	30.3444	3.78747	18	-	37.9194	**
						IC
CU	5.36673	.274512	17	-	5.91576	**
CU	10.5254	.264233	18	-	11.0539	**
CU	15.9336	.401565	15	-	16.7367	**
						AAGF
AIH4	1193.62	88.4432	9	-	1370.51	
AIH1	112.915	9.42891	10	-	131.773	
AIH1	355.4	28.397	9	-	412.194	
HY	100.592	4.68925	13	-	109.971	**
HY	205.538	9.16166	13	-	223.862	
HY	101.1	3.76153	12	-	108.623	
						SPEC
						SPEC

REPORT: ID Code 20

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	9.64667	.352271	15	-	10.3512	9.3
CL	69.8646	3.6778	24	-	77.2202	71.3
CL	29.9444	1.74195	18	-	33.4283	29.3
FL	50.2933	3.24619	15	-	56.7857	45
FL	144.627	5.18286	15	-	154.992	140
FL	30.3444	3.78747	18	-	37.9194	25
B	1.00262E+06	7309.1	13	-	1.01723E+06	1.095E+06 **
B	2.98458E+06	37483.4	18	-	3.05955E+06	3.074E+06 **
B	4.93482E+06	34950.4	17	-	5.00472E+06	5.004E+06
ACH	1193.62	88.4432	9	-	1370.51	1050
ACH	112.915	9.42891	10	-	131.773	116
HY	100.592	4.68925	13	-	109.971	97.5
HY	20.5538	.916166	13	-	22.3862	19.8
HY	50.55	1.88076	12	-	54.3115	49.5

OUTLIER

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	48.2333	1.76136	15	-	51.7561	58
CL	69.8646	3.6778	24	-	77.2202	105
CL	59.8889	3.4839	18	-	66.8567	86.7
FL	20.1173	1.29847	15	-	22.7143	24.3
FL	144.627	5.18286	15	-	154.992	151
FL	30.3444	3.78747	18	-	37.9194	34
B	501308	3654.55	13	-	508617	502500
B	596917	7496.68	18	-	611910	600300
B	1.64494E+06	11650.1	17	-	1.66824E+06	1.636E+06
FE	1017.91	54.3438	15	-	1126.6	1000
FE	1976.16	81.9249	19	-	2140.01	2040
FE	2900.06	93.5524	18	-	3087.16	2940
CU	1073.35	54.9024	17	-	1183.15	1140
CU	2105.08	52.8467	18	-	2210.77	2140
CU	3186.72	80.313	15	-	3347.35	3160
HI	1023.17	26.8817	10	-	1076.94	1060
HI	2061.29	38.9127	11	-	2139.11	2090
HI	3044.08	149.832	12	-	3343.75	3150
CR	991.06	61.6459	10	-	1114.35	810
CR	1910.41	161.475	12	-	2233.36	1610
CR	2958.45	186.569	11	-	3331.59	2310
AIN	596.811	44.2216	9	-	685.254	533.3
AIN	112.915	9.42891	10	-	131.773	127
AIN	177.7	14.1985	9	-	206.097	173.3
HY	100.592	4.68925	13	-	109.971	101
HY	20.5538	.916166	13	-	22.3862	20
HY	50.55	1.88076	12	-	54.3115	50.3

**

**

**

MERT

**

ELEC

ITIR

AA

AA

AA

AA

ELEC

SPEC

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	9.64667	.352271	15	-	10.3512	10.1
CL	69.8646	3.6778	24	-	77.2202	73.7
CL	29.9444	1.74195	18	-	33.4283	30.2
FL	10.0587	.649237	15	-	11.3571	9.2
FL	72.3133	2.59143	15	-	77.4962	72.2
FL	30.3444	3.78747	18	-	37.9194	26.1
B	1.00262E+06	7309.1	13	-	1.01723E+06	1.042E+06
B	2.98458E+06	37483.4	18	-	3.05955E+06	3.006E+06
B	4.93482E+06	34950.4	17	-	5.00472E+06	4.752E+06
FE	1017.91	54.3438	15	-	1126.6	1040
FE	1976.16	81.9249	19	-	2140.01	2010
FE	2900.06	93.5524	18	-	3087.16	3020
CU	1073.35	54.9024	17	-	1183.15	1030
CU	2105.08	52.8467	18	-	2210.77	1910
CU	3186.72	80.313	15	-	3347.35	2870
AMM	1193.62	88.4432	9	-	1370.51	1247
AMM	112.915	9.42891	10	-	131.773	113.3
AMM	355.4	28.397	9	-	412.194	370
HY	100.592	4.68925	13	-	109.971	103.7
HY	20.5538	.916166	13	-	22.3862	21.7
HY	50.55	1.88076	12	-	54.3115	49.3

* OUTLIER

RECORD: 21 Code 17

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	19.2933	.704543	15	17.8842	20.7024	**
CL	69.8646	3.6778	24	62.509	77.2202	72
CL	29.9444	1.74195	18	26.4606	33.4283	**
						SLEC
B	100262	730.91	13	98799.7	101723	99100
B	59691.7	749.668	18	58192.3	61191	59700
B	246741	1747.52	17	243246	250236	241700
						**
FE	1017.91	54.3438	15	909.221	1126.6	1010
FE	1976.16	81.9249	19	1812.31	2140.01	1950
FE	2900.06	93.5524	18	2712.95	3087.16	2930
						TTR
CU	1073.35	54.9024	17	963.541	1183.15	1020
CU	2105.08	52.8467	18	1999.39	2210.77	2090
CU	3186.72	80.313	15	3026.09	3347.35	2930
						**
NI	1023.17	26.8817	10	969.411	1076.94	1000
NI	2061.29	38.9127	11	1983.46	2139.11	2060
NI	3044.08	149.832	12	2744.42	3343.75	3100
						AA
CR	991.06	61.6459	10	867.769	1114.35	900
CR	1910.41	161.475	12	1587.46	2233.36	1850
CR	2958.45	186.569	11	2585.32	3331.59	2770
						AA

* OUTLIER

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	9.64667	.352271	15	-	10.3512	9.8
CL	69.8646	3.6778	24	-	77.2202	72.4
CL	29.9444	1.74195	18	-	33.4283	28.7
FL	40.2347	2.59695	15	-	45.4286	37.7
FL	72.3133	2.59143	15	-	77.4962	73.7
FL	30.3444	3.78747	18	-	37.9194	27.7
B	1.00262E+06	7309.1	13	-	1.01723E+06	1.029E+06
B	2.98458E+06	37483.4	18	-	3.05955E+06	3.002E+06
B	4.93482E+06	34950.4	17	-	5.00472E+06	4.952E+06
FE	814.327	43.475	15	-	901.277	790
FE	1580.93	65.5399	19	-	1712.01	1640
FE	1160.02	37.421	18	-	1234.86	1210
CU	858.677	43.9219	17	-	946.521	890
CU	1684.06	42.2773	18	-	1768.62	1730
CU	1274.69	32.1252	15	-	1338.94	1320
NI	818.539	21.5054	10	-	861.55	830
NI	1649.03	31.1302	11	-	1711.29	1680
NI	1217.63	59.9327	12	-	1337.5	1290
CR	792.848	49.3167	10	-	891.482	860
CR	1528.33	129.18	12	-	1786.68	1670
CR	1183.38	74.6275	11	-	1332.64	1270
AMM	596.811	44.2216	9	-	685.254	597
AMM	225.83	18.8578	10	-	263.546	238
AMM	355.4	28.397	9	-	412.194	340
HY	100.592	4.68925	13	-	109.971	108
HY	20.5538	.916166	13	-	22.3862	21
HY	50.55	1.88076	12	-	54.3115	53.7

AA OUTLIER

RECORD: 23 Code 18

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	19.2933	.704543	15	-	20.7024	18.4
CL	69.8646	3.6778	24	-	77.2202	64.6
CL	29.9444	1.74195	18	-	33.4283	23.8
						ELEC
B	1.00262E+06	7309.1	11	-	1.01723E+06	1.017E+06
B	2.98458E+06	37483.4	11	-	3.05955E+06	3E+06
B	4.93482E+06	34950.4	11	-	5.00472E+06	4.93E+06
FE	254.477	13.586	15	-	281.649	223
FE	494.039	20.4812	19	-	535.002	469
FE	725.014	23.3881	18	-	771.79	725
						TITR
CU	268.337	13.7256	17	-	295.780	257
CU	526.27	13.2117	18	-	552.694	508
CU	796.68	20.0782	15	-	836.826	782
						DCPL
HI	255.794	6.72043	10	-	269.234	249
HI	515.322	9.72818	11	-	534.779	497
HI	761.021	37.4579	12	-	835.927	748
						DCPL
CR	247.765	15.4115	10	-	278.588	241
CR	477.602	40.3686	12	-	558.339	476
CR	739.614	46.6422	11	-	832.898	766
						DCPL

* OUTLIER

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	19.2933	.704543	15	-	17.8842	27
CL	69.8646	3.6778	24	-	62.509	54.7
CL	29.9444	1.74195	18	-	26.4606	30.7
						ELEC
FL	100.587	6.49237	15	-	87.6019	105.7
FL	144.627	5.18286	15	-	134.261	146
FL	60.6389	7.57495	18	-	45.539	53
						ELEC
B	1.00262E+06	7305.1	13	-	987997	1.008E+06
B	2.98458E+06	37483.4	18	-	2.90962E+06	3E+06
B	4.93482E+06	34950.4	17	-	4.86492E+06	4.946E+06
						TITR
FE	5.08954	.271719	15	-	4.5461	5.6
FE	9.88078	.409625	19	-	9.06154	9.64
FE	14.5003	.467762	18	-	13.5648	14.87
						AAGF
CU	5.36673	.274512	17	-	4.81771	4.61
CU	10.5254	.264233	18	-	9.99694	10.86
CU	15.9336	.401565	15	-	15.1305	15.5
						AAGF
AIH	596.811	44.2216	9	-	508.368	670
AIH	112.915	9.42891	10	-	94.0572	120
AIH	355.4	28.397	9	-	298.606	410
						SPEC
HY	100.592	4.68923	13	-	91.2138	101
HY	20.5538	.916166	13	-	18.7215	20.7
HY	50.55	1.88076	12	-	46.7885	49.7
						SPEC

A OUTLIER

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	19.2933	.704543	15	-	20.7024	18.8
CL	69.8646	3.6778	24	-	77.2202	69.2
CL	29.9444	1.74195	18	-	33.4283	30.7
FL	20.1173	1.29847	15	-	22.7143	19.6
FL	72.3133	2.59143	15	-	77.4962	68.1
FL	36.3444	3.78747	18	-	37.9194	36.5
B	1.00262E+06	7309.1	13	-	1.01723E+06	1.009E+06
B	2.98458E+06	37483.4	18	-	3.05955E+06	2.977E+06
B	4.93482E+06	34950.4	17	-	5.00472E+06	4.945E+06
FE	304.763	16.2706	15	-	337.304	330
FE	591.664	24.5284	19	-	640.721	630
FE	868.28	28.0097	18	-	924.299	950
CU	321.361	16.4379	17	-	354.237	320
CU	630.264	15.8224	18	-	661.908	620
CU	954.108	24.0458	15	-	1002.2	920
AMM	1193.62	88.4432	9	-	1370.51	1223
AMM	112.215	9.42891	10	-	431.733	337
AMM	35.24	28.397	9	-	412.194	
HY	100.592	4.68925	13	-	109.971	94.7
HY	20.5538	.916166	13	-	22.3862	16.7
HY	50.55	1.88076	12	-	54.3115	42

OUTLIER

REPORT: 26 Code 5

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	9.64667	.352271	15	8.94212	10.3512	9.7
CL	69.8646	3.6778	24	62.509	77.2202	70.9
CL	29.9444	1.74195	18	26.4606	33.4283	30.9
FL	50.2933	3.24619	15	43.801	56.7857	47.3
FL	144.627	5.18286	15	134.261	154.992	126
FL	60.6889	7.57495	18	45.539	75.8388	57
						IC
						**
						ELEC
B	2.98458E+06	37483.4	18	2.90962E+06	3.05955E+06	3.054E+06
B	4.93482E+06	34950.4	17	4.86492E+06	5.00472E+06	4.99E+06
FE	610.99	32.6193	15	545.731	676.228	608
FE	591.664	24.5284	19	542.607	640.721	617
FE	483.343	15.5921	18	452.158	514.527	469
						TITR
CU	644.265	32.9546	17	578.256	710.175	627
CU	630.264	15.8224	18	598.619	661.908	624
CU	531.12	13.3855	15	504.349	557.893	461
						PLAS
NI	614.15	16.1355	10	581.879	646.421	640
NI	617.152	11.6505	11	593.851	640.453	627
NI	507.347	24.972	12	457.403	557.293	470
						PLAS
CR	594.874	37.0023	10	520.87	668.879	594
CR	571.578	48.3457	12	475.287	668.669	607
CR	493.076	31.0948	11	430.886	555.265	451
						PLAS
AMM	1193.62	88.4432	9	1016.74	1370.51	1215
AMM	225.83	18.8578	10	188.114	263.546	231.7
AMM	355.4	28.397	9	298.606	412.194	310
HY	100.592	4.68925	13	91.2138	109.971	93.7
HY	20.5538	.916166	13	18.7215	22.3862	19
HY	50.55	1.88076	12	46.7885	54.3115	47.7
						SPEC
						SPEC

AA OUTLIER

RECORD: 27 Code 29

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	9.64667	.352271	15	8.94212	22.7	**
CL	69.8646	3.6778	24	62.509	76	
CL	29.9444	1.74195	18	26.4606	52	**
						ELEC
FL	40.2347	2.59695	15	35.0408	43	
FL	72.3133	2.59143	15	67.1305	78.7	**
FL	30.3444	3.78747	18	22.7695	36	
						ELEC

* OUTLIER

RECORD: 28 Code 32

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	4.82333	.176136	15	4.47106	5.59	★
CL	6.98646	.36778	24	6.2509	3.29	★
CL	2.99444	.174195	18	2.64606	7.1	★

IC

★ ★ OUTLIER

RECORD: 29 Code 11

ANALYTE	MEAN	STD DEV	N	STD DEV	RANGE	REPORTED VALUE
CL	69.8646	3.6778	24	62.509	-	74
CL	29.9444	1.74195	18	26.4606	-	37
						**
						ELEC

** OUTLIER

RECORD: 30

Code 3

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
CL	19.2933	.704543	15	-	26	**
CL	69.8646	3.6778	24	-	76.3	**
CL	29.5444	1.74195	18	-	38.3	**

SPEC

** OUTLIER

RECORD: 31 Code 22

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE
CL	19.2933	.704543	15	-	21.7
CL	69.8646	3.6778	24	-	65.9
CL	29.9444	1.74195	18	-	28.9

**

IC

** OUTLIER

RECORD: 32 Code 6

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE
CL	9.64667	.352271	15	-	9.7
CL	69.8646	3.6778	24	-	70.9
CL	29.9444	1.74195	18	-	30.9

IC

AA OUTLIER

RECORD: 33 Code 13

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE
FL	40.2347	2.59695	15	-	37.7
FL	72.3133	2.59143	15	-	73.7
FL	30.3444	3.78747	18	-	27.7

* OUTLIER

RECORD: 34 Code 4

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE
CL	9.64667	.352271	15	-	10
CL	69.8646	3.6778	24	-	74
CL	29.9444	1.74195	18	-	24.3
					**
					ELEC
FL	72.3133	2.59143	15	-	76.3
FL	30.3444	3.78747	18	-	27.3

OUTLIER

RECORD: 35 (GML)

ANALYTE	MEAN	STD DEV	N	RANGE	REPORTED VALUE	
ACL	9.64667	.352271	15	8.94212	10.3512	10.3
RCL	69.8646	3.6778	24	62.509	77.2202	69.7
CCL	29.9444	1.74195	18	26.4606	33.4283	27.7
						ELEC
AFL	100.587	6.49237	15	87.6019	113.571	96.1
RFL	144.627	5.18286	15	134.261	154.992	149
CFL	303.444	37.8747	18	227.695	379.194	329
						ELEC
DR	1.00262E+06	7309.1	13	987997	1.01723E+06	1.014E+06
EB	2.98458E+06	37483.4	18	2.90962E+06	3.05955E+06	3.047E+06
FB	4.93482E+06	34950.4	17	4.86492E+06	5.00472E+06	5.04E+06
						**
GFE	1272.39	67.9298	15	1136.53	1408.25	1280
HFE	2470.2	102.406	19	2265.38	2675.01	2390
IFE	3625.07	116.941	18	3391.19	3858.95	3430
						TITR
CCU	1341.68	68.628	17	1204.43	1478.94	1330
HCU	2631.35	66.0583	18	2499.23	2763.47	2600
ICU	3983.4	100.391	15	3782.62	4184.18	3840
						AA
CNI	1278.97	33.6022	10	1211.76	1346.17	1320
HNI	2576.61	48.6409	11	2479.33	2673.89	2580
INI	3805.1	187.29	12	3430.53	4179.68	3790
						AA
GCR	1238.83	77.0573	10	1084.71	1392.94	1200
HCR	2388.01	201.843	12	1984.32	2791.7	2690
ICR	3698.07	233.211	11	3231.65	4164.49	3740
						AA
HAMH	1193.62	88.4432	9	1016.74	1370.51	1166.9
HAMH	112.915	9.42891	10	94.0572	131.773	119.7
OAMH	355.4	28.397	9	298.606	412.194	355
						ELEC
PHY	100.592	4.68925	13	91.2138	109.971	100
QUY	20.5538	.916166	13	18.7215	22.3862	19.3
RHY	50.55	1.88076	12	46.7885	54.3115	52.4
						SPEC