

Duquesne Light Company

Beaver Valley Power Station
P.O. Box 4
Shippingport, PA 15077-0004

JOHN D. SIEBER
Vice President - Nuclear Group

(412) 643-5255

August 14, 1989

U. S. Nuclear Regulatory Commission
Attn: Document Control Desk
Washington, DC 20555

Reference: Beaver Valley Power Station, Unit No. 1 and No. 2
BV-1 Docket No. 50-334, License No. DPR-66
BV-2 Docket No. 50-412, License No. NPF-73
Proposed Operating License Change Request No. 161/21
(Revision 1)


Gentlemen:

Enclosed are:

1. An Application for Amendment to the above operating licenses to revise the Reactor Coolant System leakage detection Technical Specification 3.4.6.1. The proposed amendment would provide an action statement for operation with both the containment particulate and gaseous radiation monitors inoperable. This amendment would also revise the channel calibration surveillance frequency for the containment particulate and gaseous radiation monitors in Table 4.3-3 of the Technical Specifications.
2. Documents designated Attachments A, B and C which set forth the change in the Technical Specifications, the safety analysis and the no significant hazards evaluation.

This change request supersedes our Operating License Change Request No. 161/21 submitted to you on January 12, 1989. This proposed change was reviewed by the OSC and ORC and determined to be safe and does not involve an unreviewed safety question.

Very truly yours,


J. D. Sieber
Vice President
Nuclear Group

8908240022 890314
PDR ADOCK 050-0334
P FDC

Aoo/
1/1
Add: NRE/DREP/RPB Lt. Encl
1 1

Beaver Valley Power Station, Unit Nos. 1 & 2
Docket No. 50-334, License No. DPR-66
Docket No. 50-412, License No. NPF-73
Page 2

cc: Mr. J. Beall, Sr. Resident Inspector
Mr. W. T. Russell, NRC Region I Administrator
Mr. P. Tam, Sr. Project Manager
Mr. T. M. Gerusky, Director BRP/DER
Mr. R. Janati, BRP/DER
Mr. R. Saunders (VEPCO)