

10 FEB 1988

License No. 20-02804-01
Docket No. 030-04625
Control No. 108258

EG&G, Inc.
ATTN: Phillip Ayers
45 William Street
Wellesley, Massachusetts 02181

Gentlemen:

Please find enclosed an amendment to your NRC Material License.

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify the Region I Material Licensing Section, (215) 337-5239, so that we can provide appropriate corrections and answers.

Please be advised that you must conduct your program involving licensed radioactive materials in accordance with the conditions of your NRC license, representations made in your license application, and NRC regulations. In particular, please note the items in the enclosed, "Requirements for Materials Licensees."

Please note that in your Radiological Safety Manual, Section 3.8, Procedures for taking air samples, page 16, you state to make a mental note of the time and run sampler for 2 hours. 10 CFR 20.401(b) requires you to maintain records of surveys performed. Since the time duration that the air sampler was used is an integral part of the calculations for air concentrations, it will be necessary to record the time when you started the air sampler and the time when you turned it off to determine the duration that it was in operation to ascertain the volume of air drawn by the air sampler. In addition, in the Manufacturing Manual for Nickel-63, Figure 14, Inspection Criteria For Factory Final and Q.C. Acceptance Inspections, (Part No. QC-025), Section 8.2, you must have a typographical error in that background radiation is not 0.10 R/hr (100 mr/hr). Please make the necessary corrections to your manuals and inform your personnel of these changes.

Since serious consequences to employees and the public can result from failure to comply with NRC requirements, the NRC expects licensees to pay meticulous attention to detail and to achieve the high standard of compliance which the NRC expects of its licensees.

You will be periodically inspected by NRC. A fee may be charged for inspections in accordance with 10 CFR Part 170. Failure to conduct your program safely and in accordance with NRC regulations, license conditions, and representations made in your license application and supplemental correspondence

OFFICIAL RECORD COPY

ML 20-02804-01/LTR - 0001.0.0
02/08/88

8903150196 880210
REG1 LIC30
20-02804-01 PNU

ML10

10 FEB 1988

with NRC will result in prompt and vigorous enforcement action against you. This could include issuance of a notice of violation, or in case of serious violations, an imposition of a civil penalty or an order suspending, modifying or revoking your license as specified in the General Policy and Procedures for NRC Enforcement Actions, 10 CFR Part 2, Appendix C.

We wish you success in operating a safe and effective licensed program.

Sincerely,

Original Signed By:
Jack Davis

Jack Davis
Nuclear Materials Safety Section B
Division of Radiation Safety
and Safeguards

Enclosures:

1. Amendment No. 28
2. Requirements for Materials Licensees

DRSS:RI
Davis/mjh

02/9/88

OFFICIAL RECORD COPY

ML 20-02804-01/LTR - 0002.0.0
02/08/88