

07 APR 1988

Docket No. 030-30494  
Control No. 108584

Vermont Testing  
ATTN: Douglas E. Williams  
RD #2, Box 400 Kneeland Flats  
Waterbury, Vermont 05676

Gentlemen:

This is in reference to your application dated March 7, 1988 for a byproduct material license. In order to continue our review, we need the following additional information:

1. In item 9. of your application, you state that the business is operating out your home. However, you did not specify if there are any zoning restrictions for having hazardous material stored in a residential area. Please provide a statement from the Zoning Board that there are no restrictions for having radioactive material stored in the driveway of your residence.
2. In item 9. of your application, you state that you will be using your vehicle for storage. However, you did not specify what you will do with the licensed material when the vehicle needs servicing or if the vehicle is involved in an accident and needs to be repaired. Please provide alternative storage for the licensed material when the vehicle is out of commission.
3. In item 9. of your application, you state the vehicle will be locked and stored in your driveway. It has been the Commission's experience with this type of gauging equipment that they have been removed from the vehicle or that the entire vehicle has been removed by unauthorized means. However, we have never had a gauge secured in a storage facility be removed by unauthorized means. Please specify how you will prevent the removal of the gauge or the removal of the vehicle by unauthorized means.
4. In your application, you did not address the caution signs that you will post on your vehicle during storage. 10 CFR Part 20.203(e)(1) requires you to post each area in which license material is stored. Please confirm that the vehicle will be so posted.

We will continue our review upon receipt of this information. Please reply in duplicate to my attention at the Region 1 office and refer to Mail Control No. 108584.

OFFICIAL RECORD COPY

ML 318 DAVIS - 0001.0.0  
04/06/88

8903150708 880504  
REG1 LIC30  
44-28180-01 PNU

ML10,

If we do not receive a reply from you within 30 calendar days from the date of this letter, we shall assume that you do not wish to pursue your application.

Sincerely,


Original Signed By  
Jack Davis

Jack Davis  
Nuclear Materials Safety Section B  
Division of Radiation Safety  
and Safeguards

Enclosures:

1. 10 CFR Part 20
2. Regulatory Guides FC 407

RI:DRSS  
Davis/mjh

04/7/88 

OFFICIAL RECORD COPY

ML 318 DAVIS - 0002.0.0  
04/06/88