



GOVERNMENT OF THE DISTRICT OF COLUMBIA
DEPARTMENT OF HUMAN SERVICES
WASHINGTON, D.C. 20002

COMMISSION ON MENTAL HEALTH SERVICES
ADULT SERVICE ADMINISTRATION
Saint Elizabeths Hospital
2700 MLK Jr., Avenue, S.E.
Washington, D.C. 20032

MS-16
P-2

IN REPLY REFER TO:

April 20, 1988

Nuclear Materials Safety Section B
Division of Radiation Safety and Safeguards
U.S. Nuclear Regulatory Commission
475 Allendale Road
King of Prussia, PA 19406

RE: Control 030-30295, 108022

Gentlemen:

The following information is submitted in response to the above referenced control:

1. All meters will be checked with a dedicated check source.
2. A meter will always be on site, including those times when meters are sent for calibration.

If you have any questions or require additional information, please do not hesitate to contact the undersigned.

Sincerely,

Robert A. Washington, Ph.D.
Commissioner

8903160242 880523
REG1 LIC30
08-28125-01 PNU

"OFFICIAL RECORD COPY"

ML10

108022
4/25/88