



Geotechnical & Material Testing, Inc.

- Geotechnical Engineering
- Construction Quality Control
- Construction Material Testing
- Pavement Design & Recycling

March 17, 1989

United States Nuclear Regulatory Commission
Region II
101 Marietta Street, N.W.
Atlanta, Georgia 30323

Attention: Mr. William E. Cline
Chief Nuclear Materials Safety and Safeguards
Branch
Division of Radiation Safety and Safeguards

Reference: REPLY TO A NOTICE OF VIOLATION
Docket No.: 030-20213
License No.: 45-23048-01
NRC Inspection Report No.: 45-23048-01/89-01

Gentlemen:

Your inspection of our activities on February 6, 1989, indicated a certain activity which violated NRC requirements. The violation was presented in your letter dated March 6, 1989, and its enclosure "Notice of Violation".

Our detailed response to the Notice of Violation is presented hereunder:

VIOLATION NO. 1

1. This violation is admitted.
2. The reason for this violation was no knowledge that the area adjacent to the nuclear gauge storage room was radiating an excessive amount of radiation.
3. Mr. Eugene M. Davis (Radiation Safety Officer) has ordered lead for the storage room which arrived on Wednesday, March 15, 1989. Arrangements are being made to have the lead installed.
4. Corrective steps as described above were taken after the NRC inspection in February.

8903240197 890317
REG2 LIC30
45-23048-01 PDC

703-631-2050

13944 B Willard Road • Chantilly, Virginia 22021
FAX 968-2930

IE07
1/0


United States Regulatory Commission
G.M.T.I. License No.: 45-23048-01

Page No. 2
March 17, 1989

5. The installation of the lead will be achieved within one (1) to two (2) weeks within the date of this letter.

We believe to the best of our knowledge that all corrective steps were taken and we are in full compliance with NRC requirements.

Respectfully submitted,
G.M.T., Incorporated



Ahmed Elrefai, Ph.D., P.E.
President

AE/mp

cc: Nuclear Regulatory Commission
Attn: Document Control Desk
Washington, D.C. 20555