

3M Medical Department
Medicine
Health Physics
Industrial Hygiene
Toxicology

Building 220-2E-02, 3M Center
St. Paul, Minnesota 55144-1000
612/733 1110

30-14999

DCB

3M

June 19, 1989

U.S. Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, IL 60137

Attention: D. J. Sreniawski, Chief
Nuclear Materials Safety Section 2

Subject: Notice of Violation, License No. 22-00057-61


Gentlemen:

This letter is in response to the Notice of Violation dated June 12, 1989 for our Brookings, South Dakota gamma sterilizer operated under license 22-00057-61.

Immediately following the NRC's May 23, 1989 inspection, corrective measures were instituted to insure that the biweekly check of the water level float switch is performed on schedule. All plant preventative maintenance activities have been reviewed to insure that all maintenance functions are supervised by the appropriate personnel. All maintenance activities involving the sterilizer are now being approved by the plant Radiation Safety Officer.

The changes to the plant preventative maintenance tracking system are now complete. Full compliance was achieved on June 16, 1989.

Sincerely,


Robert G. Wissink
Manager, Health Physics Services

RGW:ckm

cc: E. M. Aho - Medical/Surgical - Brookings 01/036

8907270197 890619
REG3 LIC30
22-00057-61 PDC

1207
JUN 28 1989