

404 Fountain Street Albert Lea, Minnesota 56007

507/373-2384

June 3, 1988

Nuclear Regulatory Commission Region III 799 Roosevelt Road Glen Ellyn, IL 60137

Cobalt 60 Teletherapy Licensing Attn:

Dear Commission:

Enclosed please find two (2) copies of our Cobalt Teletherapy renewal application for NRC license #22-15252-02.

Also enclosed is the application fee of \$350.00.

Sincerely,

Teny Maxhimen by 2 an WRZOMY
PRESIDENT Terry Maxhimer

Executive Director

TM:ao

8904110247 880623 REG3 LIC30 22-15252-02 PN PNU

JUN 6 1988

CONTROL NO 8 - 5 7 4


- H. Refer to reference #030-11464, Amendment #03, deleting Dr. David J. Nelson M.D. because of retirement.
- I. Dr. Salma Mikhail, M.D.
- J. We have adopted the training program described in Appendix of Draft Regulatory Guide FC 414-4.
- K. We have established and agree to follow written procedure for personnel monitoring that include as requirements the criteria specified in Item 10.12 of Draft Regulatory Guide FC 414-4.
- L. There has been no change in the information previously submitted to the NRC regarding other aspects of the radiation protection program or teletherapy program.
- M.l.We continue to follow the guidelines set at the Radiation Safety Committee meeting dated November 26, 1986. Copy of the minutes attached.
  - 2. The Radiation Safety Committee duties to include teletherapy.
  - 3. We do maintain records of the membership of the Radiation Safety Committee.

June 3, 1988 Page 2

- N. ALARA program is followed as in Appendix J of Draft Regulatory Guile FC 414-4.
- O. There has been no source change.
- P. Refer to Reference #030-11464, Amendement #03, deleting Dr. David J. Nelson, M.D. because of retirement.

## REPORT OF TELETHERAPY TESTS AND SURVEYS

Licensee	Naeve Hospital Association
Address	404 Fountain St, Albert Lea, Mn 56007
License #	22-15252-02
	4/5/86
	3. Thomas Payne Pt
	TELETHERAPY TESTS
Yes 📈	The interlock on the door(s) to the teletherapy room was tested and found to function properly. When a door was opened with the source "ON", the source returned to the "OFF" position and could not be turned "ON" again until the door was closed and the system reset at the control panel.
Yes /2 .	The teletherapy source "ON-OFF" indicators, both at the source housing and on the teletherapy machine control panel, were tested and found to function properly.
Yes No /_/	The teletherapy treatment timing device was tested and found to be accurate and to return the source to the "OFF" position when the preset time elasped.
Yes No / /	Electrical and/or mechanical stops installed to limit the orientation of the teletherapy head with the source "ON" were tested and found to function properly. The limitations are:
NO /	Mechanical limit such that the unit
	can operate from vertical beam down
	can operate from vernical
	to 90 toward west wall.
	Guntry side
	west, 900 Limit
	wallcontrol NO. 8553 4 Vertical down



## RADIATION SAFETY COMMITTEE

The Radiation Safety Committee met November 26, 1986.

Members present: Dr. Salma Mikhail, M.D., Radiologist, Thomas Payne, PhD, Radiation Physicist, Sandy Norby.

A complete discussion was held regarding the Radiation Safety Committee responsibilities. Meetings will be held quarterly. Members will include the Radiologist and the Director of Radiology, the Risk Management and Safety Director, nursing representative, administrative representative. Dr. Wogensen, Pathologist, will also be asked to serve on the committee.

Dr. Payne reviewed our nuclear medicine requirements and paper work and made some suggestions in that regard.

Dr. Payne, also, reviewed our entire cobalt therapy procedures, paper work, etc.

All radiation badge readings were reviewed and are in acceptable ranges.

There have been no accidents or radiation problems.

Submitted by

Bill Riche

BR:ws