

IN REPLY REFER TO
OIR:LFO

Oak Ridge, Tennessee
December 13, 1955

Mr. Allen M. Goldstein
Isotopes Specialties Company
703 South Main Street
Burbank, California

*our Waste disposal service
letter of May 12, 1955 and
November 23, 1955 pulled
from this file & placed in
auth. # 37803 on May 9, 1956
by J. N. M.*

Subject: MODIFICATION OF MANUFACTURER'S LEAK TESTING
REQUIREMENT

Dear Mr. Goldstein:

Our manufacturer's leak testing requirements for all beta emitting sources manufactured by Isotopes Specialties Company are hereby modified to permit such tests to be performed by wiping with a "Q" tip as stated on Application No. 37691 and related correspondence. This modification is made in consideration of information furnished by Isotopes Specialties Company to the effect that the aforementioned "Q" tip test is adequate to detect contamination or leakage in excess of 0.05 microcuries.

Very truly yours,

for E. Eugene Fowler, Chief
Allocations Branch
Isotopes Division

clinta/jc

NOT DISPATCHED BY MAIL AND RECORDS

9806030136 980520
PDR FOIA
FRIEDMAN98-19 PDR

A183

OFFICE ▶							
SURNAME ▶	L P Q	881					
DATE ▶							

9806030136