



UNITED STATES
NUCLEAR REGULATORY COMMISSION
WASHINGTON, D. C. 20555

JAN 20 1988


MS = 20
PS
20-17131-01

MEMORANDUM FOR: John E. Glenn, Chief
Nuclear Materials Safety
Section B, Region I

FROM: Vandy L. Miller, Chief
Medical, Academic, and
Commercial Use Safety Branch, IMNS

SUBJECT: TECHNICAL ASSISTANCE REQUEST REGARDING
BRIGHAM AND WOMEN'S HOSPITAL (20-17131-01)

The applicant requested an exemption from the requirement to label syringes pursuant to 10 CFR 35.60(b). The request should be denied. This matter was discussed in the analysis of comments for the final 10 CFR Part 35 rulemaking.


Vandy L. Miller, Chief
Medical, Academic, and
Commercial Use Safety Branch, IMNS

"OFFICIAL RECORD COPY"

ML10

8904130334 880422
REG1 LIC30
20-17131-01 PNU

Rec'd. 29 JAN 1988

Licensee: Brigham & Women's Hospital

License No.: 20-17131-01

TECHNICAL ASSISTANCE REQUEST

DATE: **19 OCT 1987**

TO: Vandy Miller, Chief, Material Licensing Branch, NMSS
, Chief, Material Certification and Procedures
Branch, NMSS

FROM: John E. Glenn, Chief, Nuclear Materials Safety Section B,
Region I

SUBJECT: REQUEST FOR TECHNICAL ASSISTANCE

XXX Control No. 107696 (enclosed)

XXX Letter dated 6/12/87 & 7/21/87 (enclosed)

Suggested change in licensing procedure (enclosed)

Specific ex. 35.19 Other (see remarks)

Requested action: Custom source/device review
 Review and comment
 Provide policy guidance
 Other (see remarks)

Remarks: Per 10 CFR 35.19 licensee requests specific exemption from syringe
labeling required by 10 CFR 35.60(b). Licensee submitted an entire new radiation
safety program therefore will be subject to new Part 35.

To be completed by NMSS

Date:

The above request has been received by MLB/MCPB and assigned to
(name)

. Please contact this individual for a status report if a
(phone number)

response is not received by .
(date)

Signature: FCML/FCMC Branch Chief

"OFFICIAL RECORD COPY"

ML10