



1200 JD Anderson Drive
Morgantown, WV 26505
304-598-1200
monhealth.com

U.S. Nuclear regulatory Commission
Materials Licensing Branch
Region I
2100 Renaissance Blvd.
King of Prussia, PA 19406

RE: Reply to Notice of Violation
Radioactive Material License Nos. 47-16259-01
Monongalia General Hospital DBA Mon Health Medical Center
Inspection Number: 03010683/2020001
Docket No. 030-10683

To Whom It May Concern:

In response, we offer the following:

A. The violation was the result of a miscommunication. The RSO was unaware of the new sources when performing the semi-annual inventory. The sources were secured at all times. The inventory was updated with the new sources added. Compliance was met on August 4, 2020

B. The violation was the result of the staff starting a new process of using the computer to log everything. The computer provides no visual reminder for the task and one member of the staff did not remember to do the survey on two days. It should be noted that only 35.200 materials were used on those days; no materials in 35.300 were used on days the surveys were missed. We have gone back to using paper to record surveys. It is harder to forget to do when the log is out. We have not missed any further surveys. Compliance was met following the survey of July 9, 2020.

If you have any questions, please contact: RSO Mark Perna, M.S. at (412) 551-9259.

Sincerely,

A handwritten signature in black ink, appearing to read "Mark Perna", written over a horizontal line.

Administrator