

INSTROTEK® COMPANIES



INSTROTEK • CPN • HMA • RAINHART

InstroTek Inc.
1 Triangle Drive
P.O. Box 13944
Research Triangle Park, NC 27709
(919) 875-8371 Fax: (919) 875-8328

Chris Morris
Terradon (Poca)
409 Jacobson Drive
Poca, WV 25159

Phone: 304.755.8291
Fax: 304.755.2636

LEAK TEST CERTIFICATE InstroTek License #092-1073-1


This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model: 3430PLUS
Gauge Serial Number: 74192

Date Swabbed: 11/19/2020

Source Serial Number	Reading in MicroCuries
NA	0.000002
NA	0.000226

Note: 0.005 MicroCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: 
Customer Signature: _____

Print Date: 11/23/2020
Date: _____

CPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 8 mCi Cs-137. Refer to the manual for other products.

INSTROTEK® COMPANIES



INSTROTEK • CPN • HMA • RAINHART

InstroTek Inc.
1 Triangle Drive
P.O. Box 13944
Research Triangle Park, NC 27709
(919) 875-8371 Fax: (919) 875-8328

Regina Francis
Terradon
401 Jacobson Drive
Poca, WV 25159

Phone: 304.755.8291
Fax: 304.755.2636

LEAK TEST CERTIFICATE InstroTek License #092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model: 3430
Gauge Serial Number: 32766

Date Swabbed: 10/13/2020

Source Serial Number	Reading in MicroCuries
47-28963	0.000002
750-8112	0.000223

Note: 0.005 MicroCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: _____
Customer Signature: _____

Print Date: 10/19/2020
Date: _____

CPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 8 mCi Cs-137. Refer to the manual for other products.

INSTROTEK® COMPANIES



INSTROTEK • CPN • HMA • RAINHART

InstroTek Inc.
1 Triangle Drive
Research Triangle Park, NC 27709
(919) 875-8371 Fax: (919) 875-8328

Chris Morris
Terradon (Poca)
409 Jacobson Drive
Poca, WV 25159

Phone: 304.755.8291
Fax: 304.755.2636

LEAK TEST CERTIFICATE InstroTek License #092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model: 3430
Gauge Serial Number: 73250

Test Date: 9/29/2020

Source Serial Number	Reading in MicroCuries
78-11482	0.000002
77-16190	0.000309

Note: 0.005 MicroCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: _____
Customer Signature: _____

Print Date: 10/5/2020
Date: _____

CPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 8 mCi Cs-137. Refer to the manual for other products.

INSTROTEK® COMPANIES



INSTROTEK • CPN • HMA • RAINHART

InstroTek Inc.
1 Triangle Drive
Research Triangle Park, NC 27709
(919) 875-8371 Fax: (919) 875-8328

Chris Morris
Terradon (Poca)
409 Jacobson Drive
Poca, WV 25159

Phone: 304.755.8291
Fax: 304.755.2636

LEAK TEST CERTIFICATE InstroTek License #092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model: 3430 Plus Test Date: 9/18/2020
Gauge Serial Number: 73249

Source Serial Number	Reading in MicroCuries
78-11481	0.000000
77-16189	0.000234

Note: 0.005 MicroCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: _____
Customer Signature: _____

Print Date: 9/25/2020
Date: _____

CPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 8 mCi Cs-137. Refer to the manual for other products.