



WESTERN MATERIALS TESTING, LLC
Materials Testing Laboratory
1401 Gold Avenue, Unit B
Bozeman, MT 59715

September 20, 2020

SUBJECT: NRC INSPECTION REPORT 030-39056/2019-001; AND NOTICE OF VIOLATION

The violation referenced in the above report has been fixed (February 10-12, 2020).

The lab I own, and operate also leases space to another company the owns and operates portable nuclear density gauges. When they obtained their license to store gauges in the lab, we stored the gauges in the same cabinet, mistakenly.

The cabinet has since been divided in to 2 completely separate cabinets.

The Top Cabinet: CTL Thompson Property- Only accessible by their employees, 2 independent locks, the keys only possessed by their(CTL Thompson) employees.

The Bottom Cabinet: Westerns Materials Testing Property- Only accessible by Western Materials Testing, 2 independent locks, the keys only possessed by Western Materials Testing employees.

I can provide pictures, and will verify during the next annual audit.

Please let me know if there are any other questions regarding this issue.

Thanks

Craig Paulson- Owner

Western Materials Testing

Bozeman, Montana

Hard Copies mailed to:

U.S. Nuclear Regulatory Commission

Washington, DC 20555-0001

ATTN: Document Control Desk

U.S. Nuclear Regulatory Commission

Regional Administrator, Region IV

1600 E. Lamar Blvd.

Arlington, Texas 76011

*IED7
RGN-IV*