



**UNITED STATES
NUCLEAR REGULATORY COMMISSION**
REGION I
2100 RENAISSANCE BLVD.
KING OF PRUSSIA, PA 19406-2713

January 7, 2021

Bette W. Blankenship, M.S., DABR, FAAPM
Radiation Safety Officer
The Hospital of Central Connecticut
100 Grand Street
New Britain, CT 06050

SUBJECT: THE HOSPITAL OF CENTRAL CONNECTICUT HOSPITAL, EXEMPTION TO
20.1701 NOT REQUIRED, MAIL CONTROL NO. 624088

Dear Ms. Blankenship:

This is in reference to your letter dated December 23, 2020 requesting an exemption to the regulations in 10 CFR 20.1701. After discussions with several staff members, we have concluded that the requested exemption is not required. We believe that you may revise your established survey procedures to allow periodic, annually when possible, verifications of air flow to confirm DACs are not exceeded.

Your cooperation is appreciated.

Sincerely,

Tara L. Weidner, Senior Health Physicist
Medical and Licensing Assistance Branch
Division of Nuclear Materials Safety
Region I

License No. 06-02388-01
Docket No. 03001250
Mail Control No. 624088

THE HOSPITAL OF CENTRAL CONNECTICUT, EXEMPTION TO 20.1701 NOT REQUIRED,
MAIL CONTROL NO. 624088 DATED JANUARY 7, 2021

DOCUMENT NAME: G:\WBL Documents\WBL License Cover Letter\L06-02388-01.624088.docx

SUNSI Review Complete: TLWeidner

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	RI:DNMS							
NAME	TLWeidner (TLW)							
DATE	1.7.21							

OFFICIAL RECORD COPY