



### CERTIFICATE OF DISPOSITION OF MATERIALS

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the FOIA, Privacy, and Information Collections Branch (T-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by e-mail to [Infocollections.Resource@nrc.gov](mailto:Infocollections.Resource@nrc.gov), and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

LICENSEE NAME AND ADDRESS  
*HollyFrontier Cheyenne Refining, LLC  
300 Morrie Avenue  
Cheyenne, WY. 82007*

LICENSE NUMBER  
*49-26892-01*

DOCKET NUMBER  
*030-30129*

LICENSE EXPIRATION DATE  
*12-31-2023*

#### A. LICENSE STATUS (Check the appropriate box)

- This license has expired.
- This license has not yet expired; please terminate it.

#### B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

- 1. No radioactive materials have ever been procured or possessed by the licensee under this license.
- 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner.

a. Transfer of radioactive materials to the licensee listed below:

*D757; H76 6@? E  
"#S!S" S#*

b. Disposal of radioactive materials:

1. Directly by the licensee:

2. By licensed disposal site:

*? S^5a` fcb^@g\_ TWI (\$8888  
6aU W@g\_ TWI, %%% #S+  
>[UWeW@g\_ TWI, &ZS\* +SZ #  
>[UWeW@S\_ W: a^k8cb` f[W5ZMW VDWX [ Y]>>5*

3. By waste contractor:

c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

#### C. SURVEYS PERFORMED AND REPORTED

- 1. A radiation survey was conducted by the licensee. The survey confirms:
  - a. the absence of licensed radioactive materials
  - b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.

2. A copy of the radiation survey results:

a. is attached; or  b. is not attached (Provide explanation); or  c. was forwarded to NRC on: \_\_\_\_\_ Date

3. A radiation survey is not required as only sealed sources were ever possessed under this license, and

a. The results of the latest leak test are attached; and/or  b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME	TITLE	TELEPHONE (Include Area Code)	E-MAIL ADDRESS
<i>Sarah Goodman</i>	<i>Safety Supervisor</i>	<i>307-771 8773</i>	<i>Sarah.Goodman@HollyFrontier.Com</i>

Mail all future correspondence regarding this license to: *HollyFrontier Cheyenne Refining, LLC  
300 Morrie Avenue  
Cheyenne, WY. 82007*

#### C. CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE	SIGNATURE	DATE
<i>James Brown / Safety Representative</i>	<i>James Brown</i>	<i>1-28-2021</i>

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

# Tracer Technologies International, Inc.

24600 Center Ridge Road, Suite 145  
Westlake, Ohio 44145  
Phone (440) 641-1133  
Fax (216) 464-6141

## CERTIFICATE OF LEAK TEST

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	2980CG
<b>Model #:</b>	GS400	<b>millicuries:</b>	30
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	vac tower top	<b>Test Number:</b>	
<b>Test Date:</b>	6/9/2020		

The total removable amount of activity was analyzed to be:

**0.0000 microcuries**

**Laboratory Analysis Number** AAU159      **Method:**     Wipe     Other

Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Dec-2020**.

Analyzed By:

Approved By:

Signature: WRL  
Name: Wallace R. Loder, Jr.  
Date: 19-Jun-2020

Signature: Wallace Loder Jr  
Name: Wallace R. Loder, Jr.  
Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.

It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072









**CERTIFICATE OF LEAK TEST**

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	76334
<b>Model #:</b>	SA-1 F37	<b>millicuries:</b>	500
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Coker, D704 middle	<b>Test Number:</b>	
<b>Test Date:</b>	6/9/2020		

**The total removable amount of activity was analyzed to be:**

**0.0000 microcuries**

**Laboratory Analysis Number** AAU163      **Method:**     Wipe     Other

Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Dec-2020**.

Analyzed By:

Approved By:

Signature: WRL  
Name: Wallace R. Loder, Jr.  
Date: 19-Jun-2020

Signature: Wallace Loder Jr  
Name: Wallace R. Loder, Jr.  
Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.  
It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072





**Tracer Technologies  
International, Inc.**

24600 Center Ridge Road, Suite 145  
Westlake, Ohio 44145  
Phone (440) 641-1133  
Fax (216) 464-6141

**C E R T I F I C A T E   O F   L E A K   T E S T**

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	76327
<b>Model #:</b>	SA-1 F37	<b>millicuries:</b>	500
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Coker, D705 top	<b>Test Number:</b>	
<b>Test Date:</b>	6/9/2020		

The total removable amount of activity was analyzed to be:

**0.0000 microcuries**

Laboratory Analysis Number AAU165      Method:  Wipe     Other


Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Dec-2020**.

Analyzed By:

Approved By:

Signature:



Name:

Wallace R. Loder, Jr.

Date:

19-Jun-2020

Signature:



Name:

Wallace R. Loder, Jr.

Title:

Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.

It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the

State of Ohio Department of Health

Bureau of Radiation Protection

License # 03225180072

**CERTIFICATE OF LEAK TEST**

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	M1452
<b>Model #:</b>	SA-1 F37	<b>millicuries:</b>	500
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Coker, D705 middle	<b>Test Number:</b>	
<b>Test Date:</b>	6/9/2020		

**The total removable amount of activity was analyzed to be:**

**0.0000 microcuries**

**Laboratory Analysis Number** AAU166      **Method:**     Wipe     Other

Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Dec-2020**.

Analyzed By:

Approved By:

Signature: \_\_\_\_\_



Name: Wallace R. Loder, Jr.  
Date: 19-Jun-2020

Signature: \_\_\_\_\_



Name: Wallace R. Loder, Jr.  
Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.  
It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072

**Tracer Technologies  
International, Inc.**

24600 Center Ridge Road, Suite 145  
Westlake, Ohio 44145  
Phone (440) 641-1133  
Fax (216) 464-6141

**CERTIFICATE OF LEAK TEST**

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	M1449
<b>Model #:</b>	SA-1 F37	<b>millicuries:</b>	500
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Coker, D705 bottom	<b>Test Number:</b>	
<b>Test Date:</b>	6/9/2020		

The total removable amount of activity was analyzed to be:

**0.0000 microcuries**

**Laboratory Analysis Number** AAU167      **Method:**     Wipe     Other

Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Dec-2020**.

Analyzed By:

Approved By:

Signature: WRL

Name: Wallace R. Loder, Jr.

Date: 19-Jun-2020

Signature: Wallace Loder Jr

Name: Wallace R. Loder, Jr.

Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.  
It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072

**Tracer Technologies  
International, Inc.**

24600 Center Ridge Road, Suite 145  
Westlake, Ohio 44145  
Phone (440) 641-1133  
Fax (216) 464-6141

**CERTIFICATE OF LEAK TEST**

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	M1450
<b>Model #:</b>	SA-1 F37	<b>millicuries:</b>	500
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Coker, D706 top	<b>Test Number:</b>	
<b>Test Date:</b>	6/9/2020		

The total removable amount of activity was analyzed to be:

**0.0000 microcuries**

**Laboratory Analysis Number** AAU168      **Method:** X Wipe    \_\_\_Other


\_\_\_ Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

X Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Dec-2020**.

Analyzed By:

Approved By:

Signature:



Name: Wallace R. Loder, Jr.  
Date: 19-Jun-2020

Signature:



Name: Wallace R. Loder, Jr.  
Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.  
It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072

## CERTIFICATE OF LEAK TEST

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	M1447
<b>Model #:</b>	SA-1 F37	<b>millicuries:</b>	500
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Coker, D706 middle	<b>Test Number:</b>	
<b>Test Date:</b>	6/9/2020		

**The total removable amount of activity was analyzed to be:**

**0.0000 microcuries**

**Laboratory Analysis Number** AAU169      **Method:**     Wipe     Other

Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Dec-2020**.

Analyzed By:

Approved By:

Signature: WRJ

Name: Wallace R. Loder, Jr.  
Date: 19-Jun-2020

Signature: Wallace Loder Jr.

Name: Wallace R. Loder, Jr.  
Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.  
It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072

# Tracer Technologies International, Inc.

24600 Center Ridge Road, Suite 145  
Westlake, Ohio 44145  
Phone (440) 641-1133  
Fax (216) 464-6141

## CERTIFICATE OF LEAK TEST

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	75291
<b>Model #:</b>	SA-1 F37	<b>millicuries:</b>	500
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Coker, D706 bottom	<b>Test Number:</b>	
<b>Test Date:</b>	6/9/2020		

**The total removable amount of activity was analyzed to be:**

**0.0000 microcuries**

**Laboratory Analysis Number** AAU170

**Method:**  Wipe  Other

\_\_\_Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Dec-2020**.

Analyzed By:

Approved By:

Signature:



Name: Wallace R. Loder, Jr.  
Date: 19-Jun-2020

Signature:



Name: Wallace R. Loder, Jr.  
Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.  
It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072

## CERTIFICATE OF LEAK TEST

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	M3366
<b>Model #:</b>	SA1	<b>millicuries:</b>	10
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Alkylation, Acid reg bot	<b>Test Number:</b>	
<b>Test Date:</b>	6/9/2020		

**The total removable amount of activity was analyzed to be:**

**0.0000 microcuries**

**Laboratory Analysis Number** AAU172      **Method:**     Wipe     Other

     Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Dec-2020**.

Analyzed By:

Approved By:

Signature:



Name: Wallace R. Loder, Jr.

Date: 19-Jun-2020

Signature:



Name: Wallace R. Loder, Jr.

Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.

It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072

# Tracer Technologies International, Inc.

24600 Center Ridge Road, Suite 145  
Westlake, Ohio 44145  
Phone (440) 641-1133  
Fax (216) 464-6141

## CERTIFICATE OF LEAK TEST

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	8653GG
<b>Model #:</b>	SA-15	<b>millicuries:</b>	100
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Alkylation, Acid storage top	<b>Test Number:</b>	
<b>Test Date:</b>	6/9/2020		

The total removable amount of activity was analyzed to be:

**0.0000 microcuries**

**Laboratory Analysis Number** AAU174      **Method:**    X Wipe    \_\_\_ Other

\_\_\_ Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

X Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Dec-2020**.

Analyzed By:

Approved By:

Signature: \_\_\_\_\_



Name: Wallace R. Loder, Jr.  
Date: 19-Jun-2020

Signature: \_\_\_\_\_



Name: Wallace R. Loder, Jr.  
Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.  
It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072



# Tracer Technologies International, Inc.

24600 Center Ridge Road, Suite 145  
Westlake, Ohio 44145  
Phone (440) 641-1133  
Fax (216) 464-6141

## CERTIFICATE OF LEAK TEST

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ohmart Vega	<b>Serial #:</b>	9843CM
<b>Model #:</b>	MDTSII	<b>millicuries:</b>	10
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Alkylation, mixer settler top	<b>Test Number:</b>	
<b>Test Date:</b>	6/9/2020		

The total removable amount of activity was analyzed to be:

**0.0000 microcuries**

**Laboratory Analysis Number** AAU173      **Method:**     Wipe     Other


Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Dec-2020**.

Analyzed By:

Approved By:

Signature: \_\_\_\_\_



Name: Wallace R. Loder, Jr.

Date: 19-Jun-2020

Signature: \_\_\_\_\_



Name: Wallace R. Loder, Jr.

Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.

It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072

**CERTIFICATE OF LEAK TEST**

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	2980CG
<b>Model #:</b>	GS400	<b>millicuries:</b>	30
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	vac tower top	<b>Test Number:</b>	2S-LE-0136B
<b>Test Date:</b>	12/9/19		

**The total removable amount of activity was analyzed to be:**

**0.0000 microcuries**

**Laboratory Analysis Number** AAU029      **Method:**  Wipe     Other

Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Jun-20**.

Analyzed By:

Approved By:

Signature: WRJ

Name: Wallace R. Loder, Jr.

Date: 17-Dec-19

Signature: Wallace Loder Jr.

Name: Wallace R. Loder, Jr.

Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.  
It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072

# Tracer Technologies International, Inc.

24600 Center Ridge Road, Suite 145  
Westlake, Ohio 44145  
Phone (440) 641-1133  
Fax (216) 464-6141

## CERTIFICATE OF LEAK TEST

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	3661CN
<b>Model #:</b>	GS400	<b>millicuries:</b>	30
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	vac tower bottom	<b>Test Number:</b>	2S-LE-0136A
<b>Test Date:</b>	12/9/19		

The total removable amount of activity was analyzed to be:

**0.0000 microcuries**

**Laboratory Analysis Number** AAU030      **Method:**     Wipe     Other

Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Jun-20**.

Analyzed By:

Approved By:

Signature: WRL

Name: Wallace R. Loder, Jr.

Date: 17-Dec-19

Signature: Wallace Loder Jr.

Name: Wallace R. Loder, Jr.

Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.

It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072



**Tracer Technologies  
International, Inc.**

24600 Center Ridge Road, Suite 145  
Westlake, Ohio 44145  
Phone (440) 641-1133  
Fax (216) 464-6141

**CERTIFICATE OF LEAK TEST**

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	M3617
<b>Model #:</b>	SA-1 F37	<b>millicuries:</b>	150
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Coker Frac tower bottom	<b>Test Number:</b>	36-LI-1701
<b>Test Date:</b>	12/9/19		

**The total removable amount of activity was analyzed to be:**

**0.0000 microcuries**

**Laboratory Analysis Number** AAU031      **Method:**      X   Wipe       Other

   Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

  X   Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Jun-20.**

Analyzed By:

Approved By:

Signature:           WRL            
Name:        Wallace R. Loder, Jr.  
Date:        17-Dec-19

Signature:           Wallace Loder Jr            
Name:        Wallace R. Loder, Jr.  
Title:        Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.  
It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072

# Tracer Technologies International, Inc.

24600 Center Ridge Road, Suite 145  
Westlake, Ohio 44145  
Phone (440) 641-1133  
Fax (216) 464-6141

## CERTIFICATE OF LEAK TEST

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	M1448
<b>Model #:</b>	SA-1 F37	<b>millicuries:</b>	500
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Coker, D704 top	<b>Test Number:</b>	36-LE-1704A
<b>Test Date:</b>	12/9/19		

The total removable amount of activity was analyzed to be:

**0.0000 microcuries**

**Laboratory Analysis Number** AAU032      **Method:**     Wipe     Other

Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Jun-20**.

Analyzed By:

Approved By:

Signature: WRL

Name: Wallace R. Loder, Jr.

Date: 17-Dec-19

Signature: Wallace Loder Jr

Name: Wallace R. Loder, Jr.

Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.

It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072

## CERTIFICATE OF LEAK TEST

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	76334
<b>Model #:</b>	SA-1 F37	<b>millicuries:</b>	500
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Coker, D704 middle	<b>Test Number:</b>	36-LE-1704B
<b>Test Date:</b>	12/9/19		

**The total removable amount of activity was analyzed to be:**

**0.0000 microcuries**

**Laboratory Analysis Number** AAU033      **Method:**     Wipe     Other


Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Jun-20**.

Analyzed By:

Approved By:

Signature: \_\_\_\_\_



Name: Wallace R. Loder, Jr.

Date: 17-Dec-19

Signature: \_\_\_\_\_



Name: Wallace R. Loder, Jr.

Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.

It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072

# Tracer Technologies International, Inc.

24600 Center Ridge Road, Suite 145  
Westlake, Ohio 44145  
Phone (440) 641-1133  
Fax (216) 464-6141

## CERTIFICATE OF LEAK TEST

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	75296
<b>Model #:</b>	SA-1 F37	<b>millicuries:</b>	500
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Coker, D704 bottom	<b>Test Number:</b>	36-LE-1704C
<b>Test Date:</b>	12/9/19		

The total removable amount of activity was analyzed to be:

**0.0000 microcuries**

**Laboratory Analysis Number** AAU034      **Method:**     Wipe     Other

Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Jun-20**.

Analyzed By:

Approved By:

Signature: WRL

Name: Wallace R. Loder, Jr.

Date: 17-Dec-19

Signature: Wallace Loder Jr

Name: Wallace R. Loder, Jr.

Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.  
It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072

**CERTIFICATE OF LEAK TEST**

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	76327
<b>Model #:</b>	SA-1 F37	<b>millicuries:</b>	500
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Coker, D705 top	<b>Test Number:</b>	36-LE-1705A
<b>Test Date:</b>	12/9/19		

**The total removable amount of activity was analyzed to be:**

**0.0000 microcuries**

**Laboratory Analysis Number** AAU035      **Method:**     Wipe     Other

Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Jun-20**.

Analyzed By:

Approved By:

Signature: WRL

Name: Wallace R. Loder, Jr.

Date: 17-Dec-19

Signature: Wallace Loder Jr.

Name: Wallace R. Loder, Jr.

Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.

It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072



**CERTIFICATE OF LEAK TEST**

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	M1452
<b>Model #:</b>	SA-1 F37	<b>millicuries:</b>	500
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Coker, D705 middle	<b>Test Number:</b>	36-LE-1705B
<b>Test Date:</b>	12/9/19		

**The total removable amount of activity was analyzed to be:**

**0.0000 microcuries**

**Laboratory Analysis Number** AAU036      **Method:**     **Wipe**     **Other**

Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Jun-20**.

Analyzed By:

Approved By:

Signature:



Name: Wallace R. Loder, Jr.  
Date: 17-Dec-19

Signature:



Name: Wallace R. Loder, Jr.  
Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.  
It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072

**CERTIFICATE OF LEAK TEST**

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	M1449
<b>Model #:</b>	SA-1 F37	<b>millicuries:</b>	500
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Coker, D705 bottom	<b>Test Number:</b>	36-LE-1705C
<b>Test Date:</b>	12/9/19		

The total removable amount of activity was analyzed to be:

**0.0000 microcuries**

**Laboratory Analysis Number** AAU037      **Method:**      X   Wipe       Other

       Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

  X   Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Jun-20**.

Analyzed By:

Approved By:

Signature:



Name:

Wallace R. Loder, Jr.

Date:

17-Dec-19

Signature:



Name:

Wallace R. Loder, Jr.

Title:

Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.  
It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072



**Tracer Technologies  
International, Inc.**

24600 Center Ridge Road, Suite 145  
Westlake, Ohio 44145  
Phone (440) 641-1133  
Fax (216) 464-6141

**CERTIFICATE OF LEAK TEST**

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	M1450
<b>Model #:</b>	SA-1 F37	<b>millicuries:</b>	500
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Coker, D706 top	<b>Test Number:</b>	36-LE-1706A
<b>Test Date:</b>	12/9/19		

**The total removable amount of activity was analyzed to be:**

**0.0000 microcuries**

**Laboratory Analysis Number** AAU038      **Method:**      X   Wipe       Other

   Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

  X   Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Jun-20**.

Analyzed By:

Approved By:

Signature:           WRL            
Name: Wallace R. Loder, Jr.  
Date: 17-Dec-19

Signature:           Wallace Loder Jr            
Name: Wallace R. Loder, Jr.  
Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.  
It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072

# Tracer Technologies International, Inc.

24600 Center Ridge Road, Suite 145  
Westlake, Ohio 44145  
Phone (440) 641-1133  
Fax (216) 464-6141

## CERTIFICATE OF LEAK TEST

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	M1447
<b>Model #:</b>	SA-1 F37	<b>millicuries:</b>	500
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Coker, D706 middle	<b>Test Number:</b>	36-LE-1706B
<b>Test Date:</b>	12/9/19		

The total removable amount of activity was analyzed to be:

**0.0000 microcuries**

**Laboratory Analysis Number** AAU039      **Method:**     Wipe     Other

Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Jun-20**.

Analyzed By:

Approved By:

Signature: WRL

Name: Wallace R. Loder, Jr.

Date: 17-Dec-19

Signature: Wallace Loder Jr

Name: Wallace R. Loder, Jr.

Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.  
It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072

**Tracer Technologies  
International, Inc.**

24600 Center Ridge Road, Suite 145  
Westlake, Ohio 44145  
Phone (440) 641-1133  
Fax (216) 464-6141

**CERTIFICATE OF LEAK TEST**

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82007

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	75291
<b>Model #:</b>	SA-1 F37	<b>millicuries:</b>	500
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	Coker, D706 bottom	<b>Test Number:</b>	36-LE-1706-C
<b>Test Date:</b>	12/9/19		

**The total removable amount of activity was analyzed to be:**

**0.0000 microcuries**

**Laboratory Analysis Number** AAU040      **Method:**      X   Wipe       Other

   Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

  X   Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Jun-20**.

Analyzed By:

Approved By:

Signature:                       
Name: Wallace R. Loder, Jr.  
Date: 17-Dec-19

Signature:                       
Name: Wallace R. Loder, Jr.  
Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.  
It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072

# Tracer Technologies International, Inc.

24600 Center Ridge Road, Suite 145  
Westlake, Ohio 44145  
Phone (440) 641-1133  
Fax (216) 464-6141

## CERTIFICATE OF LEAK TEST

Holly Frontier Refining  
2700 East 5th Street  
Cheyenne, WY 82003

This sealed source was tested for leakage of its radioactive material as follows:

<b>Manufacturer:</b>	Ronan	<b>Serial #:</b>	75297
<b>Model #:</b>	SA-1 F37	<b>millicuries:</b>	500
<b>Radioisotope:</b>	Cs-137	<b>Tested By:</b>	James Brown
<b>Source Location:</b>	FCCU, Frac tower bottom	<b>Test Number:</b>	34-HS-0106
<b>Test Date:</b>	12/9/19		

The total removable amount of activity was analyzed to be:

**0.0000 microcuries**

**Laboratory Analysis Number** AAU041      **Method:**  Wipe     Other

Analysis indicated 0.005 microcurie or more of removable radioactivity. NRC/Agreement State regulations require that the source be removed from service and decontaminated or repaired or replaced as appropriate. Notify the proper regulatory agency about leaking source.

Analysis indicated less than 0.005 microcurie of removable radioactivity. The source may continue to be properly used. The source is due for leak testing on or before: **09-Jun-20**.

Analyzed By:

Approved By:

Signature:



Name: Wallace R. Loder, Jr.  
Date: 17-Dec-19

Signature:



Name: Wallace R. Loder, Jr.  
Title: Radiation Safety Officer

**This certificate should be retained for inspection by appropriate regulatory agency.**

The above analysis was performed on a Packard Instruments 2300TR Liquid Scintillation Counter.  
It is calibrated by a NIST traceable standard for the isotope defined above.

This Leak Test Evaluation is authorized by the  
State of Ohio Department of Health  
Bureau of Radiation Protection  
License # 03225180072

**From:** [Brown, James](#)  
**To:** [Hill, Carol](#)  
**Subject:** [External\_Sender] Radiation Closure Form  
**Date:** Wednesday, January 27, 2021 12:21:00 PM  
**Attachments:** [SP-CHY01921012708590.pdf](#)  
[2020-1.pdf](#)  
[2020.pdf](#)

---

Please see attachments

Good morning Carol, please let me know if I've missed anything.

**Thank You,**  
**James Brown**  
**Safety Representative II**  
**Office – 307.771.8739**  
**Cell – 307.286.1077**

**From:** P\_CHY019@hollyfrontier.com <P\_CHY019@hollyfrontier.com>  
**Sent:** Wednesday, January 27, 2021 8:59 AM  
**To:** Brown, James <James.Brown@HollyFrontier.com>  
**Subject:** Message from P-CHY019

CONFIDENTIALITY NOTICE: This e-mail, and any attachments, may contain information that is privileged and confidential. If you received this message in error, please advise the sender immediately by reply e-mail and do not retain any paper or electronic copies of this message or any attachments. Unless expressly stated, nothing contained in this message should be construed as a digital or electronic signature or a commitment to a binding agreement.



**ACKNOWLEDGEMENT - RECEIPT OF CORRESPONDENCE**

<b>Name and Address of Applicant and/or Licensee</b>  Steve Hollis Safety Manager HollyFrontier Cheyenne Refining LLC 300 Morrie Avenue Cheyenne, WY 82007	<b>Date</b> 025/01/2021
	<b>License Number(s)</b> 49-26892-01
	<b>Mail Control Number(s)</b> 624444
	<b>Licensing and/or Technical Reviewer or Branch</b> C. Hill

This is to acknowledge receipt of your:  Letter and/or  Application Dated: 01/28/2021

The initial processing, which included an administrative review, has been performed.  
 Amendment  Termination  New License  Renewal

There were no administrative omissions identified during our initial review.

This is to acknowledge receipt of your application for renewal of the material(s) license identified above. Your application is deemed timely filed, and accordingly, the license will not expire until final action has been taken by this office.

Your application for a new NRC license did not include your taxpayer identification number. Please complete and submit NRC Form 531, Request for Taxpayer Identification Number, located at the following link: <http://www.nrc.gov/reading-rm/doc-collections/forms/nrc531.pdf>  
 Follow the instructions on the form for submission.

The following administrative omissions have been identified:  
 [Empty box for listing omissions]

Your application has been assigned the above listed MAIL CONTROL NUMBER. When calling to inquire about this action, please refer to this control number. Your application has been forwarded to a technical reviewer. Please note that the technical review, which is normally completed within 180 days for a renewal application (90 days for all other requests), may identify additional omissions or require additional information. If you have any questions concerning the processing of your application, our contact information is listed below:

<b>Region IV</b> <b>U. S. Nuclear Regulatory Commission</b> <b>DNMS/NMSB - B</b> <b>1600 E. Lamar Boulevard</b> <b>Arlington, TX 76011-4511</b> <b>(817) 200-1103 or (817) 200-1140</b>
--



BETWEEN:

Accounts Receivable/Payable  
and  
Regional Licensing Branches

[ FOR ARPB USE ]  
INFORMATION FROM WBL

Program Code: 03120  
Status Code: Pending Termination  
Fee Category: 3P  
Exp. Date:  
Fee Comments:  
Decom Fin Assur Req: N

## License Fee Worksheet - License Fee Transmittal

### A. REGION

#### 1. APPLICATION ATTACHED

Applicant/Licensee: HollyFrontier Cheyenne Refining LLC  
Received Date: 01/27/2021  
Docket Number: 3030129  
Mail Control Number: 624444  
License Number: 49-26892-01  
Action Type: Termination

#### 2. FEE ATTACHED

Amount:       N/A      

Check No.:       N/A      

#### 3. COMMENTS

Signed:       Carol L. Hill      

Date:       02/01/2021      

### B. LICENSE FEE MANAGEMENT BRANCH (Check when milestone 03 is entered / / )

1. Fee Category and Amount: \_\_\_\_\_

#### 2. Correct Fee Paid. Application may be processed for:

Amendment: \_\_\_\_\_

Renewal: \_\_\_\_\_

License: \_\_\_\_\_

#### 3. OTHER \_\_\_\_\_

\_\_\_\_\_

Signed: \_\_\_\_\_

Date: \_\_\_\_\_

Agency: NRC

WBL WORKSHEET

DOCKET NUMBER: 3030129 LICENSE NUMBER: 49-26892-01 STATUS: Pending Termination

MAIL CONTROL NUMBER: 624444 RECEIPT DATE: 01/27/2021 ACTION TYPE: Termination

DUE DATE: 04/27/2021 INST. CODE: 26892 LICENSE REGION: Region 4

LICENSE TYPE: 30 ENTITY TYPE: C LICENSE GROUP: Industrial

ISSUE DATE: ORIGINAL DATE: 09/24/1987 EXPIRATION DATE:

DECOMMISSIONING CATEGORY: Group 1 LAST ISSUE DATE:

LICENSEE NAME: HollyFrontier Cheyenne Refining LLC DECOM FIN ASSUR REQD: N  
SUBM: N

MAILING ADDRESS LINE1: 300 Morrie Avenue CONT PLAN REQD: N APPRV: N

MAILING ADDRESS LINE 2:

CITY: Cheyenne STATE: WY ZIP: 82007

CONTACT PERSON: PREFIX: FIRST NAME: Steve MIDDLE INITIAL:

LAST NAME: Hollis SUFFIX:

JOB TITLE: PHONE: 307-771-8925 FAX: 307-771-8794 EMAIL: Steven.Hollis@hollyfror

BILLING ADDRESS LINE 1:

BILLING ADDRESS LINE 2:

CITY: STATE: Wyoming ZIP:

BILLING CONTACT PERSON: FIRST NAME: MIDDLE INITIAL: LAST NAME:

PHONE: EMAIL: FAX:

PRIMARY PGM CODE: 03120 SECONDARY PGM CODE:

INSPECTION REGION: Region 4 PRIORITY: 5

RSO: PREFIX: FIRST NAME: James MIDDLE INITIAL: LAST NAME Brown

SUFFIX: Jr. RSO JOB TITLE: RSO

RSO PHONE: 307-771-8823 RSO FAX: 307-771-8739 RSO EMAIL: james.brown@hollyfrontier.com

STATES WHERE USE IS AUTHORIZED: 1  
0- ALL LISTED STATES  
1- SAME AS STATE IN ADDRESS  
2- ALL STATES  
3- NON-AGREEMENT-STATES

AUTHORIZED STATES (USE ONLY IF ABOVE IS ZERO):