
From: Martin, Michael <mimart@iu.edu>
Sent: Tuesday, February 02, 2021 9:30 AM
To: Tran, Frank <Frank.Tran@nrc.gov>
Subject: [External_Sender] RE: Request additional information for NRC License No. 13-02752-03

Hello Mr. Tran,

I hope this finds you well. I apologize for the delay in replying to you, as I have been out of office.

To clarify, we wish to retain line item 6.M. until the new source and unit are installed in order to minimize down time and delays in treatment. I have spoken with the vendor, who will provide source removal and decommissioning paperwork (including leak test) after the service engineer removes the source on the first day of the upgrade. We will forward this to you as soon as the source is no longer at our site and under our control, which we anticipate to occur a day or two after removal.

Please let me know if this adequately answers your request.

Kind regards,

Michael

T. Michael Martin, PhD, DABHP

Director of Health Physics & Radiation Safety Officer
IUPUI/IUMC, IU Health Academic Health Center, Eskenazi Hospital
Adjunct Assistant Professor of Radiology & Imaging Sciences
Indiana University Environmental Health and Safety
Gatch Hall, CL159
1120 W. Michigan St.
Indianapolis, IN 46202
Office: (317) 274-0331
Pager: (317) 312-1519
mimart@iu.edu
tmartin24@iuhealth.org
<https://ehs.iupui.edu/radiation/index.html>



[Please Click Here to Complete an EHS Customer Satisfaction Survey](#)