

From: [Cender, Laura](#)
To: [Kerry D. Randles](#)
Subject: RE: Additional Requirements requested for License Renewal
Date: Friday, February 12, 2021 11:50:00 AM

Hello Kerry,

I hope you are doing well. I reviewed the files that you sent over and I have a few additional questions outline below. If you could give me a call when you have a chance I would be happy to discuss each item with you.

1. Please provide a cover letter for your previous submission. Please ensure that the cover letter includes the following:
 - a. A statement confirming that erosion of the facility shielding is evaluated as part of the regular surveying of the facility. Please clarify if this evaluation is performed annually or quarterly.
 - b. A statement confirming that that staff do not transport sources in vehicles, even for shipping or maintenance. Confirm that sources are picked up on site by the shipping company when returned to the vendor or sent for disposal.
2. Please provide the google maps image that shows the surrounding area. The attachment did not appear to come through on your previous response.
3. Please provide a letter from QSA Global, Inc. confirming their receipt of any QSA Global Inc. Model 680AE, 680BE, or Model 880 Delta (IR-192) sealed sources. Please ensure that the serial numbers indicated on the leak tests can be tied to the transfer record. Please note that a shipping form is not sufficient to demonstrate source transfer.
 - a. If no QSA Global, Inc. Model 880 Delta Ir-192 radiography cameras have ever been possessed under the license please provide a statement confirming such in the cover letter.

Please provide your response to the above items by no later than Friday, February 26, 2021.

Thank you,
Laura

Laura Cender
Pronouns: She/Her
U.S. Nuclear Regulatory Commission
Materials Licensing Branch
E-mail: Laura.Cender@nrc.gov
Phone: (630) 829-9712