

From: [Gryglak, Magdalena](#)
To: ["Anzalone, Sebastiano"](#)
Subject: Request to add areas of use, NRC license no. 24-02704-02
Date: Tuesday, April 06, 2021 9:53:00 PM

Good evening Mr. Anzalone,

I have reviewed the licensee's request dated 12/8/20 to add areas of use.

In a signed by Radiation Safety Officer and dated letter, please provide the following:

1. Specify the type of material to be used in these areas of use (please confirm you will use 10 CFR 35.100 and 35.200 material);
2. Confirm no Positron Emission Tomography (PET) material will be used;
3. If 10 CFR 35.300 material will be used, please confirm that the rooms will not be in patient rooms (all patients will be released in accordance with 10 CFR 35.75);
4. Please provide dimensions for Echo Room 1, Echo Room 3, Echo Room 4, Vascular Room 1, and Vascular Room 2 (the scale/dimensions on the floor plans were not legible).

Please provide the information as a pdf file via email directly to me.

Thank you

Magdalena R. Gryglak
Health Physicist
U.S. NRC Region III