

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

<b>1. LICENSEE/LOCATION INSPECTED</b>  Camden-Clark Memorial Hospital Corporation 800 Garfield Avenue P.O. Box 718 Parkersburg, WV 26102	<b>2. NRC/REGIONAL OFFICE</b>  Region 1 2100 Renaissance Blvd Suite 100 King of Prussia, PA 19406-2713	
<b>REPORT NUMBER(S) 2021001</b>		
<b>3. DOCKET NUMBER(S)</b> 030-03390	<b>4. LICENSE NUMBER(S)</b> 47-09772-02	<b>5. DATE(S) OF INSPECTION</b> 03/01/2021 - 03/02/2021 on site, 3/30/2021 exit

**LICENSEE:**

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC ) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

\_\_\_\_\_ Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with the NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11. (Violations and Corrective Actions)

10 CFR 35.61(a) requires, in part, that licensees annually calibrate the survey instruments used to show compliance with this part and 10 CFR Part 20.

Contrary to the above, in 2017, Camden-Clark Memorial Hospital did not annually calibrate a survey instrument used to show compliance with this part and 10 CFR Part 20. Specifically, a Ludlum 9DP survey meter, serial number 25018912, was calibrated on August 13, 2019, not calibrated again until September 20, 2020, and was used to perform post implant surveys in Radiation Oncology on August 27, 2020 and September 4, 2020 when the calibration was expired.

This is a Severity Level IV violation. (Enforcement Policy Section 6.3)

Corrective actions included calibrating the meter as well as revising the form used to record the survey to include the calibration date of the instrument used to conduct the post implant survey.

**Statement of Corrective Actions**

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE	Daniel Berkley, RSO		4/7/21
NRC INSPECTOR	Robin Elliott	Robin L. Elliott <small>Digitally signed by Robin L. Elliott Date: 2021.04.05 20:39:24 -04'00'</small>	
BRANCH CHIEF	Donna Janda		