

From: [Gryglak, Magdalena](#)
To: [Laura T. Smith- Physics](#)
Subject: RE: renewal for Bronson Methodist Hospital
Date: Friday, May 7, 2021 1:32:00 PM

Good afternoon,

I have reviewed your response and have just a few questions:

1. Regarding Dr. William Campbell, I believe I have no choice but to remove him from the license since his medical license has expired. I assume he is not practicing medicine.
2. Because Y-90 TheraSphere material is listed on the license, Ms. Ritchlin needs to receive training in radiation safety, regulatory issues, and emergency procedures (similar to SirSpheres). The form 313 A(RSO) indicates that Ms. Ritchlin was only trained in SirSpheres. Was this the case (she never worked with TheraSpheres in any of the other hospitals before?) or does the form need to be corrected/resubmitted to also include TheraSpheres?
3. Please confirm in the email that you do not have any in-patient treatment rooms (i.e. all patients are released in accordance with 10 CFR 35.75).
4. Also, I need you to be aware that all the procedures that you submitted in your original renewal application dated 1/15/21 will be tied down in the license with the application.
5. Please provide Ms. Ritchlin's contact info (email and phone number) so I can update our records for the new RSO.

Thank you

Magdalena R. Gryglak

U.S NRC Region III