

From: [Nguyen, Jan](#)
To: [Kelly Stoneberg](#)
Cc: mhelyv@srmcwv.org
Subject: NRC Request for Additional Information for Summersville Regional Medical Center (Mail Control Number 625488)
Date: Monday, June 07, 2021 8:10:00 PM

Licensee: The West Virginia Health Care Cooperative, Inc.
d/b/a Summersville Regional Medical Center
License No: 47-31126-01
Docket No: 030-37131
Control No: 625488

Dear Mr. Roe,

Could you please reply back to this email to confirm receipt?

This is in reference to your letters dated April 12, and May 21, 2021, requesting to amend Nuclear Regulatory Commission License No. 47-31126-01.

You have requested to add Braxton County Hospital, located at 100 Hoylman Drive, Gassaway, WV, as a location of use. In order to continue our review, we need the following additional information:

1. Please indicate if this will be a permanent location of use with all equipment, radiopharmaceuticals, sealed sources, and all associated wastes remaining secured on-site and dosages of byproduct material being delivered directly to the hot lab (i.e., not serviced as a mobile medical provider).
2. Please confirm that no PET studies will be conducted at Braxton County Hospital.
3. Please provide the location and room numbers (if applicable) of each room on the facility diagram.
4. Please provide a description of the instrumentation (e.g., gamma counter, solid-state detector, portable or stationary count-rate meter, portable or stationary dose-rate or exposure-rate meter, single or multichannel analyzer, liquid scintillation counter, proportional counter) that will be used to perform required surveys at Braxton County Hospital.
5. Please provide a description of the equipment used to measure the dosages at Braxton County Hospital. Alternatively, you may confirm that you are only using unit dosages and will determine activity by decay correction.
6. In your letter dated April 15, 2021, you submitted a list of seven sealed sources that you would be taking possession of from CDL Nuclear Technologies, Inc. In your subsequent letter dated May 21, 2021, you confirmed that you took possession of eight sealed sources from CDL and attached the inventory record. The eight sealed sources did not match the seven sources from the April 15, 2021 letter (i.e., different serial numbers, activities, isotopes, etc.). Please confirm that the list of eight sealed

sources in the May 21, 2021 letter are the correct sources that were received and entered into your inventory.

7. Please confirm that there was no contamination at the time that the site was transferred from CDL Nuclear Technologies, Inc.
8. Please confirm you are leasing the space at Braxton County Hospital. If you are leasing the space and the location of use will be controlled by an entity other than Summersville Regional Medical Center, please provide documentation of a clear, contractual agreement concerning access to your location of use for the purpose of decontamination or removal of licensed material from the location of use in the event of disharmony between you and the owner entity. This documentation should consist of signed certification from both parties, and should also address the permitted use of byproduct material at the client's address and the authority and responsibilities of the mobile service provider and the client.
9. Please confirm that you will have full responsibility over the site, in order to ensure compliance with all applicable regulatory requirements.
10. Please describe the controls (e.g., control of keys, badge access) that will be implemented to ensure that licensed materials are under Summersville Regional Medical Center's control and not accessible to unauthorized individuals (i.e. hospital staff). Furthermore, will non-licensee staff have unescorted access to the licensed areas of use or will Summersville Regional Medical Center have control of these areas?
11. Please confirm that instructions will be given to supervised individuals in written radiation protection procedures, regulations, and license conditions with respect to the use of byproduct material in accordance with 10 CFR 35.27, in addition to the requirements in 10 CFR 19.12.

We will continue our review upon receipt of the requested information. You may respond to my attention in writing by letter, email (if letter is signed by senior management and scanned into a pdf format), or fax (610-337-5269), referencing mail control number 625488. Please submit a reply within 30 days. If you have any technical questions regarding this deficiency letter, please call me at (610) 337-5006.

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Med, Ind, & Academic Uses**; then **Licensee Toolkits, see our toolkit index page**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 a.m. to 5:30 p.m. EST, Monday through Friday (except Federal holidays).

Thank you in advance for your help!

Jan

Janice Nguyen
Senior Health Physicist
Division of Radiological Safety and Security (DRSS)
Medical and Licensing Assistance Branch (MLAB)
U.S. Nuclear Regulatory Commission Region I
2100 Renaissance Boulevard, Suite 100
King of Prussia, PA 19406
Office (610) 337-5006
FAX (610) 337-5269