



UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
REGION IV  
1600 E. LAMAR BLVD  
ARLINGTON TX 76011-4511

June 10, 2021

Kyle T. Jones  
Radiation Safety Officer  
Providence Health & Services-Washington  
d/b/a Providence Sacred Heart Medical Center  
& Children's Hospital, Cardiovascular Imaging Centers  
101 W. 8<sup>th</sup> Avenue  
Spokane, WA 99204

SUBJECT: TERMINATION OF NRC RADIOACTIVE MATERIALS LICENSE NO. 46-27732-01

By letter dated March 22, 2021, you informed the U.S. Nuclear Regulatory Commission (NRC) that Providence Health & Services-Washington wished to terminate NRC radioactive materials license number 46-27732-01. The NRC Form 314 dated May 5, 2021, stated that all radioactive material had been transferred to a licensed entity and disposed by decay-in-storage method.

The supporting documentation to the NRC Form 314 provided the most recent leak test results and final status survey. The leak test results demonstrate no evidence of leakage of radioactive material from the sealed sources. In addition, a review of the historical file demonstrates that only radioactive material with a radioactive decay half-life of less than 120 days were possessed by the licensee. Based on the documentation submitted and a review of your historical file, the staff concluded that all licensable radioactive material has been removed from your facility. The facility located at Valley Medical Center, 2315 8<sup>th</sup> Street, 2<sup>nd</sup> floor, Lewiston, Idaho 83501, is suitable for unrestricted use in accordance with 10 CFR 20.1402, "Radiological Criteria for Unrestricted Use." Based on these conclusions, no further remediation with respect to NRC regulated material is required.

Please find enclosed Amendment No. 7 to License No. 46-27732-01 authorizing your license termination. An environmental assessment for this action is not required since this action is categorically excluded under 10 CFR 51.22(c)(20)(i) and (ii). You can contact me at 817-200-1189 if you have questions on any of the information stated above.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Sincerely,

Roberto J. Torres, M.S., Senior Health Physicist  
Materials Licensing and Decommissioning Branch

Docket: 030-36016  
License: 46-27732-01  
Control: 626953