

From: [Tran, Frank](#)
To: bcaudle@mpcphysics.com
Subject: Request for additional information for NRC License No. 13-18992-01
Date: Thursday, May 20, 2021 2:37:00 PM

Dear Mr. Caudle:

We have reviewed the renewal application for NRC License No. 13-18992-01 for Decatur County Memorial Hospital in accordance with the NRC regulations, guidance, and policies, including but not limited to NUREG-1556, Volumes 9 and 20. Based on NUREG-1556, Volume 9, Revision 3 (<https://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/index.html>), we will need the following information to complete our review.

NUREG-1556, Volume 9, Revision 3, Item 9, "Facility Diagram", discusses information needed for the facility diagram. It asks for the diagram scale or facility dimensions. We reviewed the diagram named "Decatur County Memorial Hospital Nuclear Medicine Department" and could not locate the scale or dimension. Please resubmit that diagram with scale or dimensions.

The licensee also stated that administration of 35.300 materials will be performed in the department. Does that mean the administration of 35.300 materials will be in the Nuclear Medicine room (or camera room)?

Please provide the response in writing with authorized signature and date on the cover letter by June 21, 2021. To facilitate proper mailing handling in our office, please reference Mail Control No. 624429 in the cover letter. If you have any questions or need a clarification, please do not hesitate to contact me at 630-829-9623 or reply to this email.

Thank you,

Frank Tran

Health Physicist/License Reviewer
NRC Region III/Division of Nuclear Materials Safety
Phone: 630-829-9623
Fax: 630-515-1078
Email: Frank.Tran@nrc.gov

